



Published in final edited form as:

Minerals (Basel). 2022 October 18; 12(10): . doi:10.3390/min12101314.

Review of Filters for Air Sampling and Chemical Analysis in Mining Workplaces

Judith C. Chow^{1,*}, John G. Watson¹, Xiaoliang Wang¹, Behrooz Abbasi², Wm. Randolph Reed³, David Parks⁴

¹Division of Atmospheric Sciences, Desert Research Institute, Reno, NV 89511, USA

²Department of Mining and Metallurgical Engineering, University of Nevada, Reno, NV 89557, USA

³Office of the Director, National Institute for Occupational Safety and Health, Pittsburgh, PA 15236, USA

⁴Spokane Mining Research Division, National Institute for Occupational Safety and Health, Spokane, WA 99207, USA

Abstract

This review considers the use of filters to sample air in mining workplace environments for dust concentration measurement and subsequent analysis of hazardous contaminants, especially respirable crystalline silica (RCS) on filters compatible with wearable personal dust monitors (PDM). The review summarizes filter vendors, sizes, costs, chemical and physical properties, and information available on filter modeling, laboratory testing, and field performance. Filter media testing and selection should consider the characteristics required for mass by gravimetry in addition to RCS quantification by Fourier-transform infrared (FTIR) or Raman spectroscopic analysis. For mass determination, the filters need to have high filtration efficiency (99% for the most penetrable particle sizes) and a reasonable pressure drop (up to 16.7 kPa) to accommodate high dust loading. Additional requirements include: negligible uptake of water vapor and gaseous

This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

*Correspondence: judith.chow@dri.edu.

Author Contributions: Conceptualization, J.C.C., B.A., X.W., J.G.W., W.R.R. and D.P.; investigation, J.C.C., X.W. and J.G.W.; writing—original draft preparation, J.C.C., X.W. and J.G.W.; writing—review and editing, J.C.C., B.A., X.W., J.G.W., W.R.R. and D.P.; project administration, X.W. and B.A.; funding acquisition, X.W. and B.A. All authors have read and agreed to the published version of the manuscript.

Conflicts of Interest: The authors declare no conflict of interest. The funders had no role in the design of the study; in the collection, analyses, or interpretation of data; in the writing of the manuscript; or in the decision to publish the results.

Disclaimer: The findings and conclusions in this report are those of the author(s) and do not necessarily represent the official position of the National Institute for Occupational Safety and Health (NIOSH), Centers for Disease Control and Prevention (CDC). Mention of any company or product does not constitute endorsement by NIOSH, CDC.

Supplementary Materials: The following are available online at <https://www.mdpi.com/article/10.3390/min12101314/s1>: Table S1. Summary of filter characteristics for sampling and analysis; Table S2. Mass concentration under uncontrolled and controlled temperature and humidity environments; Table S3. Major infrared bands of the principal polymorphs in crystalline silica; Figure S1. Historical examples of mine safety equipment; Figure S2. Coal mine dust personal sampler unit (CMDPSU); Figure S3. Personal dust monitor (PDM3700); Figure S4. Scanning electron microscopic images for different filters; Figure S5. Penetration as a function of particle diameter for different filters; Figure S6. Pressure drop as a function of sampling flow rate for silver membrane filters; Figure S7. Absorption of blank membrane filters; Figure S8. FTIR spectra for blank Teflon-coated glass-fiber filters.

volatile compounds; adequate particle adhesion as a function of particle loading; sufficient particle loading capacity to form a stable particle deposit layer during sampling in wet and dusty environments; mechanical strength to withstand vibrations and pressure drops across the filter; and appropriate filter mass compatible with the tapered element oscillating microbalance. FTIR and Raman measurements require filters to be free of spectral interference. Furthermore, because the irradiated area does not completely cover the sample deposit, particles should be uniformly deposited on the filter.

Keywords

filter; respirable coal mine dust; respirable crystalline silica; personal dust monitor; FTIR; Raman; chemical speciation

1. Introduction

Filters are used to remove solid and liquid particles from air. The most widespread application is for air purification. Most heating and ventilation systems have some sort of filter that is regularly changed, as do air intakes on mobile and stationary engines. The earliest record of air filtration dates back to Roman times related to dust in Egyptian mines. Loose bladders were used to prevent dust inhalation and minimize health risks [1]. Use of filtering respirators [2] started in the 1800s for the removal of airborne microorganisms [3]. Early applications also included filters to protect firefighters from smoke inhalation and wetted fabrics or gas masks as protection from fumes derived from chemical warfare [4].

Filtration science was systematized and advanced in the mid-1960s by Fuchs [5] and Spurný [6] and later by Davies [7,8], who documented the history, development, and application of aerosol filtration. Filtration by Brownian motion of small particles and by impaction and interception of large particles was formulated by Langmuir [9,10] and Kaufman [11]. Later refinements applied 3D computer modeling and laboratory filter tests that revealed differences between the structures of fiber and membrane filters [1,12].

Aerosol filtration has been applied in personal protective equipment (PPE, www.who.int, accessed on 1 October 2022), industrial hygiene [13], occupational safety and health (www.osha.gov, accessed on 1 October 2022), liquid purification (by sterilization of heat-sensitive materials) [14], microplastic extraction [15,16], bioaerosols [17,18], pathogen and microorganism removal [19,20], electrospun nanofiber filtration [21], indoor air filtration, vehicle exhaust filtration, and industrial emission controls (removing particles from process gas streams). These applications emphasize particle removal rather than particle characterization. Particle characterization imposes additional constraints on the filter media in terms of blank levels, distributions across the filter surface and within the filter thickness, resistance and acquiescence to chemical extraction, and transparency to electromagnetic probes.

This review considers the use of filters to sample air in mining workplace environments for subsequent analysis of hazardous contaminants. Objectives are to: (1) survey and critically evaluate information on filters amenable to particulate matter (PM) mass and further

laboratory analysis, with an emphasis on quantifying respirable crystalline silica (RCS) on filters compatible with wearable workplace dust monitors; (2) summarize filter vendors, sizes, costs, chemical and physical properties, and information available on filter modeling, laboratory testing, and field performance; (3) identify knowledge gaps and methods to fill them; and (4) recommend filter media suitable for both mass measurement by a personal dust monitor (PDM) and RCS quantification by Fourier-transform infrared (FTIR) and/or Raman spectroscopy analyses.

Several reviews of filter media for aerosol characterization have been published in the past [22–29], and these were examined first. While containing useful information, these reviews were found to be dated. Many of the manufacturers, vendors, and filter materials have changed since their publication, and more recent resources on filter characteristics and specific applications are available. This review intends to update the knowledge base with a specific focus on workplace applications in mining environments.

2. Ambient and Workplace Aerosol Sampling and Analysis

PM mass concentrations are used as indicators of human exposure in both ambient air and workplace environments. Samplers typically consist of a size-selective inlet, such as an impactor or cyclone, a filter holder, a collection filter through which air is drawn and onto which the PM is deposited, a pump or fan that moves the air, and a flow controller that maintains a constant volumetric sampling rate [30]. These samplers are operated for a set period of time, 24 h for ambient air and a normal work shift (8 to 12 h) for workplace monitoring. Filters are weighed before and after sampling [31], with the difference in weights divided by the flow rate times the sample duration to provide PM concentrations in $\mu\text{g}/\text{m}^3$ or mg/m^3 . Continuous in situ methods used are: (1) filter tapes that advance when loaded with detection by beta-particle attenuation [32]; (2) replaceable filters that are continuously sensed by an inertial microbalance [33]; (3) counting of particles in different size ranges based on individual particle light scattering [34]; or (4) light scattering by an ensemble of particles calibrated against an aerosol mixture similar to that of the monitored environment [35,36].

Ambient air is sampled at fixed outdoor locations for comparison with the U.S. Environmental Protection Agency's National Ambient Air Quality Standards (NAAQS), while workplace monitoring intends to determine individual exposure using wearable devices. Ambient PM monitoring networks [37,38] also include a subset of chemical speciation monitors for elements, soluble ions, and carbon fractions that are used to quantify source contributions, evaluate effects of PM components on visibility and health, and track air quality improvements with emission reduction strategies. Owing to the need for such characterization, a large body of information is available for ambient PM filter sampling and analysis that is relevant to the workplace. There is little information in the ambient PM literature on RCS, however, and this review provides greater scrutiny of filter media amenable to RCS analyses.

The Occupational Safety and Health Administration (OSHA, created in 1970) and Mine Safety and Health Administration (MSHA, established in 1977) develop and enforce

workplace safety and health regulations. MSHA programs and activities are specifically related to mining, while OSHA addresses a broader range of workplace environments. Both agencies intend to reduce mortality, injuries, and morbidity by monitoring worker inhalation exposure. Figure S1 (Supplemental Material) shows examples of mine rescue breathing equipment that has evolved over time to meet workplace safety needs. In particular, inhalation of coal mine dust has been associated with underground coal miners' pneumoconiosis and silicosis and is a motivation for exposure reduction.

To determine workers' exposure to silica-bearing dust in underground and surfacemines, the MSHA established methods for coal mine sampling [39]. Table 1 summarizes the specifications for two MSHA-approved coal mine sampling devices: the coal mine dust personal sampler unit (CMDPSU) and the continuous personal dust monitor (PDM3700), as shown in Figures S2 and S3, respectively. Both devices operate at low volumetric flow rates (2.0–2.2 L/min) to measure respirable coal mine dust (RCMD, $<4\ \mu\text{m}$ aerodynamic diameter, also termed $\text{PM}_{4.0}$) under local temperature and pressure conditions [40]. CMDPSU samples are acquired with filter cartridges that are weighed before and after personal sampling over a mining work shift [31].

The PDM is a tapered element oscillating microbalance (TEOM) [33,41], a miniaturized version of the ambient sampling system [42] that is a Federal Equivalent Method (FEM) for determining compliance with the 24 h PM_{10} (particles with aerodynamic diameters $<10\ \mu\text{m}$) NAAQS of $150\ \mu\text{g}/\text{m}^3$ (not to be exceeded more than once per year over three years). The PDM3700 (Figure S3) measures RCMD mass concentrations in real time for compliance with the regulatory mining exposure limit of $1.5\ \text{mg}/\text{m}^3$ over a full working shift. It is designed for mining applications with a worker-wearable (2 kg) system and PC-based software data retrieval.

Laboratory and field performance tests have found individual PDM increments within $\pm 25\%$ of reference sampler measurements for mass loadings between 0.5 and $4\ \text{mg}/\text{m}^3$. Coefficients of variation (CV) for collocated PDMs ranged from 4.2% for concentrations between 1.5 and $2\ \text{mg}/\text{m}^3$ to 6.0% for lower concentrations between 0.2 and $0.5\ \text{mg}/\text{m}^3$ [43–45]. The largest paired comparison ($n = 955$) between CMDPSUs and PDMs in coal mines showed no significant statistical differences between the two systems [46]. In-mine testing produced linear relationships between CMDPSU and PDM mass concentrations [47].

Pairwise evaluations of PDM3700s with cyclone samplers in real-world mines by Belle [48] found $\sim 50\%$ – 60% higher gravimetric mass concentration for the personal dust monitors. The large “measurement bias” between PDM3700 and gravimetric samples using a Higgins-Dewell (HD) cyclone [49] needs to be further investigated. Large variations in measured respirable dust concentrations were also reported by Verpaele and Jouret [50], who attributed the differences to $\sim 50\%$ oversampling by the SKC conductive black plastic sampler. These levels are double the $\pm 25\%$ comparability with the reference method specified by the manufacturer and reported by Belle [48]. The sampling effectiveness of different size-selective inlets [51] and the effects of sampler surfaces on particle collection should be considered when selecting appropriate filter media.

In addition to mass concentrations by gravimetry, the MSHA analyzes ~15,000 to 20,000 filter samples per year for RCS by FTIR spectrometry to enforce the RCMD standard [52]. RCS FTIR procedures are specified by MSHA [53] method P-7 and NIOSH [54] method 7603. Method P-7 includes ashing of the exposed polyvinyl chloride (PVC) filters and redepositing of the remainder onto PVC-acrylic copolymer membrane filters (DM-450 or DM-800) by isopropanol suspension prior to FTIR analysis. RCS mass is adjusted for interference from kaolinite, which is sometimes present in mine dust. The ashing and residue redepositing may introduce operational errors, but it intends to eliminate organic materials in both the coal dust deposit and the filters that might interfere with the FTIR spectra [29,55]. Additional measurement uncertainties may also result from inhomogeneities of the redeposited material on the new filter [52,56], as the FTIR beam is directed through only a section of the filter. To expedite RCS quantification, NIOSH has been developing a field-based RCS monitoring approach using a portable FTIR that can provide end-of-shift measurements on mining sites [57–63]. The fibrous filter mat in the PDM3700 consists of borosilicate fiberglass with a polytetrafluoroethylene (PTFE) polymer binder cured at 370 °C and backed with a woven glass fiber support (EMFAB™ TX40HI20WW, Pall Laboratory) that interferes with RCS quantification by FTIR and Raman spectroscopy. The PDM3700 filter holder is not designed to accommodate filter media with chemical stability and low absorbance for spectroscopic determination of chemical components, including RCS. Because the current PDM filter assembly is not amenable to RCS quantification [29,64], the CMDPSU with PVC filters is also needed for field-based FTIR analysis. Reduced sampling effort and increased monitoring efficiency can be achieved when both RCMD and RCS concentrations can be determined from the PDM filter.

3. Filter Characteristics

Table S1 summarizes chemical and physical characteristics of 12 filter types along with compatible physical and chemical analyses. The 11 vendors identified in Table S1 provide multiple filter media types and it is not clear that the vendors are also media manufacturers. When contacted for this review, vendors were not forthcoming on filter origins for proprietary reasons. To achieve mass closure with measured chemical components [65], it is often necessary to sample concurrently on multiple substrates.

Air sampling filters vary in material, structure, filter diameter, pore size, thickness, mechanical and temperature stability, chemical compatibility, blank concentrations, flow resistance, particle loading capacity, and collection efficiency [27]. With the exception of the fluorinated ethylene propylene (FEP) membrane film used as an impaction surface for cascade impactor sampling, filters are porous structures that accommodate different flow rates, flow pathways, residence times, and applications.

Table S1 categorizes the main classes of filters that are commonly used for aerosol sampling, including: (1) six types of membrane (also termed “porous membrane”) filters (i.e., PTFE, polypropylene, PVC, nylon, silver, and mixed cellulose esters [MCE]); (2) one type of capillary pore filter (i.e., polycarbonate); and (3) five types of fibrous filters (i.e., cellulose fiber, pure and mixed quartz fiber, Teflon-coated glass fiber, and glass fiber) [26,66]. The following sections: (1) describe materials and structures of these filter types;

(2) summarize past tests on filter collection efficiency; (3) discuss potential atmospheric artifacts; (4) examine effects of particle deposit inhomogeneities; and (5) tabulate filter costs and availability.

3.1. Filter Material and Structure

Figure S4 [67] compares the structures among porous membrane, capillary pore, and fibrous filters, showing different surface topographies. The largest contrast is found between glass- and quartz-fiber filters with randomly crossed fibers and those with uniform passages in the capillary pore filter. Porous membrane filters are gels formed from a colloidal solution having interconnected pores with uniform microstructures that capture particles on the filter surface while allowing the passage of air through the filter volume. They consist of different synthetic materials, including: (1) PTFE membrane, a microporous membrane made of a synthetic fluoropolymer of PTFE; (2) polypropylene, also known as polypropene, a thermoplastic polymer produced by chain-growth polymerization from monomer propylene; (3) PVC membrane, produced by free-radical polymerization of PVC, a polymer similar to polyethylene but with one of the hydrogen atoms replaced by chlorine atoms; (4) nylon membrane, made of diacid chlorides, diamines, polyamide, or thermoplastic polymers; (5) silver membrane, consisting of sintered pure metallic silver (~99.97%); and (6) MCE membrane, made of different cellulose molecules containing hydrocarbon polymers (e.g., carbon, hydrogen, and β -glucose). MCE is a mixture of cellulose esters, cellulose acetate, and cellulose nitrate with compositions varying among manufacturers. Most membrane filter disks are thin films (~30–70 μm) manufactured with various pore sizes (0.2–5 μm). These filters have adequate porosity (>85%) for high particle collection efficiency, but they require powerful flow movers to overcome resistance across the filter. Smaller pores usually require larger pumps for a given flow rate.

After the promulgation of the $\text{PM}_{2.5}$ (particulate matter with aerodynamic diameter < 2.5 μm) NAAQS [68], 37 and 47 mm-diameter PTFE thin-film membrane filters stretched across a support ring (e.g., polymethylpentene [PMP] and FEP with polypropylene) have become the most commonly used substrates for gravimetric analysis [31,69]. These ringed-membrane filters are used in U.S. $\text{PM}_{2.5}$ compliance and speciation networks and for elemental analysis [70] in addition to mass concentration. Nylon membrane filters with upstream denuders are used in a separate channel for sampling and analysis [71] of major ionic species, such as sulfate (SO_4^{2-}), nitrate (NO_3^-), and ammonium (NH_4^+), in speciation networks [72] as they minimize the loss of semivolatile ammonium nitrate (NH_4NO_3) during and after sampling [73]. Pure quartz-fiber filters are used to measure carbonaceous aerosols [74–76] because their composition is carbon-free.

Capillary pore filters consist of polycarbonate with polyester. They are manufactured from a polycarbonate sheet in contact with uranium (U) in a nuclear reactor. The neutron flux from U-235 fission creates uniform holes in the plastic [66] that are perpendicular to the filter surface. These holes are acid-etched for different durations to obtain a wide variety of pore sizes. During the early 1980s, large-pore (~8 μm) Nuclepore polycarbonate membrane filters were used as size-selective inlets in the stacked filter unit (SFU) [77] for sampling PM of different size fractions in the U.S. National Parks; smaller particles passed through the pores

while larger particles remained on the filter surface. Polycarbonate membrane filters have low porosity (5%–10%) with adequate collection efficiency, depending on pore size. These filters are commonly used for bioaerosol sampling and image processing [17,18,20]. They are especially amenable to automated single particle analyses for particle size, shape, and color using scanning electron microscopy (SEM) [78], transmission electron microscopy (TEM), and/or optical microscopy, as pattern recognition techniques can separate the pores from the particles. These filters rapidly acquire static charges [79–81] that must be neutralized prior to gravimetric analysis.

Fibrous filters are composed of a mat or weave of randomly oriented individual fibers. They are thicker than membrane filters (~200–500 μm) with various porosities (60%–90%) and sampling efficiency. Glass- and quartz-fiber filters were commonly used for high-volume air sampling during the 1960s–1990s [27]. These filters are suitable for limited elemental [70,82,83], ionic [71,84], and carbon [85–89] speciation, depending on the analysis method and filter purity. Mixed quartz-fiber filters consist of quartz fibers with ~5% borosilicate that has a lower (~500 $^{\circ}\text{C}$) melting point, causing measurement uncertainties for carbon analysis and contributing some trace elements to the blank. Pure quartz-fiber filters have high temperature resistance, but they are brittle and portions of their edges may become attached to the filter holder or flake off, thereby negatively biasing mass measurements. Ultrapure quartz-fiber filters are used in the U.S. long-term Interagency Monitoring of Protected Visual Environments (IMPROVE) network and Chemical Speciation Network (CSN) for carbon analysis [74,76,85]. These filters adsorb organic vapors [90,91], and field blanks or backup filters are needed to correct for this organic carbon bias.

Cellulose-fiber filters are made of a thick layer of paper fibers. As this material is carbon-based, these filters are inappropriate for carbon analysis. Cellulose-fiber filters have low and variable filtration efficiency (~70%) [92] and they absorb water, which can cause filter-weighing biases unless the balance environment maintains a constant relative humidity (RH) [93]. These filters are best used for absorbing gases after being impregnated with acid or base solutions. The impregnated filters can be placed behind a Teflon-membrane or quartz-fiber filter to capture precursor gases, such as sulfur dioxide (SO_2), nitrogen dioxide (NO_2), nitric acid (HNO_3), and ammonia (NH_3) [94,95]. Citric acid, oxalic acid, and phosphoric acid have been used for sampling of NH_3 , while sodium carbonate and potassium carbonate have been used for collecting SO_2 and organic acids [96]. Triethanolamine (TEA) has been used as an absorbent for collecting NO_2 , peroxyacetyl nitrate (PAN), organic nitrates, and SO_2 .

Most filters in Table S1 list a pore size. While the capillary pore filters have well-defined pore diameters, membrane and fibrous filters do not have simple pore structures. For these filters, the pore size refers to an “equivalent pore diameter,” which is determined by a bubble-point test [97]. For the same type of filter, filtration efficiency and pressure drop vary systematically with pore size. However, filtration characteristics for the same pore sizes of different filter types are not expected to be the same [98].

Few manufacturers listed in Table S1 provide filter ash contents. As noted, MSHA [53] method P-7 ashes PVC filters prior to FTIR analysis, so this is an important specification.

Ash content is expressed as a percentage of the original filter mass remaining after the filter is baked at temperatures $> 500\text{ }^{\circ}\text{C}$ (typically $800\text{ }^{\circ}\text{C}$). Low ash contents are desired when heat-resistant minerals such as quartz are to be isolated from the collection substrate. Carbon-containing filters have low ash contents as the carbon is easily combustible, while the glass- and quartz-fiber filters have high ash contents. Lippmann [66] summarized ash contents for a variety of filters, reporting values on the order of $<0.001\%$ for MCE and 0.01% each for cellulose-fiber and polycarbonate filters. Ash contents are not reported for Teflon-, nylon-, cellulose acetate-, cellulose nitrate-, PVC-, or silver membrane- filters. The glass- and quartz-fiber filters have ash contents $> 95\%$, as expected owing to their mineral compositions.

3.2. Filter Collection Efficiency

Except for selective filtration using large-pore polycarbonate filters, air sampling filters should capture more than 99% of suspended particles, regardless of particle size or flow rate [66]. Membrane filters have been used in air sampling for over 60 years [99–101]. Lower porosities and pore sizes generally result in higher sampling efficiency but increases in flow resistance. The filtration process includes a variety of collection mechanisms that alter the filter collection efficiency for various particle sizes under different sampling face velocities, which are assumed to reflect the particle velocities [22]. Decreases in particle size enhance particle collection by Brownian motion, whereas increases in particle size lead to an increase in filtration by interception, inertial impaction, and gravitational settling [12,21], as illustrated in Figure 1. Particle collection efficiency is usually at a minimum for particles with $\sim 0.3\text{ }\mu\text{m}$ aerodynamic diameters [102–104], and many filtration test methods limit themselves to this size of test aerosol.

Diffusion is the primary mechanism for collecting ultrafine particles ($<0.1\text{ }\mu\text{m}$), with higher efficiency for smaller particles. Impaction, interception, and gravitational settling are main mechanisms for collecting larger ($\geq 0.5\text{ }\mu\text{m}$) particles, with higher efficiency for larger particles. The combined effects in Figure 1 show that both small and larger particles have high filtration efficiency, while an intermediate size range (typically $0.05\text{--}0.5\text{ }\mu\text{m}$) has lower efficiency, where none of the mechanisms are most effective. Filtration performance has commonly been determined in the laboratory by generating and measuring the number concentrations of monodisperse particles or the size distributions of polydisperse solid particles (e.g., sodium chloride [NaCl]) or plasticizer (e.g., $0.3\text{ }\mu\text{m}$ dioctyl phthalate [DOP] droplet) before and after the filter for selected face velocities [75].

Liu et al. [22] reports collection efficiency for 76 filters of various pore sizes tested with monodisperse particles (i.e., 0.035 , 0.01 , 0.30 , and $1.0\text{ }\mu\text{m}$ diameters) at various pressure drops. Values relevant to CMDPSU and PDM 3700 sampling are summarized in Table 2. Filter permeability is characterized by the face velocity measured at a pressure drop of 1.3 kPa across the filter.

The CMDPSU uses a $5\text{ }\mu\text{m}$ pore PVC filter, and Table 2 shows good collection efficiency ($96.7\%\text{--}99.9\%$) for Millipore (www.emdmillipore.com, accessed on 1 October 2022) filters but with lower collection efficiency ($49\%\text{--}98.8\%$) with the more permeable Metrical (us.vwr.com, accessed on 1 October 2022) $5\text{ }\mu\text{m}$ pore PVC filters. The Metrical $0.8\text{ }\mu\text{m}$

pore PVC filter shows higher collection efficiency (99.96%–>99.99%). Different Teflon-membrane filter types (e.g., Gelman, Ghia, and Zefluor) have high collection efficiency (85%–99.99% for 2–5 μm pore sizes). Collection efficiency for silver membrane filters increases with decreasing pore size. For 5 μm pore filters, efficiency is as low as 25% for smaller particles, while 0.45 μm pore filters have efficiency >93.6% for all tested particle sizes.

The two lots of Teflon-coated glass-fiber filters (TX40HI20 used in the PDM3700) have collection efficiency of 92.6%–99.6% and 98.9%–99.99%, higher than the Pallflex 2500QAO quartz-fiber filters (84%–99.99%). Low collection efficiency was found for Whatman 41 (a commonly used cellulose-fiber filter) (43%–99.5%) and Whatman 40 (77%–99.99%). Filtration efficiency for glass-fiber filters is high (98.5%–99.99%). Kim et al. [105] noted that filtration efficiency for glass-fiber filters can be affected by particle charge, but independently of RH. As noted for polycarbonate, filter materials may acquire electrostatic charges that bias the mass determination unless adequately neutralized. Modern weighing facilities use ionizing blowers or polonium-210 charge neutralizers to eliminate this bias [31].

While filter efficiency determines the fraction of sampled particles that are retained on the collection media, filter penetration denotes the fraction that passes through the filter (i.e., 100% minus collection efficiency). Zíková et al. [106] reported large variations in size-dependent penetration for ~20–300 nm diameter particles among the five filter types. Figure 2 shows the lowest penetration (0.001%–0.1%) for glass fiber, ~0.1% for MCE and quartz fiber, ~10% for PTFE Teflon filters, and the highest penetration for 0.4 and 8 μm polycarbonate filters. The percentage penetration maximum varied by four orders of magnitude among different filter media with most penetrable particle sizes in the 20–86 nm range, mostly <50 nm. As shown in Figure S5, particle penetrations vary among filter types, but to a much lesser degree for different batches of the same type of filter.

Figure 3 [107] compares small-particle (<300 nm) collection efficiencies for five 5 μm pore membrane filters. The highest collection efficiency was found for MCE filters with rapid and variable efficiency decreases with increasing particle size for the polycarbonate and silver membrane filters. There is a ~5%–10% reduction in PTFE collection efficiency for 20–200 nm particles. Table 3 [107] shows descriptive statistics from testing a range of pore sizes (0.4–5 μm) and flow rates (1.7–11.2 L/min) including four of the five 5 μm pore filters in Figure 3. For the range of filter pore sizes 0.45–5 μm , median collection efficiency integrated over the 10.4 to 412 nm size range of the NaCl test aerosol was lowest (96%) for silver membrane but comparable among the PTFE, PVC, and MCE membrane filters (99.7%–99.9%). Millipore 5 μm pore polycarbonate at 3.11 cm/s face velocity exhibited the lowest collection efficiency (22.48%) with large variations ($85.3 \pm 22.2\%$), consistent with that for 5 μm Nuclepore filters, with efficiency as low as 6%, reported in Table 2 [22].

Large collection efficiency variations were also found for silver membrane filters with an average and standard deviation of $86.5\% \pm 20.3\%$ for a mixture of 0.45, 0.8, 1.2 and 5 μm pore sizes. The lowest collection efficiency (42.1%) was found for the 5 μm -pore Sterlitech (www.sterlitech.com, accessed on 1 October 2022) silver membrane filter tested at a face

velocity of 8.06 cm/s, consistent with those reported by Liu et al. [22] with a low 25% collection efficiency for a 5 µm Flotronics filter.

Individual tests of 5 µm-pore PVC from three manufacturers (i.e., SKC, Gelman, and Millipore) yielded 92.98%–99.95% collection efficiency [107]. The SKC (www.skinc.com, accessed on 1 October 2022) 5 µm PVC filter tested at a face velocity of 20.5 cm/s showed the lowest collection efficiency (92.98%) among the three vendors. This is comparable to the >96.7% for Millipore but much higher than the >49% for Metricel shown in Table 2. Farcas et al. [52] concluded that a 5 µm PVC filter is a suitable replacement for the DM-450 filter used for coal dust quartz determination by FTIR using the MSHA P-7 and NIOSH 7603 methods, but they noted that further experiments are needed to confirm the high collection efficiency of small particles. Overall, Soo et al. [107] confirmed the high collection efficiency (>99.7%) of PTFE, PVC, and MCE filters with higher variations found for polycarbonate and silver membrane filters for experimental particles smaller than their pore sizes and for different flow rates.

Collection efficiency and pressure drop vary with filter pore sizes [98,108]. PDMs are designed for pressure drops up to 16.7 kPa (125 mmHg) to accommodate high dust loadings [29]. For the same filter type, collection efficiency and pressure drop increase with decreasing pore size. A power-law relationship was found between the pressure drop and flow rate with higher pressure drops for small pores. Figure S6 shows large variations in initial pressure drop from 0.137 to 2.5 kPa as pore sizes decrease from 5 µm to 0.45 µm at 1.7 L/min [107]. Large increases in pressure drop were also found as flow rates increased from 1.7 to 11.2 L/min.

Collection efficiency and pressure drop change with particle loading [109]. Due to formation of a dust cake on the filter surface, the filtration efficiency will initially increase with added dust loading, especially for capillary pore filters that rely on surface interception [110]. Soo et al. [107] found collection efficiency increases for silver membrane and polycarbonate filters with increasing sample durations (sampling after 270 and 540 s) while collection efficiency for PTFE, PVC, and MCE did not show noticeable differences in particle loadings. Longer sample durations in heavily polluted environments, however, may result in filter clogging and increased pressure drop, leading to insufficient airflow through the filter [26].

The collected particles should not clog filters over a specified sampling period (e.g., 24 h or an 8 or 12 h shift). Table 1 shows that the PDM3700 can collect particle mass from 0.1 to 4 mg. Membrane filters used in samplers with low- and medium-volume inlets generally have higher flow resistance and lower loading capacity than fiber filters [13]. Lower resistance and higher capacity can be attained by increasing the filter exposed area or pore size and reducing the filter thickness. Lower flow resistance is often gained at the expense of decreased filtration efficiency.

3.3. Potential Environmental Artifacts

High temperature and moisture environments in underground mines may result in sampling artifacts that bias mass and chemical measurements. Filters should be chemically inert and

retain their porosity and structure during sampling. However, the physical and chemical properties of filters listed in Table S1 may lead to: (1) adsorption and desorption of gases in the sampled airstream; (2) evaporation of volatile and semivolatile materials; (3) reactions with the water vapor; and (4) particle losses due to lack of adhesion to the filter surface during sampling and handling.

Gases adsorbed by the filter material result in positive biases to mass and chemical concentrations. Depending on face velocities, some particles change to gases, or volatilize, when temperatures, RH, and precursor gas concentrations change during field sampling, filter handling, transport, and storage. Volatilization causes a negative bias to mass and chemical composition and is more dependent on environmental variables than on the filter composition.

It is well known that adsorption of SO₂ and nitrogen oxides (NO_x) on borosilicate glass fibers results in positive sulfate and nitrate artifacts [111–115]. This was recently confirmed in laboratory and field tests by Gilbert et al. [116]. Some studies use quartz-fiber filters to collect samples from SO₂-containing sampling streams. Ambient RH and concentrations of NH₃ and HNO₃ gases affect the gas–particle equilibrium of NH₄NO₃, but temperature is the most important variable [117]. Evaporation of NH₄NO₃ during warm-season sampling has been documented [73,118,119]. Febo et al. [120] found that cellulose-fiber filters retain both HNO₃ and nitrate (positive artifact), in contrast to Teflon-membrane filters (negative artifact). Keck and Wittmaack [121,122] reported adequate sampling of NH₄NO₃ and ammonium chloride (NH₄Cl) with MCE membrane filters that retained large fractions of particle evaporation. The chemical stability of the MCE needs to be further examined. Instead of using a backup filter to evaluate particle evaporation [96], Keck and Wittmaack [123] measured semivolatile inorganic PM using a denuded cellulose-fiber filter.

Adsorption of SO₂, NO_x, water vapor, and diesel exhaust fumes are potential interferents for mass and chemical measurements in underground coal mines. The use of nonacidic and nonalkaline filters (e.g., Teflon membrane or quartz fiber) largely eliminates these artifacts. The adsorption of organic gases by quartz-fiber filters is still an interferent for mass and organic carbon concentrations [90,91,124]. Organic gas adsorption (positive bias) may counteract organic particle volatilization (negative bias). However, as sample durations increase, the proportion of the adsorption bias decreases because the adsorbed gases reach equilibrium with the collected particles and the filter may become saturated. The composition of the organic gases and particles in the sampled air may affect the magnitude of the artifact.

During sampling, water vapor can also adsorb to or desorb from the filter. Allen et al. [125] reported large water uptake by glass-fiber filters with less RH influence for Teflon-coated glass-fiber filters in woodsmoke emissions testing. Comparison of gravimetric mass between uncontrolled (24–28.5 °C and 40%–65% RH) and controlled (22–26 °C and 39.5%–41% RH) environments by Tsai et al. [126] found a higher CV of 0.27% for MCE (Table S2). Particle mass measured on Teflon-membrane and PVC filters is more stable than measurements on glass-fiber filters, irrespective of equilibration condition. CVs for Teflon-membrane filters (0.0019%–0.002%) are an order of magnitude lower than those of PVC

filters (0.016%–0.017%). The U.S. EPA regulations for $\text{PM}_{2.5}$ mass concentration require environmental equilibration of filters within temperature (21.5 ± 1.5 °C) and RH ($35\% \pm 5\%$) ranges for at least 24 h prior to gravimetric analysis [31] to minimize moisture effects.

Since the PDM vibrates the filter to determine mass concentrations, the particles must not be moved or dislodged due to the filter motion [42]. A fibrous filter is currently used with the assumption that the fibers hold the particles in place. However, particle loss due to lack of filter adhesion during vibration or filter transport has not been extensively tested for any of the filters. Under normal sampling conditions, particle bounce, in which particles deposited on a substrate are removed by collisions with incoming particles, has been observed during cascade impactor sampling. This becomes more common at higher face velocities.

MSHA method P-7 employs low temperature ashing and redeposition onto another filter before spectroscopic analysis. Because the current PDM filter holder cannot be disassembled, the entire assembly needs to be ashed. Tuchman et al. [29] shows that the polypropylene filter assembly containing titanium dioxide produces a strong and broad IR absorption spectrum for wave numbers of $450\text{--}850\text{ cm}^{-1}$ that interferes with RCS quantification. On the other hand, the ashed clear polypropylene shows much lower absorbance by FTIR analysis. Potential interference of filter holder on RCS analysis by FTIR should be considered when designing filter holders.

3.4. Inhomogeneous Sample Deposits and Filter Cassette Assembly

Filters should lie flat on top of the filter cassette to prevent leakage. Open-faced filter holders with no upstream constrictions provide homogeneous sample deposits, but in-line filter holders without a mixing zone above the filter surface cause more particles to be collected in the filter's center [127]. This is not an issue if the entire filter is digested and analyzed, but it biases results when only parts of the filter are examined. The PDM filter assembly (Figure 4) with a circular polypropylene base on top of a hollow axial stem results in an inhomogeneous particle deposit. Figure S7 shows an inhomogeneous sample deposit for $0.45\text{ }\mu\text{m}$ nylon filters used for redeposition of laboratory-generated coal dust compared to polypropylene and PVC [52].

Nondestructive spectroscopic analysis irradiates a small particle deposit area, then normalizes the concentration to the entire exposed area, so a uniform surface deposit is needed [27]. Miller et al. [60] used three different systems sampling laboratory-generated minusil (Min-U-Sil, a standard source of pure silica with a 92% α -quartz content) [128] and coal dust. Analyses of silica by FTIR were directed to nine different spots across each filter by moving the filter holder in 3 mm increments. CVs of 4%–51% were found, depending on the number of replicate analyses and the type of sampling device. Miller et al. [60] found that samples collected with three-piece cassettes using a single-spot FTIR analysis produced adequate CVs of ~15%. For the CMDPSU, the 37 mm 4-piece black polypropylene conductive housing (Figure S2) features additional spacer rings to increase the distance between the cyclone inlet and filter, resulting in a more homogeneous particle deposit. A redesign of 13 mm filter cassette assembly for the PDM needs to employ an open-face sampling concept to ensure homogeneous particle deposits.

3.5. Cost and Availability

Some of the filter media documented during the 1980s–1990s are no longer commercially available. Consistency in filter quantity and manufacturer production is needed for long-term coal mine sampling. Table S1 documents a variety of filter disks from 13 to 142 mm in diameter that accommodate different filter cassettes and sampling systems, with the most common being 37 and 47 mm. Large rectangular 20.3 cm × 25.4 cm quartz- and glass-fiber filters have been used for high-volume PM₁₀ sampling [129].

Filter quantities vary between 25 and 100 per box with various pore sizes. For comparison, Table S1 separates costs into low (US\$300 per 100 filters), medium (\$301 to \$500 per 100 filters), and high (>\$500 per 100 filters) categories. Filter costs have increased over time with variations by filter type and pore size. The highest costs are for silver membrane filters, ranging from \$7 to \$20 per filter, and PTFE Teflon-membrane filters, ranging from ~\$5 to \$12 per filter. Nylon membrane filter costs are low (~\$2 to \$3 per filter). The lowest cost is found for polycarbonate, cellulose-fiber, and glass-fiber filters in the range of \$1 to \$1.50 per filter.

To accommodate PDM3700 sampling, 13 mm-diameter filters are only available from three vendors for Teflon, silver membrane, and polycarbonate filters with various pore sizes. These include: (1) Pall Corporation ([us.VWR.com](https://us.vwr.com), accessed on 1 October 2022) TF filter (0.45 µm); (2) SKC (www.skcin.com, accessed on 1 October 2022) PTFE (5 µm); (3) SKC silver membrane (0.2, 0.45, 0.8, 1, 2, 3, and 5 µm); (4) Sterlitech Corporation (www.sterlitech.com, accessed on 1 October 2022) silver membrane (0.2, 0.45, 0.8, 1.2, 3, and 5 µm); and (5) Sterlitech polycarbonate (3 µm). These 13 mm filters are ~\$3 to \$4 per filter. Some filters without commercially available 13 mm sizes can be punched from larger sizes to produce 13 mm disks.

4. Membrane Filters for Respirable Crystalline Silica (RCS) Quantification

Silica, or silicon dioxide (SiO₂), is present in both crystalline and amorphous (polymorphic) forms. Crystalline silica is a mineral commonly found in the forms of sand, soil, stone, brick, and concrete, whereas RCS is mechanically generated when cutting, clipping, grinding, drilling, sanding, sawing, and crusting rocks and stones that present health risks in workplaces.

Table S3 shows three major infrared bands of crystalline silica, including α-quartz, cristobalite, and tridymite. Of these, quartz is a naturally occurring crystalline mineral that consists primarily of silica with some impurities in α-quartz, a low-temperature phase quartz that is the most abundant and thermodynamically stable form of crystalline silica. Cristobalite and tridymite are rare: they can occur in igneous rocks, but generally in very small amounts. Tridymite can also occur in highly metamorphosed impure limestones and arkoses, and there can be trace amounts of cristobalite in soils [130]. Other crystalline silica minerals include keatite, coesite, stishovite, and moganite. Coesite and stishovite are rare in nature. Keatite is a synthetic form not normally found in nature [130]. Published literature uses the terms silica, quartz, α-quartz, crystalline silica, and RCS interchangeably. OSHA established standards to protect workers' exposure to RCS (29 CFR 1926.1153).

The following sections discuss spectroscopic methods, potential interference, and detection limits for RCS analysis by FTIR and Raman spectrometry.

4.1. Spectroscopic Analysis Methods for Coal Mine Dust

The most common analytical methods for determining RCS in dust samples are infrared (IR) spectrometry and X-ray diffraction (XRD) [59,131–134]. Currently, IR is used for RCS analysis in coal dust [53,54], while XRD is used for RCS analysis in metal and nonmetal mining dust [54,135]. XRD determines the structure and composition of crystalline substances in dust and rocks semiquantitatively. It is less affected by interference by distinguishing different types of crystalline silica (e.g., quartz, cristobalite, and tridymite). However, XRD analysis is labor intensive and not commonly applied to filter samples compared to the IR or FTIR methods. IR spectrometry is less costly with good precision [57,61], but it is also a less specific chemical characterization technique. Comparisons between XRD and IR spectrometry for the analysis of α -quartz show agreement on average (within 2%) with large variabilities (with differences up to 41.3%) in individual sample comparisons [132].

4.1.1. FTIR Spectrometry—FTIR spectrometry has been applied to explore functional groups (e.g., alcohols, amines, carboxylic acids, and ketones) and to examine chemical structures of coal mine dust in the workplace. It is nondestructive and usually applied to membrane filters. FTIR spectrometry identifies molecular vibrations and the resulting absorption by illuminating the sample with multiple-wavelength radiation from an IR-emitting source [64,136]. The concentration is proportional to the absorbance as determined by Beer's law [137]. The IR is absorbed when incident radiation is equivalent to the energy of a particular molecular vibration. The transmitted radiation is detected and processed based on a "Fourier-transform algorithm" to generate absorbance spectra. Similarly to other spectroscopic techniques, FTIR absorbance can be estimated by: (1) peak height after subtraction of estimated filter background spectra; (2) peak height after spectral blank subtraction; and (3) integrated absorbance after blank corrections [55].

Both MSHA [53] P-7 and NIOSH [54] 7603 methods use a low-temperature radio-frequency ashers to destroy the organic matrix of the filter sample, but kaolinite remains structurally unaltered and can interfere with the silica measurement [138]. MSHA method P-7 integrates the absorbance peak area with a baseline between 770 and 815 cm^{-1} for quartz (plus kaolinite) and between 900 and 930 cm^{-1} for kaolinite, whereas NIOSH method 7603 uses a peak height at 800 cm^{-1} with baseline between 670 and 820 cm^{-1} for quartz (plus kaolinite) and a peak height at 915 cm^{-1} with baseline between 860 and 960 cm^{-1} for kaolinite. The absorption peak centered around 800 cm^{-1} is corrected for the kaolinite interference using the absorption centered around 915 cm^{-1} . For samples containing >20% of calcite or graphite, a muffle furnace heated at 600 to 800 $^{\circ}\text{C}$ can also be used in NIOSH method 7603. Ashing at these high temperatures allows the conversion of kaolinite to amorphous meta-kaolin that does not interfere with the RCS measurement [138,139]; the kaolinite correction is not needed under this condition. A consistent approach to integrate FTIR absorption peaks is needed to determine deviations among these methods.

FTIR absorption bands for α -quartz at 780 and 799 cm^{-1} shown in Table S3 are similar to the 778 and 797 cm^{-1} identified by Abbasi et al. [131]. Secondary absorbance of cristobalite at 796 cm^{-1} may overlap with wave numbers of 798 and 799 cm^{-1} for quartz. However, minerals that interfere with RCS analysis are typically not found in RCMD, with the exception of kaolinite [138].

Table 4 summarizes infrared absorbance spectra for five membrane filter types [140]. The lowest absorbance was 0.057 at 695 cm^{-1} for polycarbonate filters. For PTFE, absorbances were 0.074 at 779 cm^{-1} and 0.063 at 798 cm^{-1} . DM-450 filter absorbances were 0.078 at 779 cm^{-1} and 0.074 at 798 cm^{-1} . Lorberau [141] reported that PVC-copolymer filters (DM-450 and DM-800, Gelman Sciences, Ann Arbor, MI, USA) had the lowest absorbance (1.3%–2.4%) and lowest associated standard deviations at wave numbers of 779 cm^{-1} and 798 cm^{-1} for fifteen 25 mm filters that were tested. The DM-450 PVC-copolymer filters have been used for redepositing in both the MSHA P-7 and NIOSH 7603 methods. However, this type of filter is no longer commercially available, and alternative filters are being sought [52]. Absorption for four types of blank membrane filters (i.e., polypropylene, nylon, DM-450, and PVC) from 12 replicate FTIR measurements by Farcas et al. [52] show smooth absorbances with the least sample-to-sample variations for DM-450 filters while PVC filters show the largest variations among measurements. Multiple absorption peaks are found for both polypropylene and nylon blank filters.

Interlaboratory comparisons between NIOSH (Spokane, WA, USA) and DRI (Reno, NV, USA) laboratories (Figure 5) also show multiple peaks for polypropylene filters. Nevertheless, good reproducibility in the peaks and valleys of the FTIR spectra was found between the two laboratories. The SKC 5 μm PVC filters (Figure 5a) show the lowest and most stable baseline for RCS determination. The spectrum for SKC 0.8 μm MCE blank filters in Figure 5b shows a sloping tail between the two black vertical lines and a large peak close to 850 cm^{-1} that might interfere with RCS quantification.

Figure S8 shows that the Teflon-coated glass-fiber filter (TX40HI20) used in PDM3700 has high background (~ 2 absorbance units) in the RCS range (767–816 cm^{-1}). Griffiths and de Haseth [136] recommend <0.7 absorbance units for FTIR. Therefore, the TX40HI20 may not be suitable for RCS quantification by FTIR. Additional tests are needed to verify the relative differences between the RCS signal magnitude and blank filter absorbance.

4.1.2. Raman Spectroscopy—Raman spectroscopy uses scattered light to determine the crystallinity and molecular interaction between light and chemical bonds by a vibrational spectrum. It is sensitive to both chemical and morphological changes in samples and capable of selectively identifying specific vibrational modes in organic and inorganic substances between polymorphs. Raman spectroscopy can effectively differentiate between polymorphs and microcrystalline silica [142]. It has been used to determine trace impurities in mineral samples with sharp bands for crystalline minerals and broad bands for amorphous phases or fluorescence [16,19]. Since the spatial resolution of FTIR spectroscopy is limited to 10–20 μm [143], Raman spectrometry complements FTIR analysis by providing better resolution down to 1 μm with micrometer-scale characterization, thereby improving the detection limit.

Stacey et al. [144] reported a quartz peak at 464 cm^{-1} and a cristobalite peak at 410 cm^{-1} when the sample was excited with a near-infrared laser (785 nm). For coal dust analysis, the two major bands are in the regions of $1355\text{--}1380\text{ cm}^{-1}$ (D-band) and $1557\text{--}1620\text{ cm}^{-1}$ (G [graphic] band) [145]. Shin and Chung [146] advocate the application of wide-area illumination to enhance reproducibility and the signal-to-noise ratio of spectra for optimal Raman spectroscopic analysis.

Preliminary laboratory tests of polycarbonate and silver membrane filters for RCS were conducted using an iRaman Plus spectrometer (B&W Tek) with 532 nm laser excitation. Figure 6a shows that polycarbonate and silver membrane filters do not cause spectral interference with the quartz peak at 465 cm^{-1} . The quartz peak intensity on the silver filter is twice that for the polycarbonate filter with a given mass loading ($6\text{ }\mu\text{g RCS}$), which is attributed to the enhanced Raman scattering from porous silver membrane filters [147]. Figure 6b shows that the iRaman Plus (532 nm) is capable of detecting quartz (characteristic band at 465 cm^{-1}) and black coal (characteristic bands at 1356 cm^{-1} and 1576 cm^{-1}) when a small amount of quartz is added to coal and that the black coal carbon does not interfere with quartz. However, a separate test with a low-ranked brown coal sample with rich volatile organics generated high fluorescence signals that saturated the detector. As the excitation laser wavelength of a Raman spectrometer affects excitation efficiency, Raman scattering intensity, fluorescence, and heating, further research is needed to select the best laser wavelength for RCS quantification in coal matrix. Limited research has been done with Raman spectroscopy [142,148–150] for coal mines, and more tests are needed on membrane filters to better understand the blank filter absorbance and potential spectral interference.

4.2. Detection Limits

Filters should have low blank levels for the targeted chemical species ($<1\text{ }\mu\text{g}$ per 47 mm filter). Background contents of major and trace elements and isotopic composition for the PTFE, nylon, polycarbonate, and glass-fiber filters are reported by Lee et al. [151]. Trace-element concentrations are the lowest in the PTFE followed by nylon.

Prior to sampling, at least 2% of each batch of 100 filters from each manufacturer should be light-inspected for pinholes, creases, discoloration, or other defects and analyzed for all species to verify the background concentrations. Quartz-fiber filters adsorb organic vapors requiring pre-firing at $900\text{ }^{\circ}\text{C}$ for 4 h prior to “acceptance testing.” Nylon and MCE membrane filters absorb HNO_3 over time that need to be acceptance tested and/or washed with distilled deionized water prior to sampling [96].

The nominal RCS limits of detection (LODs) are 4 and $10\text{ }\mu\text{g}$ per 37 mm PVC sample for MSHA P-7 and NIOSH 7603 methods, respectively [53,54]. Linear calibration curves between 10 and $1000\text{ }\mu\text{g}$ for FTIR can be obtained with standard reference materials (SRM 1878a for quartz and SRM 1879a for cristobalite [152]). Lorberau [141] reported LOD levels (triple the standard deviation of the average blanks) for DM-450 and DM-800 of $5.3\text{ }\mu\text{g}$ at 779 cm^{-1} and $2.9\text{ }\mu\text{g}$ at 798 cm^{-1} per 25 mm filter with 1.3%–2.4% variations within and between the two filter types. These levels are higher than the $0.59\text{--}0.99\text{ }\mu\text{g}$ per 47 mm filter in Table 5, reported by Farcas et al. [52]. This may be due to the fact that filters are prewashed in isopropyl alcohol prior to FTIR analysis [52]. Laboratory tests should be

conducted to determine if rinsing or prewashing in isopropyl alcohol is needed prior to field sampling.

Both Zefon 5 μm PVC and Sterlitech 0.45 μm polypropylene reported lower LODs (0.52–0.72 μg per 47 mm filter) by MSHA [53] P-7 than by NIOSH [54] 7603 (1.2–1.5 $\mu\text{g}/\text{filter}$). Note that LODs varied by around threefold for 5 μm PVC filters between the MSHA P-7 and NIOSH 7603 methods. However, the trend is reversed for DM-450 and Nylon filters, with NIOSH 7603 having lower LODs than MSHA P-7. The limit of quantification (LOQ, 10 times the standard deviation of the average blanks) in Table 5 are lower than LODs of ~ 7 $\mu\text{g}/\text{filter}$ for quartz and 20 $\mu\text{g}/\text{filter}$ for coal dust reported in Tuchman et al. [29] and 3–10 $\mu\text{g}/\text{filter}$ for quartz reported by Abbasi et al. [131] and NIOSH [153]. The LOD should account for filter variations by batches as well as measurement variations due to changes in environmental parameters and FTIR performances.

The LOD by Raman spectrometry can be improved with longer collection times for several locations on each filter. In general, detection limits for Raman spectroscopy can be one to two orders of magnitude lower than those for XRD and FTIR. For five 13 mm PVC (5 μm pore size) filters, Stacey et al. [144] reported LOD of 0.049 $\mu\text{g}/\text{filter}$ for quartz and 0.02 $\mu\text{g}/\text{filter}$ after accounting for the variability of the background scatter. For an analysis area of 100 μm^2 , Stacey et al. [144] reported an LOQ of 0.066–0.161 $\mu\text{g}/\text{filter}$ for quartz by analyzing 50 locations on 5 mm-diameter silver-membrane-filter particle deposits. The measurement precision was 10%–25% for 0.25–10 μg quartz. Using an aerosol microconcentrator to obtain 400 to 1000 μm diameter deposits on silver membrane filters, Zheng et al. [147] reported 8 to 55 ng α -quartz LODs using a portable Raman spectrometer at 465 cm^{-1} over 60 s integration times.

5. Major Findings and Recommendations

Mine safety regulations require monitoring dust and RCS concentrations in mining environments. CMDPSUs are used to collect dust samples for gravimetric and RCS quantification in offsite laboratories. However, it typically takes days or weeks before the data are available. The PDM measures dust concentrations in near real time. However, the current filter used in the PDM and the filter holder design are not suitable for RCS measurement. This study reviews characteristics of different filter media with an aim to identify filters that can be used for both dust mass concentration and chemical composition analysis, especially for RCS quantification by FTIR and Raman spectrometry.

Chemical and physical characteristics of 12 types of commercially available filter substrates are evaluated (Table S1). Among the three main classes of filters that have been commonly applied for aerosol sampling, the five fibrous filters, including the Teflon-coated glass fiber (TX40HI20) used in the PDM3700, are not further considered for mass and quartz determination owing to their low and variable collection efficiency (Table 2) and/or interference with quartz measurement. Adsorption or desorption of SO_2 and NO_2 on glass fibers [120–122] and adsorption and volatilization of organic vapors in quartz-fiber filters [96] may bias mass determination. Quartz-fiber filters have been applied to long-term

sampling in PM_{2.5} speciation networks, but they contain quartz material in the filter mat and cannot be used for RCS quantification.

Performance of capillary pore filters (i.e., polycarbonate) shows variable collection efficiency (as low as ~1%–6% for 5 µm and 8 µm pore sizes) [22]; high penetration (0.4 and 8 µm pore sizes) for particles between 10 and 100 nm [106,107]; and variable blank filter absorbance by FTIR spectrometry [52].

Different test results are reported for the six porous membrane filters. Silver membrane has variable collection efficiency (~25%–90%) and high costs per filter. For 5 µm silver membrane filters, as low as 25% collection efficiency was reported by Liu et al. [22] and as low as 42% efficiency was found by Soo et al. [107]. Pressure drops and filtration efficiency increase with decreasing pore size. Silver membrane filters are not transparent by FTIR and are not suitable for FTIR measurements using the transmission mode; further testing is needed to verify if they can be used in reflectance mode. The ability of these filters to enhance Raman scattering makes them good candidates for Raman measurement [147].

MCE membrane filters show the highest collection efficiency (98%–99.99%) [107] and low penetration (~0.1%) [106]. However, MCE filters absorb nitric acid. Keck and Wittmaack [121,122] reported adequate sampling of ammonium salts with MCE filters, as it retained a large fraction of volatilized gases. Comparing gravimetric mass among Teflon, PVC, MCE membrane filters and glass-fiber filters, Tsai et al. [126] reported the largest coefficient of variation (CV, 0.27%) for MCE (Table S2), one to two orders of magnitude higher than those of Teflon-membrane and PVC filters, an indication of potential water uptake by MCE filters. Therefore, the chemical stability of MCE filters warrants further examination. Preliminary laboratory tests show a sloping tail close to 850 cm⁻¹ that might interfere with RCS quantification by FTIR. MCE filters are made of cellulose nitrate and small amounts of cellulose acetate. It is not clear if consistent mixtures are used among the five different vendors. For example, Table S1 shows that the Thermo Scientific Nalgene MCE contains a mixture of cellulose diacetate and triacetate. More tests need to be conducted to select an MCE vendor and to ensure blank filter consistency among batches.

No collection efficiency was reported for nylon filters. Nylon filters are known to absorb nitric acid. Farcas et al. [52] reported higher RCS limit of detection (1.5–1.8 µg per 47 mm filter) in 0.45 µm nylon filters than those of 0.45 µm DM-450 or polypropylene, and 5 µm PVC filters. Blank nylon filters show multiple peaks over the RCS analysis region on FTIR. Inhomogeneous particle deposition was reported when redeposit of ashed nylon filters [52].

Little testing has been conducted on polypropylene filters. In laboratory tests of blank filters, multiple absorption peaks with large variations were found on polypropylene filters by Farcas et al. [52] as well as at NIOSH (Spokane, WA, USA) and DRI (Reno, NV, USA) laboratories that might interfere with RCS quantification.

Collection efficiency for 5 µm PVC membrane filters used in the CMDPSU are generally high, ranging 92.98%–99.95% on individual tests among filters from three vendors (i.e., SKC, Gelman, and Millipore), with the lowest (92.98%) found at 20.5 cm/s face velocity for SKC filters [107]. Very low collection efficiency (as low as 49%) was also found for the 5

µm Metrical PVC filter at a higher face velocity (>51 cm/s) by Liu et al. [22]. Collection efficiency of 5 µm PVC for PDM3700 at a high face velocity of 27.6 cm/s needs to be tested for mass and RCS determination by PDM. Farcas et al. [52] cautioned that 5 µm PVC filters may replace the DM-450 filter for coal dust quartz analysis by FTIR, but their ability to retain small particles warrants additional experiments.

Large variations in FTIR baseline absorption of blank PVC filters were reported by Farcas et al. [52]. However, preliminary laboratory testing at both NIOSH (Spokane, WA, USA) and DRI (Reno, NV, USA) laboratories showed stable and low baselines for the RCS spectrum of 767–816 cm⁻¹. Low detection limits of 0.52–1.5 µg per 47 mm 5 µm Zefon PVC filters were reported by Farcas et al. [52], lower than the LOD of ~7 µg/filter for quartz noted by Tuchman et al. [29] and 3–10 µg/filter for quartz reported in Abbasi et al. [131] and NIOSH [153].

Overall performance is best for PTFE Teflon-membrane filters, with an average of 99% ± 2% collection efficiency for the 171 tests [107]. The lowest collection efficiency of 94.76% at 3.11 cm/s was found for 5 µm Pall PTFE filters. Although earlier tests by Liu et al. [22] reported >88% collection efficiency for the 3 µm Zefluor filters and >85% for 5 µm Gelman filters, the 2 µm PTFE produced adequate collection efficiency with >94.6% (Zefluor) and >99.89% (Ghia). Lorberau [140] reported low IR absorbance for PTFE Teflo by FTIR (0.074 to 0.078 at 779 cm⁻¹ and 0.063 to 0.074 at 798 cm⁻¹), at similar magnitude to the 0.074–0.078 level found in DM-450 (a PVC-acrylic copolymer).

Two candidate filters, PTFE and PVC, warrant additional tests. Potential vendors should be contacted to confirm the availability and cost of 13 mm diameter filters to ensure long-term consistency and stability. Currently, Pall Corporation supplies 13 mm diameters of 0.45 µm and SKC supplies 5 µm PTFE membrane filters. Given the high face velocity in PDM3700, the 2 µm Teflo™ from Pall Corporation with thickness of 46 µm, most commonly applied for aerosol sampling in U.S. networks, may be considered. The challenge is to manufacture a 13 mm PTFE filter with thinner polymethylpentene (PMP) support ring to prevent leaks. It may be possible to stack two PTFE filters without a support ring. The thickness of 2 µm Zefluor (152 µm) is three times that of Teflo, which may not transmit light for FTIR analysis.

Filter media testing and selection should consider the characteristics required for mass, FTIR, and Raman measurements. For mass determination by the PDM, the filters need to have high filtration efficiency (99% for the most penetrable particle sizes) and a reasonable pressure drop (up to 16.7 kPa) to accommodate high dust loadings [29]. Additional requirements include: negligible uptake of water vapor and gaseous volatile compounds; adequate particle adhesion as a function of particle loading; sufficient particle loading capacity to form a stable particle deposit layer during sampling in wet and dusty environments; mechanical strength to withstand vibrations and pressure drops across the filter; and appropriate filter mass compatible with the tapered element. FTIR and Raman measurements require filters to be free of spectral interference. Furthermore, because the irradiated area does not completely cover the sample deposit, particles should be uniformly deposited on the filter.

For a broader application of potential chemical characterization, two out of 100 filters should be submitted for light inspection and acceptance testing of trace elements and ions. Standard operating procedures for FTIR and Raman spectroscopy need to be developed. Detection limits and blank filter absorption and potential interference by FTIR and Raman spectrometry need to be determined.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgments:

The authors thank Arthur L. Miller of NIOSH (ret) for his insightful review and recommendations. Bankole Osho and Mohammadreza Elahifard of DRI generated the original figures.

Funding:

This research was funded by National Institute for Occupational Safety and Health (NIOSH), grants 75D30119C0759 and 75D30121C11871.

Data Availability Statement:

Relevant data are contained in the cited articles and reports.

References

1. Spurny KR The history of dust and aerosol filtration. In *Advances in Aerosol Filtration*; Spurny KR, Ed.; CRC Press LLC: Boca Raton, FL, USA, 1998; pp. 3–12.
2. Davies CN; Aylward M Photoelectric measurement of coal dust stains of filter paper. *Br. J. Ind. Med.* 1949, 6, 254–258. [PubMed: 15390513]
3. Feldhaus GM Schutzmasken in vergangenen Jahrhunderten. *Gasmaske* 1929, 1, 104–114.
4. Fieldner AC; Oberfell GG; Teague MC; Lawrence JN Methods of testing gas masks and absorbents. *J. Ind. Eng. Chem.* 1919, 11, 519–540.
5. Fuchs NA *The Mechanics of Aerosols*; Pergamon Press Ltd.: New York, NY, USA, 1964.
6. Spurný K Membrane filters in aerosology. II. Filtration mechanisms in membrane filters. *Zent. Biol. Aerosolforsch.* 1965, 12, 530–545.
7. Davies CN *Air Filtration*; Academic Press: London, UK, 1973.
8. Davies CN Filtration of aerosols. *J. Aerosol Sci.* 1983, 14, 147–161.
9. Langmuir I Report on Smokes and Filters, Section I; U.S. Office of Scientific Research and Development: Washington, DC, USA, 1942.
10. Langmuir I *The Collected Works of Irving Langmuir*; Pergamon Press: London, UK, 1962.
11. Kaufman A Die Faserstoffe für Atemschutzfilter–Wirkungsweise und Verbesserungsmöglichkeiten. *VDI* 1936, 80, 593–599.
12. Jung S; Kim J Advanced Design of Fiber-Based Particulate Filters: Materials, Morphology, and Construction of Fibrous Assembly. *Polymers* 2020, 12, 1714. [PubMed: 32751674]
13. Barrett LW; Rousseau AD Aerosol loading performance of electret filter media. *Am. Ind. Hyg. Assoc. J.* 1998, 59, 532–539.
14. Lee H; Segets D; Sub S; Peukert W; Chen SC; Pui DYH Liquid filtration of nanoparticles through track-etched membrane filters under favorable and different ionic strength conditions: Experiments and modeling. *J. Membr. Sci.* 2017, 524, 682–690.
15. Cai H; Chen M; Chen Q; Du F; Liu J; Shi H Microplastic quantification affected by structure and pore size of filters. *Chemosphere* 2020, 257, 127198. [PubMed: 32512329]

16. Wright SL; Levermore JM; Kelly FJ Raman spectral imaging for the detection of inhalable microplastics in ambient particulate matter samples. *Environ. Sci. Technol.* 2019, 53, 8947–8956. [PubMed: 31293159]
17. Mainelis G Bioaerosol sampling: Classical approaches, advances, and perspectives. *Aerosol Sci. Technol.* 2020, 54, 496–519. [PubMed: 35923417]
18. Haig CW; Mackay WG; Walker JT; Williams C Bioaerosol sampling: Sampling mechanisms, bioefficiency and field studies. *J. Hosp. Infect.* 2016, 93, 242–255. [PubMed: 27112048]
19. Wigginton RK; Vikesland PJ Gold-coated polycarbonate membrane filter for pathogen concentration and SERS-based detection. *Analyst* 2010, 135, 1320–1326. [PubMed: 20498881]
20. Kildeso J; Nielsen BH Exposure assessment of airborne microorganisms by fluorescence microscopy and image processing. *Ann. Occup. Hyg.* 1997, 41, 201–216. [PubMed: 9155240]
21. Zhu M; Han J; Wang F; Shao W; Xiong R; Zhang Q; Pan H; Yang Y; Samal SK; Zhang F; et al. Electrospun nanofibers membranes for effective air filtration. *Macromol. Mater. Eng.* 2016, 302, 1600353.
22. Liu BYH; Pui DYH; Rubow KL Characteristics of air sampling filter media. In *Aerosols in the Mining and Industrial Work Environments*; Ann Arbor Science: Ann Arbor, MI, USA, 1983; pp. 989–1037.
23. Li XW; Hoff SJ; Bundy DS; Harmon J; Xin H; Zhu J Biofilter—A malodor control technology for livestock industry. *J. Environ. Sci. Health-Part A-Toxics/Haz. Subst. Env. Eng.* 1996, 31, 2275–2285.
24. Spurny KR (Ed.) *Advances in Aerosol Filtration*; CRC Press LLC: Boca Raton, FL, USA, 1998.
25. Lippmann M Filters and filter holders. In *Air Sampling Instruments for Evaluation of Atmospheric Contaminants*, 9th ed.; Cohen BS, McCammon CS Jr., Eds.; American Conference of Governmental Industrial Hygienists (ACGIH): Cincinnati, OH, USA, 2001; pp. 281–314.
26. Raynor PC; Leith D; Lee KW; Mukund R Sampling and analysis using filters. In *Aerosol Measurement: Principles, Techniques and Applications*, 3rd ed.; Kulkarni P, Baron PA, Willeke K, Eds.; Wiley: Hoboken, NJ, USA, 2011; pp. 107–128.
27. Chow JC Critical review: Measurement methods to determine compliance with ambient air quality standards for suspended particles. *J. Air Waste Manag. Assoc.* 1995, 45, 320–382. [PubMed: 7773805]
28. Tuchman DP Research toward direct analysis of quartz dust on filters using FTIR spectroscopy. *Bur. Mines Inf. Circ.* 1992, 9309, 1–17.
29. Tuchman DP; Volkwein JC; Vinson RP Implementing infrared determination of quartz particulates on novel filters for a prototype dust monitor. *J. Environ. Monit.* 2008, 10, 671–678. [PubMed: 18449405]
30. Watson JG; Chow JC Ambient aerosol sampling. In *Aerosol Measurement: Principles, Techniques and Applications*, 3rd ed.; Kulkarni P, Baron PA, Willeke K, Eds.; Wiley: Hoboken, NJ, USA, 2011; pp. 591–614.
31. Watson JG; Tropp RJ; Kohl SD; Wang XL; Chow JC Filter processing and gravimetric analysis for suspended particulate matter samples. *Aerosol Sci. Eng.* 2017, 1, 193–205.
32. Lilienfeld P; Dulchinos J Portable instantaneous mass monitoring for coal mine dust. *Am. Ind. Hyg. Assoc. J.* 1972, 33, 136–145. [PubMed: 5074670]
33. Patashnick H On-line, real-time instrumentation for diesel particulate testing. *Diesel Prog. N. Am.* 1987, 53, 43–44.
34. Sorensen CM; Gebhart J; O'Hern TJ; Rader DJ Optical measurement techniques: Fundamentals and applications. In *Aerosol Measurement: Principles, Techniques and Applications*, 3rd ed.; Kulkarni P, Baron PA, Willeke K, Eds.; Wiley: Hoboken, NJ, USA, 2011; pp. 269–312.
35. Chow JC; Watson JG; Park K; Lowenthal DH; Robinson NF; Magliano KL Comparison of particle light scattering and PM_{2.5} mass in central California. *J. Air Waste Manag. Assoc.* 2006, 56, 398–410. [PubMed: 16681205]
36. Wang XL; Chancellor G; Evenstad J; Farnsworth JE; Hase A; Olson GM; Sreenath A; Agarwal JK A novel optical instrument for estimating size segregated aerosol mass concentration in real time. *Aerosol Sci. Technol.* 2009, 43, 939–950.

37. Kleinman MT; Head SJ; Morris RE; Stevenson ED; Altshuler SL; Chow JC; Watson JG; Hidy GM; Mueller PK Air quality measurements: From rubber bands to tapping the rainbow: Critical review discussion. *J. Air Waste Manag. Assoc.* 2017, 67, 1159–1168. [PubMed: 29028427]
38. Hidy GM; Mueller PK; Altshuler SL; Chow JC; Watson JG Critical review: Air quality measurements—From rubber bands to tapping the rainbow. *J. Air Waste Manag. Assoc.* 2017, 67, 637–668. [PubMed: 28333580]
39. Mine Safety and Health Administration (MSHA). 30 CFR Part 74: Coal mine dust sampling devices. *Fed. Regist.* 2010, 70, 17523–17529.
40. Page SJ; Volkwein JC A revised conversion factor relating respirable dust concentrations measured by 10 mm Dorr-Oliver nylon cyclones operated at 1.7 and 2.0 L min⁻¹. *J. Environ. Monit.* 2009, 11, 684–689. [PubMed: 19280048]
41. Patashnick H; Hemenway CL Oscillating fiber microbalance. *Rev. Sci. Instrum.* 1969, 40, 1008–1011.
42. Patashnick H; Rupprecht EG Continuous PM10 measurements using the tapered element oscillating microbalance. *J. Air Waste Manag. Assoc.* 1991, 41, 1079–1083.
43. Volkwein JC; Tuchman DP; Vinson RP Performance of a prototype personal dust monitor for coalmine use. In *Mine Ventilation*; AA Balkema: Lisse, The Netherlands, 2002.
44. Volkwein JC; Vinson RP; McWilliams LJ; Tuchman DP; Mischler SE Performance of a New Personal Respirable Dust Monitor for Mine Use; National Institute of Occupational Safety and Health: Pittsburgh, PA, USA, 2004.
45. Volkwein JC; Vinson RP; Page SJ; McWilliams LJ; Joy GJ; Mischler SE; Tuchman DP Laboratory and Field Performance of a Continuously Measuring Personal Respirable Dust Monitor; Department of Health and Human Services, Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health: Washington, DC, USA, 2006.
46. Kohler J NIOSH analysis of comments questioning the use of the CPDM. In *MSHA Public Hearings on the Proposed Rule to Limit Miners' Exposure to Coalmine Dust*; AB64-COMM-93; Mine Safety and Health Administration: Arlington, VA, USA, 2010.
47. Page SJ; Volkwein JC; Vinson RP; Joy GJ; Mischler SE; Tuchman DP; McWilliams LJ Equivalency of a personal dust monitor to the current United States coal mine respirable dust sampler. *J. Environ. Monit.* 2008, 10, 96–101. [PubMed: 18175022]
48. Belle B Pairwise evaluation of PDM3700 and traditional gravimetric sampler for personal dust exposure assessment. In *Proceedings, The Australian Mine Ventilation Conference*; The Australasian Institute of Mining and Metallurgy: Brisbane, QLD, Australia, 2017.
49. Lidén G; Gudmundsson A Optimization of a cyclone to the 1993 international sampling convention for respirable dust. *Appl. Occup. Environ. Hyg.* 1996, 11, 1398–1408.
50. Verpaele S; Jouret J A comparison of the performance of samplers for respirable dust in workplaces and laboratory analysis for respirable quartz. *Ann. Occup. Hyg.* 2013, 57, 54–62. [PubMed: 22826536]
51. Watson JG; Chow JC; Shah JJ; Pace TG The effect of sampling inlets on the PM10 and PM15 to TSP concentration ratios. *J. Air Pollut. Control Assoc.* 1983, 33, 114–119.
52. Farcas D; Lee T; Chisholm WP; Soo JC; Harper M Replacement of filters for respirable quartz measurement in coal mine dust by infrared spectroscopy. *J. Occup. Environ. Hyg.* 2016, 13, D16–D22. [PubMed: 26375614]
53. Mine Safety and Health Administration (MSHA). Infrared Determination of Quartz in Respirable Coal Mine Dust-Method No. MSHA P7; Mine Safety and Health Administration: Pittsburgh, PA, USA, 2013.
54. The National Institute of Occupational Safety and Health (NIOSH). Quartz in Coal Mine Dust, by IR (redeposition)—NIOSH Method 7603; The National Institute for Occupational Safety and Health (NIOSH): Washington, DC, USA, 2003.
55. Lorberau CD; Carsey TP; Fischbach TJ; Mulligan KJ Evaluation of direct-on-filter methods for the determination of respirable α -quartz. *Appl. Occup. Environ. Hyg.* 1990, 5, 27–35.
56. Miller AL; Drake PL; Murphy NC; Cauda EG; Lebouf RF; Markevicius G Deposition uniformity of coal dust on filters and its effect on the accuracy of FTIR analyses for silica. *Aerosol Sci. Technol.* 2013, 47, 724–733. [PubMed: 26719603]

57. Ashley EL; Cauda E; Chubb LG; Tuchman DP; Rubinstein EN Performance comparison of four portable FTIR instruments for direct-on-filter measurement of respirable crystalline silica. *Ann. Work. Expo. Health* 2020, 64, 536–546. [PubMed: 32266371]
58. Cauda E; Miller A; Drake P Promoting early exposure monitoring for respirable crystalline silica: Taking the laboratory to the mine site. *J. Occup. Environ. Hyg.* 2016, 13, D39–D45. [PubMed: 26558490]
59. Hart JF; Autenrieth DA; Cauda E; Chubb L; Spear TM; Wock S; Rosenthal S A comparison of respirable crystalline silica concentration measurements using a direct-on-filter Fourier transform infrared (FT-IR) transmission method vs. a traditional laboratory X-ray diffraction method. *J. Occup. Environ. Hyg.* 2018, 15, 743–754. [PubMed: 29985762]
60. Miller AL; Drake PL; Murphy NC; Noll JD; Volkwein JC Evaluating portable infrared spectrometers for measuring the silica content of coal dust. *J. Environ. Monit.* 2012, 14, 48–55. [PubMed: 22130611]
61. Miller AL; Weakley AT; Griffiths PR; Cauda EG; Bayman S Direct-on-filter α -quartz estimation in respirable coal mine dust using transmission Fourier Transform Infrared Spectrometry and partial least squares regression. *Appl. Spectrosc.* 2017, 71, 1014–1024. [PubMed: 27645724]
62. Stach R; Barone T; Cauda E; Krebs P; Pejčić B; Daboss S; Mizaikoff B Direct infrared spectroscopy for the size-independent identification and quantification of respirable particles relative mass in mine dusts. *Anal. Bioanal. Chem.* 2020, 412, 3499–3508. [PubMed: 32285183]
63. Pampena JD; Cauda EG; Chubb G; Meadows JJ Use of the field-based silica monitoring technique in a coal mine: A case study. *Min. Metall. Explor.* 2020, 37, 717–726. [PubMed: 35836821]
64. Miller AL; Murphy NC; Bayman SJ; Briggs ZP; Kilpatrick AD; Quinn CA; Wadas MR; Cauda EG; Griffiths PR Evaluation of diffuse reflection infrared spectrometry for end-of-shift measurement of α -quartz in coal dust samples. *J. Occup. Environ. Hyg.* 2015, 12, 421–430. [PubMed: 25636081]
65. Chow JC; Lowenthal DH; Chen L-WA; Wang XL; Watson JG Mass reconstruction methods for PM_{2.5}: A review. *Air Qual. Atmos. Health* 2015, 8, 243–263. [PubMed: 26052367]
66. Lippmann M Sampling aerosols by filtration. In *Air Sampling Instruments for Evaluation of Atmospheric Contaminants*, 7th ed.; Hering SV, Ed.; American Conference of Governmental Industrial Hygienists: Cincinnati, OH, USA, 1989; pp. 305–336.
67. Millette JR; Brown RS Chapter 13—Environmental forensic microscopy. In *Introduction to Environmental Forensics*, 3rd ed.; Murphy BL, Morrison RD, Eds.; Academic Press: San Diego, CA, USA, 2015; pp. 487–511.
68. U.S. Environmental Protection Agency (U.S. EPA). National ambient air quality standards for particulate matter: Final rule. *Fed. Regist.* 1997, 62, 38651–38760.
69. Rattigan OV; Carpenter AC; Felton HD; Civerolo KL Optical carbon analysis on Teflon filters from the FRM network in New York. *Atmos. Pollut. Res.* 2021, 12, 101163.
70. Watson JG; Chow JC; Frazier CA X-ray fluorescence analysis of ambient air samples. In *Elemental Analysis of Airborne Particles*; Landsberger S, Creatchman M, Eds.; Advances in Environmental, Industrial and Process Control Technologies; Gordon and Breach Science: Amsterdam, The Netherlands, 1999; Volume 1, pp. 67–96.
71. Chow JC; Watson JG Enhanced ion chromatographic speciation of water-soluble PM_{2.5} to improve aerosol source apportionment. *Aerosol Sci. Eng.* 2017, 1, 7–24.
72. Solomon PA; Crumpler D; Flanagan JB; Jayanty RKM; Rickman EE; McDade CE US National PM_{2.5} Chemical Speciation Monitoring Networks-CSN and IMPROVE: Description of networks. *J. Air Waste Manag. Assoc.* 2014, 64, 1410–1438. [PubMed: 25562937]
73. Chow JC; Watson JG; Lowenthal DH; Magliano KL Loss of PM_{2.5} nitrate from filter samples in central California. *J. Air Waste Manag. Assoc.* 2005, 55, 1158–1168. [PubMed: 16187585]
74. Chow JC; Watson JG; Chen L-WA; Chang M-CO; Robinson NF; Trimble DL; Kohl SD The IMPROVE_A temperature protocol for thermal/optical carbon analysis: Maintaining consistency with a long-term database. *J. Air Waste Manag. Assoc.* 2007, 57, 1014–1023. [PubMed: 17912920]

75. Chow JC; Watson JG; Robles J; Wang XL; Chen L-WA; Trimble DL; Kohl SD; Tropp RJ; Fung KK Quality assurance and quality control for thermal/optical analysis of aerosol samples for organic and elemental carbon. *Anal. Bioanal. Chem.* 2011, 401, 3141–3152. [PubMed: 21626190]
76. Chow JC; Wang XL; Sumlin BJ; Gronstal SB; Chen L-WA; Trimble DL; Kohl SD; Mayorga SR; Riggio GM; Hurbain PR; et al. Optical calibration and equivalence of a multiwavelength thermal/optical carbon analyzer. *Aerosol Air Qual. Res.* 2015, 15, 1145–1159.
77. Cahill TA; Ashbaugh LL; Barone JB; Eldred RA; Feeney PJ; Flocchini RG; Goodart C; Shadoan DJ; Wolfe G Analysis of respirable fractions in atmospheric particulates via sequential filtration. *J. Air Pollut. Control Assoc.* 1977, 27, 675–678. [PubMed: 881526]
78. Wittmaack K Combustion characteristics of water-insoluble elemental and organic carbon in size selected ambient aerosol particles. *Atmos. Chem. Phys.* 2005, 5, 1905–1913.
79. Engelbrecht DR; Cahill TA; Feeney PJ Electrostatic effects on gravimetric analysis of membrane filters. *J. Air Pollut. Control Assoc.* 1980, 30, 391–392.
80. Kröger CR; Morales JR Charge effect in nuclepore filter. *Phys. Scr.* 1988, 37, 270–273.
81. Romo-Kröger CM Measurable electrostatic effects in Nuclepore filters. *J. Air Pollut. Control Assoc.* 1989, 39, 1465–1466.
82. Okuda T; Fujimori E; Hatoya K; Takada H; Kumata H; Nakajima F; Hatakeyama S; Uchida M; Tanaka S; He KB; et al. Rapid and simple determination of multi-elements in aerosol samples collected on quartz fiber filters by using EDXRF coupled with fundamental parameter quantification technique. *Aerosol Air Qual. Res.* 2013, 13, 1864–1876.
83. Watson JG; Chow JC; Chen L; Wang XL; Merrifield TM; Fine PM; Barker K Measurement system evaluation for upwind/downwind sampling of fugitive dust emissions. *Aerosol Air Qual. Res.* 2011, 11, 331–350.
84. Chow JC; Watson JG Ion chromatography in elemental analysis of airborne particles. In *Elemental Analysis of Airborne Particles*; Landsberger S, Creatchman M, Eds.; Advances in Environmental, Industrial and Process Control Technologies; Gordon and Breach Science: Amsterdam, The Netherlands, 1999; Volume 1, pp. 97–137.
85. Chow JC; Watson JG; Pritchett LC; Pierson WR; Frazier CA; Purcell RG The DRI Thermal/Optical Reflectance carbon analysis system: Description, evaluation and applications in U.S. air quality studies. *Atmos. Environ.* 1993, 27, 1185–1201.
86. Chow JC; Watson JG; Crow D; Lowenthal DH; Merrifield TM Comparison of IMPROVE and NIOSH carbon measurements. *Aerosol Sci. Technol.* 2001, 34, 23–34.
87. Chow JC; Watson JG; Chen L-WA; Arnott WP; Moosmüller H; Fung KK Equivalence of elemental carbon by Thermal/Optical Reflectance and Transmittance with different temperature protocols. *Environ. Sci. Technol.* 2004, 38, 4414–4422. [PubMed: 15382872]
88. Chow JC; Yu JZ; Watson JG; Ho SSH; Bohannon TL; Hays MD; Fung KK The application of thermal methods for determining chemical composition of carbonaceous aerosols: A review. *J. Environ. Sci. Health-Part A* 2007, 42, 1521–1541.
89. Raja S; Chandrasekaran SR; Lin L; Xia XY; Hopke PK; Valsaraj KT Analysis of beta attenuation monitor filter rolls for particulate matter speciation. *Aerosol Air Qual. Res.* 2017, 17, 14–23.
90. Watson JG; Chow JC; Chen L-WA; Frank NH Methods to assess carbonaceous aerosol sampling artifacts for IMPROVE and other long-term networks. *J. Air Waste Manag. Assoc.* 2009, 59, 898–911. [PubMed: 19728484]
91. Chow JC; Watson JG; Chen L-WA; Rice J; Frank NH Quantification of PM_{2.5} organic carbon sampling artifacts in US networks. *Atmos. Chem. Phys.* 2010, 10, 5223–5239.
92. Lindeken CL; Morgin RL; Petrock KF Collection efficiency of Whatman 41 filter paper for submicron aerosols. *Health Phys.* 1963, 9, 305–308. [PubMed: 13930769]
93. Demuyneck M Determination of irreversible absorption of water by cellulose filters. *Atmos. Environ.* 1975, 9, 523–528.
94. Chow JC; Watson JG; Bowen JL; Frazier CA; Gertler AW; Fung KK; Landis D; Ashbaugh LL A sampling system for reactive species in the western United States. In *Sampling and Analysis of Airborne Pollutants*; Winegar ED, Keith LH, Eds.; Lewis Publishers: Ann Arbor, MI, USA, 1993; pp. 209–228.

95. Chow JC; Watson JG; Lowenthal DH; Hackney R; Magliano KL; Lehrman DE; Smith TB Temporal variations of PM_{2.5}, PM₁₀, and gaseous precursors during the 1995 Integrated Monitoring Study in Central California. In Proceedings, PM_{2.5}: A Fine Particle Standard; Chow JC, Koutrakis P, Eds.; Air & Waste Management Association: Pittsburgh, PA, USA, 1998; pp. 59–77.
96. Chow JC; Watson JG Chemical analyses of particle filter deposits. In Aerosols Handbook: Measurement, Dosimetry, and Health Effects, 2nd ed.; Ruzer L, Harley NH, Eds.; CRC Press; Taylor & Francis: New York, NY, USA, 2013; pp. 179–204.
97. ASTM F316–03; Standard Test Methods for Pore Size Characteristics of Membrane Filters by Bubble Point and Mean Flow Pore Test. American Society for Testing Materials International: West Conshohocken, PA, USA, 2011.
98. Lindsley WG Filter pore size and aerosol sample collection. In NIOSH Manual of Analytical Methods (NMAM), 5th ed.; National Institute for Occupational Safety and Health (NIOSH): Cincinnati, OH, USA, 2016.
99. Sherwood R; Greenhalgh D A personal air sampler. *Ann. Occup. Hyg.* 1960, 2, 127–132. [PubMed: 14445915]
100. Sherwood R Historical perspectives: Realization, development, and first applications of the personal air sampler. *Appl. Occup. Environ. Hyg.* 1997, 12, 229–234.
101. First MW; Silverman L Air sampling with membrane filters. In A.M.A. Archives of Industrial Hygiene and Occupational Medicine; American Medical Association (AMA): Chicago, IL, USA, 1953; Volume 7, pp. 1–11. [PubMed: 13007196]
102. Hinds WC Aerosol Technology: Properties, Behavior, and Measurement of Airborne Particles, 2nd ed.; John Wiley and Sons, Inc.: New York, NY, USA, 1999.
103. Lee KW; Liu BYH On the minimum efficiency and the most penetrating particle size for fibrous filters. *J. Air Pollut. Control Assoc.* 1980, 30, 377–381.
104. Heikkinen MSS; Harley NH Experimental investigation of sintered porous metal filters. *J. Aerosol Sci.* 2000, 31, 721–738.
105. Kim SC; Harrington MS; Pui DYH Experimental study of nanoparticles penetration through commercial filter media. *J. Nanopart. Res.* 2006, 9, 117–125.
106. Zikova N; Ondráček J; Ždímal V Size-resolved penetration through high-efficiency filter media typically used for aerosol sampling. *Aerosol Sci. Technol.* 2015, 49, 239–249.
107. Soo JC; Monaghan K; Lee T; Kashon M; Harper M Air sampling filtration media: Collection efficiency for respirable size-selective sampling. *Aerosol Sci. Technol.* 2016, 50, 76–87. [PubMed: 26834310]
108. Breuer D Analytical performance issues: Flow resistance of samplers for personal monitoring in work areas and requirements for sampling pump performance. *J. Occup. Environ. Hyg.* 2012, 9, D25–D32. [PubMed: 22225399]
109. Yamamoto N; Fujii M; Kumagai K; Yanagisawa Y Time course shift in particle penetration characteristics through capillary pore membrane filters. *J. Aerosol Sci.* 2004, 35, 731–741.
110. Wang Q; Lin XL; Chen DR Effect of dust loading rate on the loading characteristics of high efficiency filter media. *Powder Technol.* 2016, 287, 20–28.
111. Appel BR; Tokiwa Y; Haik M; Kothny EL Artifact particulate sulfate and nitrate formation on filter media. *Atmos. Environ.* 1984, 18, 409–416.
112. Batterman SA; Osak I; Gelman C SO₂ sorption characteristics of air sampling filter media using a new laboratory test. *Atmos. Environ.* 1997, 31, 1041–1047.
113. Hsu YM; Kollett J; Wysocki K; Wu CY; Lundgren DA; Birky BK Positive artifact sulfate formation from SO₂ adsorption in the silica gel sampler used in NIOSH method 7903. *Environ. Sci. Technol.* 2007, 41, 6205–6209. [PubMed: 17937303]
114. Yamashita T; Sasaki T; Fujimura M; Hashimoto Y Adsorption of sulfur dioxide and nitrogen oxides on glass fiber filters for aerosol collection. *Bunseki Kagaku* 1978, 27, T1–T5.
115. Byers RL; Davis JW Sulfur dioxide adsorption and desorption on various filter media. *J. Air Pollut. Control Assoc.* 1970, 20, 236–238.

116. Gilbert J; Sartim R; Suringar ME; Richards J Positive bias in particulate matter emissions data due to sulfur dioxide adsorption and oxidation on glass fiber filters. *J. Air Waste Manag. Assoc.* 2021, 71, 1076–1084. [PubMed: 33709882]
117. Watson JG; Chow JC; Lurmann FW; Musarra S Ammonium nitrate, nitric acid, and ammonia equilibrium in wintertime Phoenix, Arizona. *J. Air Waste Manag. Assoc.* 1994, 44, 405–412.
118. Hering SV; Cass GR The magnitude of bias in the measurement of PM_{2.5} arising from volatilization of particulate nitrate from Teflon filters. *J. Air Waste Manag. Assoc.* 1999, 49, 725–733. [PubMed: 26355375]
119. Harrison RM; Kitto AMN Field intercomparison of filter pack and denuder sampling methods for reactive gaseous and particulate pollutants. *Atmos. Environ.* 1990, 24, 2633–2640.
120. Febo A; Perrino C; Allegrini I Field intercomparison exercise on nitric acid and nitrate measurement (Rome, 1988)—A critical approach to the evaluation of the results. *Sci. Total Environ.* 1993, 133, 39–71.
121. Keck L; Wittmaack K Laboratory studies on the retention of nitric acid, hydrochloric acid and ammonia on aerosol filters. *Atmos. Environ.* 2005, 39, 2157–2162.
122. Keck L; Wittmaack K Effect of filter type and temperature on volatilisation losses from ammonium salts in aerosol matter. *Atmos. Environ.* 2005, 39, 4093–4100.
123. Keck L; Wittmaack K Simplified approach to measuring semivolatile inorganic particulate matter using a denuded cellulose filter without backup filters. *Atmos. Environ.* 2006, 40, 7106–7114.
124. Turpin BJ; Huntzicker JJ; Hering SV Investigation of organic aerosol sampling artifacts in the Los Angeles Basin. *Atmos. Environ.* 1994, 28, 3061–3071.
125. Allen G; Rector L; Butcher T; Trojanowski R Evaluation of alternative filter media for particulate matter emission testing of residential wood heating devices. *J. Air Waste Manag. Assoc.* 2017, 67, 1055–1060.
126. Tsai CJ; Chang CT; Shih BH; Aggarwal SG; Li SN; Chein HM; Shih TS The effect of environmental conditions and electrical charge on the weighing accuracy of different filter materials. *Sci. Total Environ.* 2002, 293, 201–206. [PubMed: 12109473]
127. Chow JC Summary of the 1995 A&WMA critical review: Measurement methods to determine compliance with ambient air quality standards for suspended particles. *EM* 1995, 1, 12–15.
128. Stacey P; Kauffer E; Moulut J-C; Dion C; Beaulant M; Fernandez P; Key-Schwartz R; Friede B; Wake D An International Comparison of the Crystallinity of Calibration Materials for the Analysis of Respirable α -Quartz Using X-Ray Diffraction and a Comparison with Results from the Infrared KBr Disc Method. *Ann. Occup. Hyg.* 2009, 53, 639–649. [PubMed: 19531809]
129. Watson JG; Bowen JL; Chow JC; Rogers CF; Ruby MG; Rood MJ; Egami RT High volume measurement of size classified suspended particulate matter. In *Methods of Air Sampling and Analysis*, 3rd ed.; Lodge JP, Ed.; Lewis Publishers, Inc.: Chelsea, MI, USA, 1989; pp. 427–439.
130. Ampian SG; Virta RL Crystalline Silica Overview: Occurrence and Analysis; U.S. Bureau of Mines: Washington, DC, USA, 1992.
131. Abbasi B; Wang XL; Chow JC; Watson JG; Peik B; Nasiri V; Riemenschnitter KB; Elahifard M Review of respirable coal mine dust characterization for mass concentration, size distribution and chemical composition. *Minerals* 2021, 11, 426.
132. Kauffer E; Masson A; Moulut JC; Lecaue T; Protois JC Comparison of direct (X-ray diffraction and infrared spectrophotometry) and indirect (infrared spectrophotometry) methods for the analysis of alpha-quartz in airborne dusts. *Ann. Occup. Hyg.* 2005, 49, 661–671. [PubMed: 15972798]
133. Soo JC; Lee T; Chisholm WP; Farcas D; Schwegler-Berry D; Harper M Treated and untreated rock dust: Quartz content and physical characterization. *J. Occup. Environ. Hyg.* 2016, 13, D201–D207. [PubMed: 27314444]
134. Key-Schwartz R; Baron P; Bartley D; Rice F; Schlecht P Chapter R: Determination of airborne crystalline silica. In *NIOSH Manual of Analytical Methods*; National Institute for Occupational Safety and Health (NIOSH): Cincinnati, OH, USA, 2003; pp. 260–280.
135. Occupational Safety and Health Administration (OSHA). OSHA Method ID-142: Crystalline Silica: Quartz and Cristobalite; Occupational Safety and Health Administration (OSHA): Washington, DC, USA, 2016.

136. Griffiths PR; de Haseth JA Fourier Transfer Infrared Spectrometry, 2nd ed.; John Wiley & Sons: New York, NY, USA, 2007.
137. Beer A Bestimmung der Absorption des rothen Lichts in farbigen Flüssigkeiten (Determination of the absorption of red light in colored liquids). *Ann. Phys. Chem.* 1852, 86, 78–88.
138. Ainsworth SM Infrared Analysis of Respirable Coal Mine Dust for Quartz: Thirty-Five Years; Harper M, Lee T, Eds.; ASTM International: West Conshohocken, PA, USA, 2013; pp. 204–221.
139. Anderson CC Collaborative Tests of Two Methods for Determining Free Silica in Airborne Dust; Final Report; SRI International: Menlo Park, CA, USA, 1983.
140. Lorberau C Investigation of the Determination of Respirable Quartz on Filter Media Using Fourier Transform Infrared Spectrophotometry; National Institute for Occupational Safety and Health: Cincinnati, OH, USA, 1989.
141. Lorberau C Investigation of the determination of respirable quartz on filter media using Fourier Transform Infrared Spectrophotometry. *Appl. Occup. Environ. Hyg.* 1990, 5, 348–350.
142. Kingma KJ; Hemley RJ Raman spectroscopic study of microcrystalline silica. *Am. Mineral.* 1994, 79, 269–273.
143. Araujo CF; Nolasco MM; Ribiero AMP; Ribiero-Claro PJA Identification of microplastics using Raman spectroscopy: Latest developments and future prospects. *Water Res.* 2018, 142, 426–440. [PubMed: 29909221]
144. Stacey P; Mader KT; Sammon C Feasibility of the quantification of respirable crystalline silica by mass on aerosol sampling filters using Raman microscopy. *J. Raman Spectrosc.* 2017, 48, 720–725.
145. Friedel RA; Carlson GL Difficult carbonaceous materials and their infra-red and Raman spectra. Reassignments for coal spectra. *Fuel* 1972, 51, 194–198.
146. Shin K; Chung H Wide area coverage Raman spectroscopy for reliable quantitative analysis and its applications. *Analyst* 2013, 138, 3335. [PubMed: 23636144]
147. Zheng LN; Kulkarni P; Birch ME; Ashley K; Wei SJ Analysis of crystalline silica aerosol using portable Raman spectrometry: Feasibility of near real-time measurement. *Anal. Chem.* 2018, 90, 6229–6239. [PubMed: 29659249]
148. Silva LFO; Sampaio CH; Guedes A; Fdez-Ortiz de Vallejuelo S; Madariaga JM Multianalytical approaches to the characterisation of minerals associated with coals and the diagnosis of their potential risk by using combined instrumental microspectroscopic techniques and thermodynamic speciation. *Fuel* 2012, 94, 52–63.
149. Guedes A; Valentim B; Prieto AC; Sanz A; Flores D; Noronha F Characterization of fly ash from a power plant and surroundings by micro-Raman spectroscopy. *Int. J. Coal Geol.* 2008, 73, 359–370.
150. Tuinstra F; Koenig JL Raman spectrum of graphite. *J. Chem. Phys.* 1970, 53, 1126–1130.
151. Lee JY; Ryu JS; Jeong S; Kim JH; Jeong HY; Ra KT; Yang MJ; Chang HJ Elemental and isotopic compositions in blank filters collecting atmospheric particulates. *J. Anal. Sci. Technol.* 2021, 12, 27.
152. Foster RD; Walker RF Quantitative determination of crystalline silica in respirable size dust samples by infrared spectrophotometry. *Analyst* 1984, 109, 1117–1127. [PubMed: 6095699]
153. The National Institute of Occupational Safety and Health (NIOSH). SILICA, Respirable Crystalline, by IR (KBr Pellet); DHHS: Cincinnati, OH, USA, 2017.




Figure 1. Illustration of the combined effects of the particle capture mechanisms for solid particles on the overall filtration efficiency [12].




Figure 2.

Size-resolved penetration of submicron particles by mixed cellulose esters (MCE), quartz, glass fiber, polycarbonate (PC, 0.4 and 8 μm), and Teflon (Zefluor 2 μm and Teflo 1 and 3 μm) at a face velocity of 40 cm/s [106]. Reprinted by permission of Taylor & Francis Ltd. on behalf of the American Association for Aerosol Research.




Figure 3.

Average and standard deviation of collection efficiency of five types of 37 mm filters with 5 μm pore size tested with a nanoparticle diameter range of 10–400 nm at 1.7 L/min. Aerosol measurements were conducted using three different filters for each filter type ($n = 3$). The two overlapping point symbols for MCE and PVC filters are denoted as solid circle and star symbols [107]. Reprinted by permission of Taylor & Francis Ltd. on behalf of the American Association for Aerosol Research.



Figure 4.

Examples of: (a) a clean PDM filter holder; (b) backside of a PDM filter holder showing structural bars; and (c) a PDM filter with a nonuniform particle deposit affected by the filter holder [64]. Reprinted by permission of Taylor & Francis Ltd.




Figure 5.

Comparison of blank membrane filters between the NIOSH (Spokane, WA, USA) and DRI (Reno, NV, USA) laboratories for: (a) PVC, (b) MCE, and (c) polypropylene filters. The two vertical black dashed lines bracket RCS absorption from 767 to 816 cm⁻¹. The vertical red dashed line represents the wave number (915 cm⁻¹) for correcting kaolinite absorption interference for RCS determination.




Figure 6. Raman spectra for: (a) crystalline silica on silver-membrane and polycarbonate (PC)-membrane filters; and (b) samples of pure coal and coal mixed with 1 μg and 2 μg of RCS on a PC filter.

Table 1.

Specification of MSHA coal mine dust monitors [39].

Sampler Type	Coal Mine Dust Personal Sampling Units (CMDPSU)	Personal Dust Monitor (PDM 3700)
Manufacturer	Zefon International, Ocala, FL, USA (zefon.com , accessed on 1 October 2022)	Thermo Fisher, Waltham, MA, USA (thermofisher.com , accessed on 1 October 2022)
Averaging Time	Integrated 8 h sample	Real time with 30 min average
Size-selective Inlet (d_{50} of 4 μm)	Dorr-Oliver nylon cyclone by Zefon	Higgins Dewell (HD) cyclone (Model BG14CP conductive plastic by Mesa Labs, Lakewood, CO, USA, mesalabs.com accessed on 1 October 2022)
Filter Cassette Assembly	2, 3, or 4 piece conductive housing cassettes	14 mm circular polypropylene base with a hollow axial stem
Flow Rate	2.0 L/min ($\pm 5\%$ in mine) for coal	2.2 L/min ($\pm 2.5\%$)
Operating Temperature	0 to 45 °C	−20 to +40 °C
Filter Type (Diameter)	37 mm, 5 μm pore size polyvinyl chloride (PVC)	13 mm Teflon-coated borosilicate glass-fiber filter (TX40HI20, Pall Corporation, East Hills, NY, USA, pall.com , accessed on 1 October 2022)
Filter Exposed Area	784 mm ²	132.7 mm ²
Face Velocity	4.25 cm/s	27.6 cm/s
Mass Loading Range	0–200 mg	0.1–4 mg
Mass Determination Method	Gravimetry	Real-time TEOM inertial microbalance
Accuracy	Accurate measurements are possible within $\pm 5\%$	$\pm 25\%$ of the reference method for concentrations > 0.2 mg/m ³
Physical Dimensions	Sampling Tube: 92 cm Cyclone Assembly: 6 cm (d) \times 15 cm (h) \times 6 cm (w) Pump: 5.7 cm (d) \times 10.8 cm (h) \times 10.2 cm (w)	Sampling Tube: 92 cm Cyclone Assembly: 5.08 cm (w) \times 4.32 cm (d) \times 9.91 cm (h) Monitor: 24.31 cm (w) \times 8.26 cm (d) \times 17.15 cm (h)
Sampler Weight	Pump: 0.65 kg with battery pack	2 kg
Pump Type	Escort ELF Pump	Internal sampling pump
Power Requirements	48 volt battery pack of 4 NiMH cells	Lithium ion battery assembly

Table 2.

Filter types, pore sizes, permeabilities, and collection efficiency for filters applicable to mine exposure monitors [22].

Filter Type	Filter Material (Manufacturer)	Pore Size (µm)	Filter Permeability Face Velocity ^a (cm/s)	Range of Collection Efficiency ^b
Teflon membrane	Fluoropore (PTFE-polyethylene reinforced, Millipore Sigma, Burlington, MA, USA)	3	23.5	98.2%–99.8%
	Teflon (Gelman Sciences, Hilliard, OH, USA)	5	56.8	85%–99.9%
	Teflon (Ghia SKC, Eighty Four, PA, USA)	2	23.4	99.89%–99.99%
		3	24.2	92%–98.98%
	Teflon (Zefluor, Millipore Sigma, Burlington, MA, USA)	2	32.5	94.6%–99.96%
		3	31.6	88%–99.9%
Silver membrane	Pure metallic silver (Sterlitech, Auburn, WA, USA)	0.45	1.8	93.6%–99.98%
		0.8	6.2	90%–99.6%
		1.2	9.2	73%–99.7%
		5	19.0	25%–99.2%
Polyvinyl chloride (PVC) membrane	PVC (Metrical)	0.8	2.7	99.96%–>99.99%
	PVC (Metrical)	5	51	49%–98.8%
	PVC (Millipore Sigma, Burlington, MA, USA)	2	5	88%–99.9%
	PVC-5 (Millipore Sigma, Burlington, MA, USA)	5	11	96.7%–>99.9%
Cellulose acetate/nitrate membrane	MF-RA (Millipore Sigma, Burlington, MA, USA)	1.2	6.2	>99.9%
	MF-SS (Millipore Sigma, Burlington, MA, USA)	3	7.5	98.5%–99.9%
	MF-SM (V)	5	10	98.1%–99.9%
	MF-SC (Millipore Sigma, Burlington, MA, USA)	8	14.1	92%–99.9%
Capillary pore membrane	Polycarbonate (Nuclepore, Whatman-Cytiva, Little Chalfont Buckinghamshire, UK)	0.4	2.9	78%–99.99%
		0.6	2.1	53%–99.5%
		5	30.7	6%–90.7%
		8	21.2	1%–90.5%
Quartz fiber	2500 QAO (Pallflex, Pall Corp., Duncan, SC, USA)	NA	41	84%–99.9%
Teflon-coated glass fiber	TX40HI20 (Pall Corp., Duncan, SC, USA)	NA	15.1	92.6%–99.6%
	TX40HI20 (2nd lot, Pall Corp., Duncan, SC, USA)	NA	9	98.9%–99.9%
Cellulose fiber	Whatman 40 (Whatman-Cytiva, Little Chalfont Buckinghamshire, UK)	NA	3.7	77%–99.99%
	Whatman 41 (Whatman-Cytiva, Little Chalfont Buckinghamshire, UK)	NA	16.9	43%–99.5%
Glass fiber	Microquartz (Gelman Sciences, Hilliard, OH, USA)	NA	14.1	98.5%–99.99%

Filter Type	Filter Material (Manufacturer)	Pore Size (µm)	Filter Permeability Face Velocity ^a (cm/s)	Range of Collection Efficiency ^b
	GF/A (Whatman-Cytiva, Little Chalfont Buckinghamshire, UK)	NA	14.5	99%–99.99%

^aFilter permeability is characterized by the face velocity measured at a pressure drop of 1.3 kPa (1 cm Hg) across the filter.

^bThe efficiency range corresponds to particle diameters in the range of 0.035–1 µm, a pressure drop of 1.3–40 kPa (1–30 cm Hg), and a face velocity range of 1–100 cm/s.

Table 3.

Summary of collection efficiency on membrane and capillary pore filters over a range of pore sizes and flow rates [107].

	Filter Type ^a				
	PTFE	PVC	Silver Membrane	Polycarbonate	MCE
No. of filters tested	171	171	168	171	162
Median	99.86%	99.74%	96.07%	98.01%	99.99%
Mean \pm standard deviation	99.02 \pm 2.25%	98.85 \pm 2.96%	86.46 \pm 20.3%	85.32 \pm 22.2%	99.5 \pm 4.76%
Lowest collection efficiency	94.76%	92.98%	42.10%	22.48%	98.82%
Pore size ^b	5 μ m	5 μ m	5 μ m	5 μ m	0.45 μ m
Flow rate ^b	1.7 L/min	11.2 L/min	4.4 L/min	1.7 L/min	2.5 L/min
Face velocity ^{b,c}	3.11 cm/s	20.5 cm/s	8.06 cm/s	3.11 cm/s	4.58 cm/s
Vendor ^b	Pall	SKC	Sterlitech	Millipore	SKC

^aPTFE: polytetrafluoroethylene (0.45, 1.2, and 5 μ m); PVC: polyvinyl chloride (0.8 and 5 μ m); MCE: mixed cellulose esters (0.45, 0.8, 1.2, and 5 μ m); polycarbonate (0.4, 0.8, 2, and 5 μ m); and silver membrane (0.45, 0.8, 1.2, and 5 μ m).

^bConditions corresponding to the lowest collection efficiency.

^cBased on exposure area of 9.1 cm² in the 37 mm cassette.

Table 4.

Infrared absorbance of five 25 mm-diameter membrane filters [140].

Filter Name	Filter Type ^a	Gravimetric Weight	Average Absorbance ^b at Selected Quartz Bands ^c		
			695 cm ⁻¹	779 cm ⁻¹	798 cm ⁻¹
Nuclepore	PE	4.90 mg	0.057 ± 0.005	0.116 ± 0.004	0.157 ± 0.003
Teflo	PTFE	4.50 mg	0.157 ± 0.021	0.074 ± 0.011	0.063 ± 0.010
DM-450 (0.45 µm pore size)	PVC/A	12.60 mg	0.163 ± 0.003	0.078 ± 0.001	0.074 ± 0.001
DM-800 (0.8 µm pore size)	PVC/A	14.13 mg	0.205 ± 0.005	0.111 ± 0.003	0.108 ± 0.003
GLA-5000	PVC/A	5.31 mg	0.222 ± 0.005	0.105 ± 0.006	0.112 ± 0.006

^aPTFE = polytetrafluoroethylene; PVC = polyvinylchloride; PVC/A = polyvinylchloride–acrylonitrile copolymer; and PE = Nuclepore polyester.

^bFTIR identify functional groups by measuring absorption of infrared radiation as a function of wavelengths. Absorbance (A) is a unitless measure of the optical density: $A = \log_{10}(I_0/I)$, where I_0 is the intensity of incident light and I is the intensity of transmitted light. Absorbance is defined by Beer's law and is linearly proportional to the concentration of light-absorbing species.

^cAnalyzed by Nicolet 60-SX FTIR spectrometer.

Table 5.

Limits of detection (LOD) and limits of quantification (LOQ) for membrane filters analyzed by FTIR [52].

47 mm Filter Type ^b (Pore Size)	LOD (µg) ^a		LOQ (µg) ^a	
	MSHA ^c	NIOSH ^c	MSHA ^c	NIOSH ^c
DM-450 (0.45 µm)	0.99	0.59	3.3	2
PVC (5 µm)	0.52	1.5	1.7	4.9
Polypropylene (0.45 µm)	0.72	1.2	2.4	4
Nylon (0.45 µm)	1.8	1.5	6	4.9

^aStandard deviations of 12 blank filters in 47 mm diameter (after rinsing with isopropyl alcohol) determined by FTIR at 800 cm⁻¹ absorbance (LOD = 3 times standard deviations of the average blanks, LOQ = 10 times standard deviations of the average blanks) [52].

^bDM-450: a vinyl/acrylic co-polymer membrane filter from Pall Life Science; polyvinylchloride (PVC 547) from Zefon; polypropylene (PP04547100) from Sterlitech; and Nylon (HNWP04700) from Millipore.

^cMSHA Method P-7 [53]; NIOSH Method 7603 [54].