

NIOSH Respiratory Protective Device Information

NIOSH CA 2022-1049

November 2022

Revocation of Seven Honeywell Int'l. Inc. SAF-T-FIT Approvals Labeled as Sperian Products



**Centers for Disease Control
and Prevention**
National Institute for Occupational
Safety and Health

Revocation of Seven Honeywell Int'l. Inc. SAF-T-FIT Approvals Labeled as Sperian Products

The National Institute for Occupational Safety and Health (NIOSH) has **revoked seven approval numbers (provided below) issued to Honeywell Int'l. Inc. that were labeled as Sperian SAF-T-FIT filtering facepiece respirators.**

Any Honeywell Int'l. Inc. respirator marked with a NIOSH approval label that includes any of the approval numbers below is no longer approved by NIOSH, including private label versions.

- TC-84A-3745
- TC-84A-3769
- TC-84A-3789
- TC-84A-3845
- TC-84A-3877
- TC-84A-3879
- TC-84A-4270

Revocation also means that respirators bearing any of the NIOSH approval numbers listed above may no longer be used, manufactured, assembled, sold, or distributed. As of October 27, 2022, respirators bearing the NIOSH approval numbers listed above are no longer considered approved and may not be used in fulfillment of any standard or statute requiring the use of a NIOSH Approved® respirator.

Honeywell Int'l. Inc. failed to control the design, labeling, and quality management for the approvals listed above as required by 42 C.F.R. Part 84. Thus, NIOSH has revoked these approvals in accordance with its authority as set forth in 42 C.F.R. § 84.34, which states, *"The Institute reserves the right to revoke, for cause, any certificate of approval issued pursuant to the provisions of this part. Such causes include, but are not limited to, misuse of approval labels and markings, misleading advertising, and failure to maintain or cause to be maintained the quality control requirements of the certificate of approval."*

NIOSH Approved® is a registered certification mark of the U.S. Department of Health and Human Services (HHS).