

Tetramethyl Succinonitrile

Analyte:	Tetramethyl Succinonitrile	Method No.: S155
Matrix:	Air	Range: 1.80-8.20 mg/cu m
OSHA Standard:	0.5 ppm (3 mg/cu m)-skin	Precision ($\overline{CV}_T$): 0.075
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 7/4/75 .

1. Principle of the Method (Reference 11.1)

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained for standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 1.80-8.20 mg/cu m at an atmospheric temperature and pressure of 22°C and 760 mm Hg, using a 55-liter sample. Under the conditions of sample size (55 liters) the probable useful range of this method is 0.5-10 mg/cu m. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of tetramethyl succinonitrile and other substances in the air. The first section of the charcoal tube was found to hold at least 0.78 mg of tetramethyl succinonitrile when a test atmosphere containing 3.63 mg/cu m of tetramethyl succinonitrile in air was sampled at 0.90 liter per minute for 240 minutes; breakthrough was not observed at this time since no tetramethyl succinonitrile was detected in the backup section of

the charcoal tube. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interferences

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When interfering compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation ($\overline{CV}_T$) for the total analytical and sampling method in the range of 1.80-8.20 mg/cu m was 0.075. This value corresponds to a 0.23 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 A collection efficiency of 1.00 was determined for the collecting medium, thus, no bias was introduced in the sample collection step. There was also no apparent bias in the sampling and analytical method for which a desorption efficiency correction was made. Thus, $\overline{CV}_T$ is a satisfactory measure of both accuracy and precision of the sampling and analytical method.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate.
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (6-ft X 1/4-in I.D. stainless steel) packed with 5% SE-30 on 80/100 mesh Chromosorb W DMCS.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 Two-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 1.0-ml delivery pipets.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 Tetramethyl Succinonitrile, reagent grade.

7.3 Hexane, reagent grade.

7.4 Acetone, reagent grade.

7.5 Purified nitrogen.

7.6 Prepurified hydrogen.

7.7 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A sample size of 50 liters is recommended. Sample at a flow of 1.0 liter per minute or less. The flow rate should be known with an accuracy of at least ±5%.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

- 8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.
- 8.3.8 With each batch of ten samples submit one tube from the same lot of tubes which was used for sample collection and which is subjected to exactly the same handling as the samples except that no air is drawn through it. Label this as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 2-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 1.0 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
1. 50 ml/min (60 psig) nitrogen carrier gas flow
 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 3. 500 ml/min (50 psig) air flow to detector
 4. 230°C injector temperature
 5. 255°C manifold temperature (detector)
 6. 90°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of a 20% (V/V) acetone-hexane solution of tetramethyl succinonitrile containing 34.75 mg/ml is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 50-liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 1.0 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight recovered (mg)}}{\text{Weight added (mg)}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/1.0 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/1.0 ml versus peak area. Note: Since no internal standard is used in this method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/1.0 ml carbon disulfide, and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P	= pressure (mm Hg) of air sampled
T	= temperature (°C) of air sampled
24.45	= molar volume (liter/mole) at 25°C and 760 mm Hg
M.W.	= molecular weight (g/mole) of analyte
760	= standard pressure (mm Hg)
298	= standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.