

BIBLIOGRAPHIC DATA SHEET		1. Report No. NIOSH-SCP-F	2.	PB 246 149	
4. Title and Subtitle NIOSH Analytical Methods for Set F			5. Report Date		
			6.		
7. Author(s)			8. Performing Organization Rept. No.		
9. Performing Organization Name and Address Stanford Research Institute 333 Ravenswood Avenue Menlo Park, CA 94025			10. Project/Task/Work Unit No.		
			11. Contract/Grant No. CDC-99-74-45		
12. Sponsoring Organization Name and Address Dr. David G. Taylor National Institute for Occupational Safety and Health 1014 Broadway Cincinnati, Ohio 45202			13. Type of Report & Period Covered		
			14.		
15. Supplementary Notes					
16. Abstracts Industrial Hygiene sampling and analytical monitoring methods validated under the joint NIOSH/OSHA Standards Completion Program for Set F are contained herein. Monitoring methods for the following compounds are included: <div style="display: flex; justify-content: space-between;"> <div> 2-Butoxy ethanol Chlorinated camphene Dipropylene glycol methyl ether Glycidol Methyl cellosolve Phenyl ether vapor Phenyl glycidyl ether Tetrahydrofuran </div> <div> n-Butyl glycidyl ether Cyclohexene Ethyl ether Isopropyl glycidyl ether Methylal Phenyl ether-biphenyl vapor mixture Propylene oxide </div> </div>					
17. Key Words and Document Analysis. 17a. Descriptors Chemical analysis Sampling <div style="display: flex; justify-content: space-between;"> <div> *2-Butoxy ethanol *Chlorinated camphene *Dipropylene glycol methyl ether *Glycidol *Methyl cellosolve *Phenyl ether vapor *Phenyl glycidyl ether *Tetrahydrofuran </div> <div> *n-Butyl glycidyl ether *Cyclohexene *Ethyl ether *Isopropyl glycidyl ether *Methylal *Phenyl ether-biphenyl vapor mixture *Propylene oxide </div> </div>					
17b. Identifiers/Open-Ended Terms					
Reproduced by NATIONAL TECHNICAL INFORMATION SERVICE U S Department of Commerce Springfield VA 22151					
17c. COSATI Field/Group					
18. Availability Statement Release unlimited			19. Security Class (This Report) UNCLASSIFIED		
			20. Security Class (This Page) UNCLASSIFIED		
			21. No. of Pages		

NIOSH Analytical Methods for Set F

A joint National Institute for Occupational Safety and Health (NIOSH)/Occupational Safety and Health Administration (OSHA) Standards Completion Program will complete standards for approximately 400 air contaminants presently listed in Tables Z-1, Z-2, and Z-3 of 29 CFR Part 1910.1000 by adding other requirements of a standard required under Section 6(b)(7) and 8(c)(3) of the Occupational Safety and Health Act of 1970 (PL 91-596). These completed standards will then contain, in addition to the permissible exposure limit given in 1910.93, appropriate provisions requiring monitoring of worker exposure, engineering control, personal protection, employee training, medical surveillance, and record keeping.

As a part of the Standards Completion Program, NIOSH is engaged in a two-year study under contract CDC-99-74-45 to validate sampling and analytical procedures for use in monitoring worker exposure to substances listed in Tables Z-1, Z-2, and Z-3. These methods have been validated and are suitable for measuring airborne concentrations of these substances and thus may be used for determining compliance with the standard or the need for control, for research, or whenever there is a need to measure airborne concentrations in the workplace. These analytical methods should not be considered the only methods which may be used to evaluate worker exposure. Other methods meeting the accuracy requirements in the standard may also be used.

These analytical methods will be periodically modified as new developments in science and technology require.

Set F

2-Butoxy ethanol	n-Butyl glycidyl ether
Chlorinated camphene	Cyclohexene
Dipropylene glycol methyl ether	Ethyl ether
Glycidol	Isopropyl glycidyl ether
Methyl cellosolve	Methylal
Phenyl ether vapor	Phenyl ether-biphenyl vapor mixture
Phenyl glycidyl ether	Propylene oxide
Tetrahydrofuran	

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE
Public Health Service
Center for Disease Control
National Institute for Occupational Safety and Health
Division of Laboratories and Criteria Development
Cincinnati, Ohio
October 1975

2-Butoxy Ethanol
(Butyl Cellosolve)

Analyte:	2-Butoxy Ethanol	Method No.: S76
Matrix:	Air	Range: 124-490 mg/cu m
OSHA Standard:	50 ppm (240 mg/cu m)	Precision (\overline{CV}_T): 0.060
Procedure:	Adsorption on charcoal, desorption with 5% methanol in methylene chloride, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with 5% methanol in methylene chloride.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 124-490 mg/cu m at an atmospheric temperature and pressure of 21°C and 766 mm Hg, using a 10 liter sample. Under the conditions of sample size (10 liters) the probable useful range of this method is 25-720 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 7 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of 2-butoxy ethanol and other substances in the air. The first section of the charcoal tube was found to hold at least 21 mg of 2-butoxy ethanol when a test atmosphere

containing 480 mg/cu m of 2-butoxy ethanol in air was sampled at 0.183 liter per minute for 240 minutes; at that time the concentration of analyte in the effluent was less than 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample. Since 5% methanol in methylene chloride is used rather than carbon disulfide to desorb the 2-butoxy ethanol from the charcoal, it would not be possible to measure methanol or methylene chloride in the sample with this desorbing solvent. If it is suspected that methanol or methylene chloride is present, a separate sample should be collected for methanol or methylene chloride analysis.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 124-490 mg/cu m was 0.060. This value corresponds to a 14 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 8% lower than the "true values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

- 6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

- 7.1 Eluent: 5% methanol in methylene chloride (both reagent grade).
7.2 2-Butoxy Ethanol, reagent grade.
7.3 Purified nitrogen.
7.4 Prepurified hydrogen.
7.5 Filtered compressed air.

8. Procedure

- 8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.
- 8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.
- 8.3 Collection and Shipping of Samples
- 8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).
- 8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.
- 8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.
- 8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.
- 8.3.5 A maximum sample size of 10 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.
- 8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.
- 8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of 5% methanol in methylene chloride is pipetted into each sample container. Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 215°C injector temperature
 - 5. 263°C manifold temperature (detector)
 - 6. 145°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 10 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of the eluent with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml of the eluent, because samples are desorbed in this amount of eluent. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml of the eluent and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

n-Butyl Glycidyl Ether

Analyte:	n-Butyl Glycidyl Ether	Method No.: S81
Matrix:	Air	Range: 133-542 mg/cu m
OSHA Standard:	50 ppm (270 mg/cu m)	Precision (\overline{CV}_T): 0.074
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 133-542 mg/cu m at an atmospheric temperature and pressure of 22°C and 767 mm Hg, using a 10 liter sample. Under the conditions of sample size (10 liters) the probable useful range of this method is 30-810 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 8 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of n-butyl glycidyl ether and other substances in the air. The first section of the charcoal tube was found to hold at least 23 mg of n-butyl glycidyl ether when

a test atmosphere containing 530 mg/cu m of n-butyl glycidyl ether in air was sampled at 0.183 liter per minute for 240 minutes; at that time the concentration of analyte in the effluent was less than 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 133-542 mg/cu m was 0.074. This value corresponds to a 20 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 1.8% lower than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 n-Butyl Glycidyl Ether, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 10 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 180°C injector temperature
 - 5. 275°C manifold temperature (detector)
 - 6. 130°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 10 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

- 11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).
- 11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.
- 11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Chlorinated Camphene
(Toxaphene)

Analyte:	Chlorinated Camphene	Method No.: S67
Matrix:	Air	Range: 0.225-1.155 mg/cu m
OSHA Standard:	0.5 mg/cu m	Precision (\overline{CV}_T): 0.076
Procedure:	Filter collection, petroleum ether extraction, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a cellulose membrane filter to collect particulate matter.
- 1.2 The filter is transferred to a screw cap bottle and extracted with petroleum ether.
- 1.3 An aliquot of the extract is analyzed by electron capture gas chromatography.
- 1.4 The area of the sample peaks is determined and compared with areas obtained from standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 0.225-1.155 mg/cu m at an atmospheric temperature and pressure of 22°C and 761 mm Hg, using a 15-liter sample. The probable useful range of this method is 0.05-1.5 mg/cu m for 15-liter samples.
- 2.2 The upper limit of the range of the method is dependent on the collection efficiency of the cellulose membrane filter. If higher concentrations than those tested are to be sampled, smaller sample volumes should be used. The collection efficiency for chlorinated camphene was considered to be 1.00 ± 0.01 when sampled for 15 minutes at 1 liter per minute from a test atmosphere containing 1 mg/cu m.

3. Interference

- 3.1 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.2 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method and a high electron capture cross section is an interference.
- 3.3 Chlorinated camphene is composed of a mixture of compounds, which chromatograph over a range of retention times. Any compound with an electron capture response in this retention time band will interfere. This would include pesticides, such as aldrin, parathion, dieldrin, DDT and its metabolites, and PCB's.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 0.225-1.155 mg/cu m was 0.076. This value corresponds to a standard deviation of 0.038 mg/cu m at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.1.
- 4.2 A collection efficiency of 1.00 ± 0.01 was determined for the collection medium; thus, no bias was introduced in the sample collection step. There were no apparent biases in the sampling and analytical method for which analytical method recovery corrections were made. Thus \overline{CV}_T is a satisfactory measure of both accuracy and precision of the sampling and analytical method.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Samples collected on filters are analyzed by means of a quick, instrumental method.
- 5.2 Other volatile compounds which give an electron capture response that are dispersed as particulates, are likely to interfere.

6. Apparatus

- 6.1 The sampling unit for the collection of personal air samples for the determination of organic aerosol has the following components:
 - 6.1.1 The filter unit consisting of the filter media (Section 6.2) and a polystyrene 37-mm 3-piece cassette filter holder. Do not use tenite filter holders.
 - 6.1.2 Personal Sampling Pump: A calibrated personal sampling pump whose flow can be determined to an accuracy of $\pm 5\%$ (Reference 11.1) at the recommended flow rate. The pump must be calibrated with a representative filter holder and filter in the line.

- 6.2 Mixed cellulose ester membrane filter, 0.8 micrometer pore size and 37 mm diameter. The filter is held in the three piece cassette supported by a cellulose backup pad.
- 6.3 Screw cap bottles. Within 1 hour after the sample has been collected, the filter is transferred to a clean screw cap bottle (a 45-mm tissue sample holder is satisfactory) for shipping.
- 6.4 Gas chromatograph equipped with an electron capture detector.
- 6.5 Column (6-ft X 1/4-in glass) packed with 3% OV-1 on 100/120 mesh Gas-Chrom Q.
- 6.6 An electronic integrator or some other suitable method for measuring peak areas.
- 6.7 Sample extraction equipment.
 - 6.7.1 Beakers, 50 ml.
 - 6.7.2 Volumetric flasks, 10 ml.
 - 6.7.3 Tweezers and funnels.
- 6.8 Microliter syringes: 10-microliter and other convenient sizes for making standard solutions.
- 6.9 Volumetric flasks: 10-ml and other convenient sizes for making standard solutions.

7. Reagents

- 7.1 Petroleum ether, 30° to 60°C boiling range, suitable for pesticide residue analysis, Mallinckrodt Nanograde[®] or equivalent.
- 7.2 Chlorinated camphene (toxaphene), reagent grade.
- 7.3 Purified nitrogen.

8. Procedure

- 8.1 Cleaning of Equipment. All glassware used for the laboratory analysis as well as the screw cap bottles should be detergent washed and thoroughly rinsed with tap water and distilled water, dried, and then rinsed with petroleum ether.
- 8.2 Calibration of Personal Sampling Pumps. Each personal sampling pump must be calibrated with a representative filter cassette in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

- 8.3.1 Assemble the filter in the three piece filter cassette holder and close firmly to insure that the center ring seals the edge of the filter. The cellulose membrane filter is held in place by a cellulose backup pad.
- 8.3.2 Remove the cassette plugs and attach to the personal sampling pump tubing. Clip the cassette to the worker's lapel.
- 8.3.3 Air being sampled should not pass through any hose or tubing before entering the filter cassette.
- 8.3.4 A sample size of 15 liters is recommended. Sample at a flow of 1.0 liter per minute. The flow rate should be known with an accuracy of $\pm 5\%$.
- 8.3.5 Turn the pump on and begin sample collection. Since it is possible for a filter to become plugged by heavy particulate loading or by the presence of oil mists or other liquids in the air, the pump rotameter should be observed frequently, and the sampling should be terminated at any evidence of a problem.
- 8.3.6 Terminate sampling at the predetermined time and note sample flow rate, collection time and ambient temperature and pressure. If pressure reading is not available, record the elevation.
- 8.3.7 The cellulose filter should be removed from the cassette filter holder and the cellulose backup pad within 1 hour of sampling and placed in a clean screw cap bottle. Care must be taken to handle the filter only with clean tweezers.
- 8.3.8 Carefully record sample identity and all relevant sampling data.
- 8.3.9 With each batch of samples submit one filter which is subjected to exactly the same handling as for the samples except that no air is drawn through it. Label this as a blank.
- 8.3.10 The screw cap bottles in which the samples are stored should be shipped in a suitable container, designed to prevent damage in transit.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. Care must be taken to handle the filter only with clean tweezers. Open the screw cap

bottle, remove the filter, and place the filter in a 50-ml beaker. Rinse the screw cap bottle with 3 ml petroleum ether. Pour the solvent into the beaker. Swirl the solvent over the filter, then decant the extract into a 10-ml volumetric flask. Repeat this process two more times. Dilute the combined extracts to a total of 10 ml with petroleum ether, stopper, and mix.

8.4.2 GC Conditions. The typical operating conditions for the gas chromatograph are:

1. 50 ml/min (60 psig) nitrogen carrier gas flow
2. 250°C detector temperature
3. 230°C injector temperature
4. 205°C column temperature

8.4.3 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

8.4.4 Measurement of area. The areas of the sample peaks are measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed in Section 9. The peaks are not resolved, and therefore the total area of the envelope of peaks is measured as a single area.

8.5 Determination of Analytical Method Recovery

8.5.1 Need for determination. It is anticipated that the extraction scheme is dependent on operator skill and perhaps on the formulation used to disperse the analyte. Thus, it is necessary to determine the recovery of the analyte.

8.5.2 Procedure for determining analytical method recovery. A known amount of petroleum ether solution containing 1.74 mg/ml of chlorinated camphene, preferably equivalent to the sample concentration expected, is added to a representative cellulose membrane filter and air-dried. The analyte is then extracted from the filter and analyzed as described in Section 8.4. An amount of the analyte equivalent to that present in a 15 liter sample at the selected level is used for the extraction studies. Six filters at each of the three levels (0.5X, 1X, and 2X the OSHA standard) are dosed accordingly. A parallel blank filter is also treated in the same manner except that no sample is added to it. All filters are then extracted and analyzed as described in Section 8.4.

Three standards are prepared by injecting the same amount of analyte into 10 ml of petroleum ether with the same syringe used in the preparation of the dosed filters and are analyzed at the same time.

Analytical method recovery (A.M.R.) equals the weight in mg found divided by the weight in mg added to the filter, or

$$\text{A.M.R.} = \frac{\text{Weight found (mg)}}{\text{Weight added (mg)}}$$

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/10 ml petroleum ether, because samples are extracted in this amount of petroleum ether. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/10 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the electron capture response.

10. Calculations

10.1 Read the weight, in mg, corresponding to each envelope of peak areas from the standard curve. No volume corrections are needed, because the standard curve is based on mg/10 ml of petroleum ether and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in sample filter

mg blank = mg found in blank filter

10.3 Divide the total weight by the analytical method recovery (A.M.R.) to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{A.M.R.}}$$

10.4 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.3)} \times 1000 \text{ (liter/cu m)}}{\text{Air Volume Sampled (liter)}}$$

11. Reference

11.1 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

Cyclohexene

Analyte:	Cyclohexene	Method No.: S82
Matrix:	Air	Range: 510-2030 mg/cu m
OSHA Standard:	300 ppm (1010 mg/cu m)	Precision (\overline{CV}_T): 0.073
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 510-2030 mg/cu m at an atmospheric temperature and pressure of 23.5°C and 768 mm Hg, using a 5 liter sample. Under the conditions of sample size (5 liters) the probable useful range of this method is 100-3030 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 15 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of cyclohexene and other substances in the air. The first section of the charcoal tube was found to hold 21 mg of cyclohexene when a test atmosphere containing

2002 mg/cu m of cyclohexene in air was sampled at 0.185 liter per minute for 56 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 510-2030 mg/cu m was 0.073. This value corresponds to a 74 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 9% higher than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate.
(Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (4-ft X 1/4-in stainless steel) packed with 50/80 mesh Porapak, Type Q.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

- 6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

- 7.1 Chromatographic quality carbon disulfide.
7.2 Cyclohexene, reagent grade.
7.3 Purified nitrogen.
7.4 Prepurified hydrogen.
7.5 Filtered compressed air.

8. Procedure

- 8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.
- 8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.
- 8.3 Collection and Shipping of Samples
- 8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).
- 8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.
- 8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.
- 8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.
- 8.3.5 A maximum sample size of 5 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least +5%.
- 8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.
- 8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 195°C injector temperature
 - 5. 260°C manifold temperature (detector)
 - 6. 205°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 5 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Dipropylene Glycol Methyl Ether

Analyte:	Dipropylene Glycol Methyl Ether	Method No.: S69
Matrix:	Air	Range: 314-1197 mg/cu m
OSHA Standard:	100 ppm (610 mg/cu m)	Precision (\overline{CV}_T): 0.064
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 314-1197 mg/cu m at an atmospheric temperature and pressure of 22°C and 765 mm Hg, using a 10 liter sample. Under the conditions of sample size (10 liters) the probable useful range of this method is 60-1830 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 18 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of dipropylene glycol methyl ether and other substances in the air. The first section of the charcoal tube was found to hold 33 mg of dipropylene glycol methyl ether

when a test atmosphere containing 1219 mg/cu m of dipropylene glycol methyl ether in air was sampled at 0.185 liter per minute for 145 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 314-1197 mg/cu m was 0.064. This value corresponds to a 39 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 1.4% higher than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 Dipropylene Glycol Methyl Ether, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 10 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 210°C injector temperature
 - 5. 270°C manifold temperature (detector)
 - 6. 160°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.

8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 10 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Ethyl Ether

Analyte:	Ethyl Ether	Method No.: S80
Matrix:	Air	Range: 606-2400 mg/cu m
OSHA Standard:	400 ppm (1210 mg/cu m)	Precision (\overline{CV}_T): 0.053
Procedure:	Adsorption on charcoal, desorption with ethyl acetate, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with ethyl acetate.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 606-2400 mg/cu m at an atmospheric temperature and pressure of 22°C and 766 mm Hg, using a 3 liter sample. Under the conditions of sample size (3 liters) the probable useful range of this method is 120-3630 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 11 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of ethyl ether and other substances in the air. The first section of the charcoal tube was found to hold 15 mg of ethyl ether when a test atmosphere containing

2470 mg/cu m of ethyl ether in air was sampled at 0.185 liter per minute for 33 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample. Since ethyl acetate is used rather than carbon disulfide to desorb the ethyl ether from the charcoal it would not be possible to measure ethyl acetate in the sample with this desorbing solvent. If it is suspected that ethyl acetate is present, a separate sample should be collected for ethyl acetate analysis.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 606-2400 mg/cu m was 0.053. This value corresponds to a 64 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 5.1% higher than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate.
(Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (4-ft x 1/4-in stainless steel) packed with 50/80 mesh Porapak, Type Q.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

- 6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.
7. Reagents
- 7.1 Eluent: Ethyl acetate, reagent grade
- 7.2 Ethyl ether, reagent grade.
- 7.3 Purified nitrogen.
- 7.4 Prepurified hydrogen.
- 7.5 Filtered compressed air.
8. Procedure
- 8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.
- 8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.
- 8.3 Collection and Shipping of Samples
- 8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).
- 8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.
- 8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.
- 8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.
- 8.3.5 A maximum sample size of 3 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least +5%.
- 8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.
- 8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of ethyl acetate is pipetted into each sample container. Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 195°C injector temperature
 - 5. 250°C manifold temperature (detector)
 - 6. 175°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 3 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of ethyl acetate with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml ethyl acetate, because samples are desorbed in this amount of ethyl acetate. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml ethyl acetate and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Glycidol

Analyte:	Glycidol	Method No.: S70
Matrix:	Air	Range: 73-310 mg/cu m
OSHA Standard:	50 ppm (150 mg/cu m)	Precision (\overline{CV}_T): 0.080
Procedure:	Adsorption on charcoal, desorption with tetra- hydrofuran, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present, and the charcoal tube is immediately refrigerated at 4°C and held at that temperature until the sample is desorbed with tetrahydrofuran.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with tetrahydrofuran.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 73-310 mg/cu m at an atmospheric temperature and pressure of 21°C and 768 mm Hg, using a 50 liter sample. Under the conditions of sample size (50 liters) the probable useful range of this method is 15-450 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 22 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of glycidol and other substances in the air. The first section of the charcoal tube was found to hold 45 mg of glycidol when a test atmosphere containing

300 mg/cu m of glycidol in air was sampled at 0.93 liter per minute for 163 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample. Since tetrahydrofuran is used rather than carbon disulfide to desorb the glycidol from the charcoal, it would not be possible to measure tetrahydrofuran in the sample with this desorbing solvent. If it is suspected that tetrahydrofuran is present, a separate sample should be collected for tetrahydrofuran analysis.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 73-310 mg/cu m was 0.080. This value corresponds to a 12 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 3.1% lower than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.
- 5.4 A disadvantage of the method is that the sample must be kept refrigerated, or recoveries are poor.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.2)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Eluent: Tetrahydrofuran, chromatographic quality.

7.2 Glycidol, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 50 liters is recommended. Sample at a flow of 1.0 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping. Capped tubes should be refrigerated during storage and shipping at 4°C.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of tetrahydrofuran is pipetted into each sample container. Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
1. 50 ml/min (60 psig) nitrogen carrier gas flow
 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 3. 500 ml/min (50 psig) air flow to detector
 4. 225°C injector temperature
 5. 260°C manifold temperature (detector)
 6. 155°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.

- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm. These tubes must be refrigerated after injection of the analyte onto the charcoal. Previous studies have shown that the desorption efficiency decreases with time if the samples are stored at room temperature.

The amount injected is equivalent to that present in a 50 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. The tubes should be kept refrigerated, and they are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of the eluent with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml tetrahydrofuran, because samples are desorbed in this amount of tetrahydrofuran. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml tetrahydrofuran and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

$$\text{mg sample} = \text{mg found in front section of sample tube}$$

$$\text{mg blank} = \text{mg found in front section of blank tube}$$

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

Isopropyl Glycidyl Ether

Analyte:	Isopropyl Glycidyl Ether	Method No.: S77
Matrix:	Air	Range: 121-484 mg/cu m
OSHA Standard:	50 ppm (240 mg/cu m)	Precision (\overline{CV}_T): 0.067
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 121-484 mg/cu m at an atmospheric temperature and pressure of 21°C and 763 mm Hg, using a 10 liter sample. Under the conditions of sample size (10 liters) the probable useful range of this method is 25-720 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 7 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of isopropyl glycidyl ether and other substances in the air. The first section of the charcoal tube was found to hold at least 21 mg of isopropyl glycidyl ether

when a test atmosphere containing 480 mg/cu m of isopropyl glycidyl ether in air was sampled at 0.183 liter per minute for 240 minutes; at that time the concentration of analyte in the effluent was less than 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 121-484 mg/cu m was 0.067. This value corresponds to a 16 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 7% lower than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 Isopropyl Glycidyl Ether, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 10 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least +5%.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 205°C injector temperature
 - 5. 270°C manifold temperature (detector)
 - 6. 115°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 10 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Methyl Cellosolve

Analyte:	Methyl Cellosolve	Method No.: S79
Matrix:	Air	Range: 44-160 mg/cu m
OSHA Standard:	25 ppm (80 mg/cu m)	Precision (\overline{CV}_T): 0.068
Procedure:	Adsorption on charcoal, desorption with 5% methanol in methylene chloride, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with 5% methanol in methylene chloride.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 44-160 mg/cu m at an atmospheric temperature and pressure of 22°C and 768 mm Hg, using a 50 liter sample. Under the conditions of sample size (50 liters) the probable useful range of this method is 8-240 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 12 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of methyl cellosolve and other substances in the air. The first section of the charcoal tube was found to hold 20 mg of methyl cellosolve when a test atmosphere containing

158 mg/cu m of methyl cellosolve in air was sampled at 0.93 liter per minute for 138 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample. Since 5% methanol in methylene chloride is used rather than carbon disulfide to desorb the methyl cellosolve from the charcoal, it would not be possible to measure methanol or methylene chloride in the sample with this desorbing solvent. If it is suspected that methanol or methylene chloride is present, a separate sample should be collected for methanol or methylene chloride analysis.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 44-160 mg/cu m was 0.068. This value corresponds to a 5.4 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 7.2% lower than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within +5% at the recommended flow rate. (Reference 11.2)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Eluent: 5% methanol in methylene chloride (both reagent grade).

7.2 Methyl Cellosolve, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 50 liters is recommended. Sample at a flow of 1.0 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of 5% methanol in methylene chloride is pipetted into each sample container. Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
1. 50 ml/min (60 psig) nitrogen carrier gas flow
 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 3. 500 ml/min (50 psig) air flow to detector
 4. 215°C injector temperature
 5. 268°C manifold temperature (detector)
 6. 95°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 50 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of the eluent with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml of the eluent, because samples are desorbed in this amount of eluent. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml of the eluent and the volume of sample injected is identical to the volume of the standards injected.

72<

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

Methylal
(Dimethoxymethane)

Analyte:	Methylal	Method No.: S71
Matrix:	Air	Range: 1684-6670 mg/cu m
OSHA Standard:	1000 ppm (3110 mg/cu m)	Precision (\overline{CV}_T): 0.060
Procedure:	Adsorption on charcoal, desorption with hexane, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with hexane.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 1684-6670 mg/cu m at an atmospheric temperature and pressure of 22°C and 771 mm Hg, using a 2 liter sample. Under the conditions of sample size (2 liters) the probable useful range of this method is 310-9330 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 19 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of methylal and other substances in the air. The first section of the charcoal tube was found to hold 23 mg of methylal when a test atmosphere containing 6250 mg/cu m

of methylal in air was sample at 0.183 liter per minute for 20 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample. Since hexane is used rather than carbon disulfide to desorb the methylal from the charcoal, it would not be possible to measure hexane in the sample with this desorbing solvent. If it is suspected that hexane is present, a separate sample should be collected for hexane analysis.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 1684-6670 mg/cu m was 0.060. This value corresponds to a 187 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 4.6% higher than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate.
(Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (4-ft x 1/4-in stainless steel) packed with 50/80 mesh Porapak, Type Q.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Eluent: Hexane, reagent grade.

7.2 Methylal, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 2 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least +5%.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of hexane is pipetted into each sample container. Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 190°C injector temperature
 - 5. 250°C manifold temperature (detector)
 - 6. 195°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 2 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of hexane with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml hexane, because samples are desorbed in this amount of hexane. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml hexane and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

$$\text{mg sample} = \text{mg found in front section of sample tube}$$

$$\text{mg blank} = \text{mg found in front section of blank tube}$$

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{\text{P}} \times \frac{\text{T} + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

- 11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).
- 11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.
- 11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Phenyl Ether

Analyte:	Phenyl Ether (vapor)	Method No.: S72
Matrix:	Air	Range: 3-13 mg/cu m
OSHA Standard:	1 ppm (7 mg/cu m)	Precision (\overline{CV}_T): 0.070
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 3-13 mg/cu m at an atmospheric temperature and pressure of 24°C and 766 mm Hg, using a 10 liter sample. Under the conditions of sample size (10 liters) the probable useful range of this method is 0.70-21 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 0.2 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of phenyl ether and other substances in the air. The first section of the charcoal tube was found to hold at least 0.6 mg of phenyl ether when a test atmosphere

containing 13.1 mg/cu m of phenyl ether in air was sampled at 0.185 liter per minute for 240 minutes; at that time the concentration of analyte in the effluent was less than 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 3-13 mg/cu m was 0.070. This value corresponds to a 0.49 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 Collection efficiency of the collecting medium was found to be 1.00 ± 0.01 , so that no bias was introduced in the sample collection step. There were no apparent biases in the sampling and analytical method for which desorption efficiency corrections were made. Thus, \overline{CV}_T is a satisfactory measure of both accuracy and precision of the sampling and analytical method.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.

5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)

6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.

6.3 Gas chromatograph equipped with a flame ionization detector.

6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.

6.5 An electronic integrator or some other suitable method for measuring peak areas.

6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.

6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.

6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 Phenyl Ether, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 10 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least +5%.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
1. 50 ml/min (60 psig) nitrogen carrier gas flow
 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 3. 500 ml/min (50 psig) air flow to detector
 4. 230°C injector temperature
 5. 265°C manifold temperature (detector)
 6. 215°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of hexane solution of phenyl ether containing 17.3 mg/ml is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 10 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Phenyl Ether - Biphenyl Mixture
(vapor)

Analyte:	Phenyl Ether-Biphenyl Mixture	Method No.: S73
Matrix:	Air	Range: 3.86-15.7 mg/cu m
OSHA Standard:	1 ppm (7 mg/cu m)	Precision (\overline{CV}_T): 0.089
Procedure:	Adsorption on silica gel, desorption with benzene, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a silica gel tube to trap the organic vapors present.
- 1.2 The silica gel in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with benzene.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The areas of the resulting peaks are determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 3.86-15.7 mg/cu m at an atmospheric temperature and pressure of 23°C and 761 mm Hg, using a 10 liter sample. Under the conditions of sample size (10 liters) the probable useful range of this method is 0.7-21 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 0.2 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the silica gel tube. This capacity varies with the concentrations of the analyte and other substances in the air. The first section of the silica gel tube was found to hold at least 0.7 mg of analyte when a test atmosphere containing

15.7 mg/cu m of analyte in air was sampled at 0.185 liter per minute for 240 minutes. (The silica gel tube consists of two sections of silica gel separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 Silica gel has a high affinity for water, so organic vapors will not be trapped efficiently in the presence of a high relative humidity. This effect may be important even though there is no visual evidence of condensed water in the silica gel tube.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample. Since benzene is used rather than carbon disulfide to desorb the analyte from the silica gel, it is not possible to measure benzene in the sample with this desorbing solvent. If it is suspected that benzene is present, a separate sample should be collected for benzene analysis.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 3.86-15.7 mg/cu m was 0.089. This value corresponds to a 0.62 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 Collection efficiency of the collecting medium was found to be 1.00 ± 0.01 , so that no bias was introduced in the sample collection step. There were no apparent biases in the sampling and analytical method for which desorption efficiency corrections were made. Thus, \overline{CV}_T is a satisfactory measure of both accuracy and precision of the sampling and analytical method.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the silica gel tube exceeds 25% of that found on the front section, the possibility of sample loss exists.

5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)

6.2 Silica gel tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh silica gel separated by a 2-mm portion of urethane foam. The adsorbing section contains approximately 150 mg of silica gel, the backup section, approximately 75 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.

6.3 Gas chromatograph equipped with a flame ionization detector.

6.4 Column (10-ft x 1/8-in stainless steel) packed with 5% OV-17 on 80/100 mesh, acid washed DMCS Chromosorb W.

6.5 An electronic integrator or some other suitable method for measuring peak areas.

6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.

6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.

6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

- 6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

- 7.1 Chromatographic quality benzene.
- 7.2 Biphenyl, reagent grade.
- 7.3 Phenyl Ether, reagent grade.
- 7.4 Phenyl Ether-Biphenyl Mixture: Ten grams of phenyl ether-biphenyl mixture is prepared by dissolving 2.65 grams solid biphenyl in 7.35 grams melted phenyl ether. (Liquid eutectic)
- 7.5 Prepurified hydrogen.
- 7.6 Filtered compressed air.
- 7.7 Purified nitrogen.

8. Procedure

- 8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.
- 8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative silica gel tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.
- 8.3 Collection and Shipping of Samples
- 8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).
- 8.3.2 The smaller section of silica gel is used as a back-up and should be positioned nearest the sampling pump.
- 8.3.3 The silica gel tube should be placed in a vertical direction during sampling to minimize channeling through the silica gel.
- 8.3.4 Air being sampled should not be passed through any hose or tubing before entering the silica gel tube.
- 8.3.5 A maximum sample size of 10 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.
- 8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

- 8.3.7 The silica gel tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.
- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the silica gel tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each silica gel tube is scored with a file in front of the first section of silica gel and broken open. The glass wool is removed and discarded. The silica gel in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of benzene is pipetted into each sample container. (All work with benzene should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. It is necessary to be able to measure the phenyl ether and biphenyl separately. Using the following operating conditions, biphenyl and phenyl ether are observed as separate chromatographic peaks.
1. 50 ml/min (60 psig) nitrogen carrier gas flow
 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 3. 500 ml/min (50 psig) air flow to detector
 4. 200°C injector temperature
 5. 250°C manifold temperature (detector)
 6. 140°C column temperature

8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

8.4.5 Measurement of area. The areas of the sample peaks for both components are measured separately by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from standard curves prepared as discussed below.

8.5 Determination of Desorption Efficiency

8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of silica gel to another. Thus, it is necessary to determine at least once the percentage of the specific compounds that are removed in the desorption process, provided the same batch of silica gel is used.

8.5.2 Procedure for determining desorption efficiency. Silica gel equivalent to the amount in the first section of the sampling tube (approximately 150 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This silica gel must be from the same batch as that used in obtaining the samples and can be obtained from unused silica gel tubes. The open end is capped with Parafilm. A known amount of hexane solution of the phenyl ether-biphenyl mixture (Section 7.4) containing 17.5 mg/ml is injected directly into the silica gel with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 10 liter air sample at the selected level.

Six tubes at each of three levels (using 2 microliters, 4 microliters, and 8 microliters of hexane solution for the 0.5X, 1X, and 2X OSHA levels) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the silica gel. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of benzene with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The weight (mg) added for each component is calculated from the amount of solution added and the weight percent of that component in the solution. For instance, addition of an 8-microliter aliquot of the hexane solution of 17.5 mg/ml to 0.5 ml benzene results in a standard containing 0.102 mg phenyl ether/0.5 ml benzene and 0.037 mg biphenyl/0.5 ml benzene.

The desorption efficiencies (D.E.) for phenyl ether and biphenyl are determined separately. For each compound, the desorption efficiency equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the silica gel. Plot the desorption efficiency versus weight of analyte found for the two components separately. These curves are used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml benzene, because samples are desorbed in this amount of benzene. The mixture of phenyl ether-biphenyl described in Section 7.4 is used for the standards. The density of the mixture (1.06 g/ml at 25°C), as given in Reference 11.4, is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, are prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. A curve for each component is established by plotting concentration in mg/0.5 ml for that component versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the

sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

The concentration in mg/cu m in air is determined for phenyl ether and biphenyl separately. The concentration of the mixture is the sum of these two concentrations.

10.1 Read the weight, in mg, corresponding to the peak area for each component from the appropriate standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml benzene and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

$$\text{mg sample} = \text{mg found in front section of sample tube}$$

$$\text{mg blank} = \text{mg found in front section of blank tube}$$

A similar procedure is followed for the backup sections.

10.3 For each component, add the weights found in the front and backup sections to get the total weight of that component in the sample.

10.4 For each component, read the desorption efficiency from the curve (See Section 8.5.2) for the amount of that component found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of each component in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

The concentration of the mixture in air is the sum of the concentrations of the two components:

$$\text{mg/cu m (mixture)} = \text{mg/cu m (phenyl ether)} + \text{mg/cu m (biphenyl)}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

- 11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).
- 11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.
- 11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.
- 11.4 Hake, C.L. and Rowe, V.K., "Phenyl Ether-Biphenyl Eutectic Mixture (Dowtherm A)," Industrial Hygiene & Toxicology, Volume II, 1701, (1962).

Phenyl Glycidyl Ether

Analyte:	Phenyl Glycidyl Ether	Method No.: S74
Matrix:	Air	Range: 31-121 mg/cu m
OSHA Standard:	10 ppm (60 mg/cu m)	Precision (\overline{CV}_T): 0.057
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 31-121 mg/cu m at an atmospheric temperature and pressure of 22°C and 766 mm Hg, using a 50 liter sample. Under the conditions of sample size (50 liters) the probable useful range of this method is 6-180 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 9 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of phenyl glycidyl ether and other substances in the air. The first section of the charcoal tube was found to hold at least 25 mg of phenyl glycidyl ether when a

test atmosphere containing 112 mg/cu m of phenyl glycidyl ether in air was sampled at 0.93 liter per minute for 240 minutes; at that time the concentration of analyte in the effluent was less than 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation ($\overline{CV_T}$) for the total analytical and sampling method in the range of 31-121 mg/cu m was 0.057. This value corresponds to a 3.4 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 8.8% lower than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.2)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (10-ft x 1/8-in stainless steel) packed with 10% FFAP on 80/100 mesh, acid washed DMCS Chromosorb W.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 Phenyl Glycidyl Ether, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 50 liters is recommended. Sample at a flow of 1.0 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.
- 8.4 Analysis of Samples
- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
1. 50 ml/min (60 psig) nitrogen carrier gas flow
 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 3. 500 ml/min (50 psig) air flow to detector
 4. 230°C injector temperature
 5. 265°C manifold temperature (detector)
 6. 90°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 50 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).

11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.

Propylene Oxide

Analyte:	Propylene Oxide	Method No.: S75
Matrix:	Air	Range: 121-482 mg/cu m
OSHA Standard:	100 ppm (240 mg/cu m)	Precision (\overline{CV}_T): 0.085
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 121-482 mg/cu m at an atmospheric temperature and pressure of 24°C and 766 mm Hg, using a 5 liter sample. Under the conditions of sample size (5 liters) the probable useful range of this method is 25-720 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 3.5 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of propylene oxide and other substances in the air. The first section of the charcoal tube was found to hold 4 mg of propylene oxide when a test atmosphere containing

484 mg/cu m of propylene oxide in air was sampled at 0.185 liter per minute for 45 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 121-482 mg/cu m was 0.085. This value corresponds to a 20 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 5.6% lower than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (4-ft x 1/4-in stainless steel) packed with 50/80 mesh Porapak, Type Q.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 Propylene Oxide, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 5 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 190°C injector temperature
 - 5. 255°C manifold temperature (detector)
 - 6. 145°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of hexane solution of propylene oxide containing 300 mg/ml is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 5 liter air sample at the selected level.

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

- 10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

- 11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).
- 11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.
- 11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

Tetrahydrofuran

Analyte:	Tetrahydrofuran	Method No.: S78
Matrix:	Air	Range: 323-1240 mg/cu m
OSHA Standard:	200 ppm (590 mg/cu m)	Precision (\overline{CV}_T): 0.055
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date: 2/14/75

1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container, and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from the injection of standards.

2. Range and Sensitivity

- 2.1 This method was validated over the range of 323-1240 mg/cu m at an atmospheric temperature and pressure of 22.5°C and 763 mm Hg, using a 9 liter sample. Under the conditions of sample size (9 liters) the probable useful range of this method is 60-1480 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 13 mg sample. The method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of tetrahydrofuran and other substances in the air. The first section of the charcoal tube was found to hold 15 mg of tetrahydrofuran when a test atmosphere containing

1140 mg/cu m of tetrahydrofuran in air was sampled at 0.185 liter per minute for 72 minutes; breakthrough was observed at this time, i.e. the concentration of analyte in the effluent was 5% of that in the influent. (The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

4. Precision and Accuracy

- 4.1 The Coefficient of Variation (\overline{CV}_T) for the total analytical and sampling method in the range of 323-1240 mg/cu m was 0.055. This value corresponds to a 32 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the values obtained using the overall sampling and analytical method were 1.3% lower than the "true" values at the OSHA standard level.

5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method.

The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within $\pm 5\%$ at the recommended flow rate. (Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame sealed, 7 cm long with a 6-mm O.D. and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (4-ft X 1/4-in stainless steel) packed with 50/80 mesh Porapak, Type Q.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 One-milliliter sample containers with glass stoppers or Teflon-lined caps.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 0.5-ml delivery pipets or 1.0-ml type graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 Tetrahydrofuran, reagent grade.

7.3 Purified nitrogen.

7.4 Prepurified hydrogen.

7.5 Filtered compressed air.

8. Procedure

8.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a back-up and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A maximum sample size of 9 liters is recommended. Sample at a flow of 0.20 liter per minute or less. The flow rate should be known with an accuracy of at least $\pm 5\%$.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.

- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 1-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 0.5 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
 - 1. 50 ml/min (60 psig) nitrogen carrier gas flow
 - 2. 65 ml/min (24 psig) hydrogen gas flow to detector
 - 3. 500 ml/min (50 psig) air flow to detector
 - 4. 190°C injector temperature
 - 5. 250°C manifold temperature (detector)
 - 6. 185°C column temperature
- 8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation

within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected.

- 8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

8.5 Determination of Desorption Efficiency

- 8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.
- 8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm.

The amount injected is equivalent to that present in a 9 liter air sample at the selected level.

100<

Six tubes at each of three levels (0.5X, 1X, and 2X of the standard) are prepared in this manner and allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 0.5 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg/0.5 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown samples. Curves are established by plotting concentration in mg/0.5 ml versus peak area. Note: Since no internal standard is used in the method, standard solutions must be analyzed at the same time that the sample analysis is done. This will minimize the effect of known day-to-day variations and variations during the same day of the FID response.

10. Calculations

10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg/0.5 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

$$\text{mg sample} = \text{mg found in front section of sample tube}$$

$$\text{mg blank} = \text{mg found in front section of blank tube}$$

A similar procedure is followed for the backup sections.

10.3 Add the weights found in the front and backup sections to get the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg/cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liter/cu m)}}{\text{Air volume sampled (liter)}}$$

10.6 Another method of expressing concentration is ppm.

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{T + 273}{298}$$

where:

P = pressure (mm Hg) of air sampled
T = temperature (°C) of air sampled
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg
M.W. = molecular weight (g/mole) of analyte
760 = standard pressure (mm Hg)
298 = standard temperature (°K)

11. References

- 11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).
- 11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.
- 11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.

E _____
 R _____
 S _____
 UP ☐ 2-UP ☐

MACHINE NO. _____
 METER FINISH _____
 METER START 316025
 METER TOTAL _____

NTIS ☐ .8929
 CENSUS ☐ .8325

ADMIN ☐ .0901
 OTHER ☐

SESSION NUMBER	PLATES	RE- JECTS	RE- MAKES	ACCESSION NUMBER	PLATES	RE- JECTS	RE- MAKES
				16			
				17			
				18			
				19			
				20			
				21			
				22			
				23			
				24			
				25			
				26			
				27			
				28			
				29			
				30			
Sub-Total				TOTALS			
SUS:				SUPERVISOR'S SIGNATURE			

OP/MRG
 TEST FORM #24
 5-71

DATE _____
 OPERATOR _____
 HOURS _____
 1-UP ☐ 2-UP ☐

MACHINE NO. _____
 METER FINISH _____
 METER START 316025
 METER TOTAL _____

NTIS ☐ .0920 ADMIN ☐ .0901
 CENSUS ☐ .0325 OTHER ☐

ACCESSION NUMBER	PLATES	RE- JECTS	RE- MAKES	ACCESSION NUMBER	PLATES	RE- JECTS	RE- MAKES
1				16			
2				17			
3				18			
4				19			
5				20			
6				21			
7				22			
8				23			
9				24			
10				25			
11				26			
12				27			
13				28			
14				29			
15				30			
Sub-Total				TOTALS			

REMARKS:

SUPERVISOR'S
SIGNATURE

OP/MRG
 TEST FORM #24
 5-71

