

This article was downloaded by: [CDC]

On: 26 June 2012, At: 09:58

Publisher: Taylor & Francis

Informa Ltd Registered in England and Wales Registered Number: 1072954 Registered office: Mortimer House, 37-41 Mortimer Street, London W1T 3JH, UK



Applied Occupational and Environmental Hygiene

Publication details, including instructions for authors and subscription information:

<http://www.tandfonline.com/loi/uaoh20>

ELPAT Program

Curtis A. Esche Column Editor & Jensen H. Groff Column Editor

Available online: 25 Feb 2011

To cite this article: Curtis A. Esche Column Editor & Jensen H. Groff Column Editor (1997): ELPAT Program, Applied Occupational and Environmental Hygiene, 12:11, 716-720

To link to this article: <http://dx.doi.org/10.1080/1047322X.1997.10387752>

PLEASE SCROLL DOWN FOR ARTICLE

Full terms and conditions of use: <http://www.tandfonline.com/page/terms-and-conditions>

This article may be used for research, teaching, and private study purposes. Any substantial or systematic reproduction, redistribution, reselling, loan, sub-licensing, systematic supply, or distribution in any form to anyone is expressly forbidden.

The publisher does not give any warranty express or implied or make any representation that the contents will be complete or accurate or up to date. The accuracy of any instructions, formulae, and drug doses should be independently verified with primary sources. The publisher shall not be liable for any loss, actions, claims, proceedings, demand, or costs or damages whatsoever or howsoever caused arising directly or indirectly in connection with or arising out of the use of this material.

ELPAT Program Background and Current Status

Curtis A. Esche and Jensen H. Groff, Column Editors

Introduction

The Environmental Lead Proficiency Analytical Testing (ELPAT) Program is administered by the American Industrial Hygiene Association (AIHA), in cooperation with researchers at the Centers for Disease Control and Prevention (CDC), National Institute for Occupational Safety and Health (NIOSH), and the U.S. Environmental Protection Agency (EPA), Office of Pollution Prevention and Toxics to evaluate and improve the performance of laboratories conducting analyses associated with lead abatement.^(1,2) Proficiency test samples are prepared by an AIHA contractor, Research Triangle Institute (RTI), using real-world paint chips, dusts, and soils. Quarterly samples are sent to participating laboratories by RTI and the performance of the laboratories is evaluated at NIOSH with sufficient time for laboratories to obtain repeat samples and to correct analytical problems before the next round of samples is sent.

The ELPAT Program is open to all interested laboratories, including laboratories outside the United States, laboratories seeking accreditation by various private or state laboratory accreditation systems, laboratories that do not intend to seek laboratory accreditation, and laboratories conducting analyses at permanent, fixed locations, in self-contained mobile facilities, and at temporary locations (e.g., abatement sites). The ELPAT Program is part of an EPA program, the National Lead Laboratory Accreditation Program (NLLAP), to recognize private and state laboratory accreditation systems.⁽³⁾ U.S. Department of Housing and Urban Development (HUD) guidelines for the evaluation and control of lead-based paint hazards in housing⁽⁴⁾ require the use of NLLAP-recognized laboratories to ensure the consistency and quality of measurements of lead in paints, soils, and dusts. NLLAP requirements include successful participation in the ELPAT Program for EPA recognition of accreditation. Two organizations, the American Association for Laboratory Accreditation (A2LA)⁽⁵⁾ and AIHA,⁽⁶⁾ are recognized as accrediting organizations under NLLAP and have in place environmental lead lab-

oratory accreditation systems. Each of these accreditation systems requires participation in ELPAT for environment lead analysis of paint chips, dusts, and soils. Information on specific A2LA or AIHA laboratory accreditation requirements can be obtained from A2LA and AIHA at the addresses listed at the end of this column.

ELPAT Performance Evaluation

The evaluation of the individual laboratories in the ELPAT Program is based on consensus values from reference laboratories and is modeled after the evaluation procedures currently used in an industrial hygiene proficiency testing program, the Proficiency Analytical Testing (PAT) Program.⁽⁷⁾ Reference laboratories are preselected to provide the performance limits for each sample. These laboratories must meet the following criteria: the laboratory was proficient in the previous ELPAT round for paint chips, soils, and dust wipes; and the laboratory is accredited by an EPA NLLAP-recognized accrediting organization.

After data from reference laboratories are collected and extreme reference laboratory data have been statistically treated, the mean ± 3 standard deviations of the treated reference laboratory data becomes the acceptable performance range. Laboratory results are acceptable if they fall within the performance limits. Results falling outside the performance limits are designated as outliers. This is the same criterion used by NIOSH to establish acceptable and outlier performance of industrial hygiene laboratories in the PAT Program.⁽⁷⁾

Laboratories are rated based on performance in the ELPAT Program over the last year (i.e., four rounds) for each lead matrix—paint chips, soil, and dust wipes. The laboratory is proficient for each lead matrix if:

1. all four results have been reported and all are designated as acceptable for the last two consecutive rounds; or
2. three-fourths or more of the results reported in the last four consecutive rounds are designated as acceptable.

However, if a laboratory does not report values for the lead matrix on the round being evaluated, the laboratory is not rated.

ELPAT Round 19, May 1997

Paint samples for round 19 were prepared from paint chips collected from a variety of sites in North Carolina and Ohio, including a school and a hospital. The chips were ground to a maximum particle size of 120 μm .

Soil samples came from driplines around North Carolina residences. Soil samples were dried, sterilized by heating the soil to 325°F for a minimum of 2 hours, and finally sieved to a maximum particle size of 150 μm .

Round 19 dust wipes were prepared from dust collected from households in North Carolina and Milwaukee, Wisconsin. Following sterilization by gamma-irradiation, the household and post-abatement dust was sieved to 150 μm and gravimetrically loaded on a premoistened PaceWipeTM, which has been shown to meet ASTM E 1792⁽⁸⁾ specifications. The loaded wipes were stored under refrigeration until shipment as an antimicrobial measure. Also, it was recommended that dust wipe samples be refrigerated until the laboratory analyses were performed, as an additional precaution to reduce/prevent the growth of mold.

A total of 379 laboratories were enrolled for the nineteenth round of the ELPAT Program, with 362 laboratories (96%) submitting results. Table 1 lists summary statistics of reference laboratories for each matrix and sample number. Agreement among reference laboratories using a variety of sample preparation techniques and analytical methods is demonstrated by relative standard deviations (RSDs) ranging from 6.1 to 9.5 percent for paint chips, 5.1 to 12.0 percent for soils, and 6.5 to 10.5 percent for dust wipes. The RSDs are similar to the findings on previous ELPAT rounds.

Table 2 shows the number of all participating laboratory analyses that were identified as outliers. The percentage of all participating laboratory analyses that were identified as outliers was less than

TABLE 1. ELPAT Program Summary Statistics of Reference Laboratories for Round 019

Sample Type	Sample	N	Mean	Minimum	Maximum	STD	RSD (%)	Acceptable Range
Paint chips (%)	1	105	0.7928	0.692	0.8859	.052	6.5	0.6379–0.9476
	2	105	5.1107	4.37	5.82	.403	7.9	3.902–6.3193
	3	105	0.202	0.1615	0.234	.019	9.5	0.144–0.2595
	4	105	0.7939	0.7046	0.8762	.048	6.1	0.649–0.9388
Soil (mg/kg)	1	105	677.6	610.3	737	36.3	5.4	568.6–786.6
	2	105	1783	1583	1945	101	5.6	1481–2085
	3	105	58.3	48.7	73.2	6.97	12.0	37.3–79.2
	4	105	1613.8	1450.2	1740	81.6	5.1	1369–1858.6
Dust wipes (μg)	1	105	80.2	62	93	8.40	10.5	54.9–105.4
	2	105	1498.5	1263.2	1635	98.0	6.5	1204.5–1792.4
	3	105	287.1	226.3	325.8	25.0	8.7	212–362.2
	4	105	127.1	104	146	12.4	9.7	89.9–164.3

8.8 percent (3.4 to 7.6% for paint chips, 5.2 to 8.8% for soils, and 2.5 to 5.8% for dust wipes). This is similar to the frequency of outliers reported on the earlier rounds of ELPAT for each matrix.

Table 3 shows a summary of acceptable results for the three lead matrices by sample preparation technique and instrumental method used by participating laboratories. Analytical methods that were not identified by laboratories were omitted from the table. Sample extraction techniques are grouped into hotplate, microwave, and other techniques reported by participants. Hotplate digestion categories are: NIOSH 7082/7105 (a nitric acid/hydrogen peroxide digestion method modified from the *NIOSH Manual of Analytical Methods*, Method 7082⁽⁹⁾, EPA SW846–3050A⁽¹⁰⁾ (an EPA nitric acid/hydrogen peroxide method), American Society for Testing Materials (ASTM) hotplate methods,^(11,12) and other hotplate techniques. Microwave digestion categories are: EPA SW846–3051⁽¹³⁾ (a nitric acid digestion method),

ASTM microwave methods,^(14,15) and other microwave techniques. The “other” category includes nonmicrowave and nonhotplate techniques, such as X-ray fluorescence sample preparation, leaching techniques, ultrasonic extraction, and Parr bomb. Instrumental methods are categorized into flame atomic absorption (FAA), inductively coupled plasma-atomic emission spectroscopy (ICP-AES), graphite furnace atomic absorption (GFAA), laboratory X-ray fluorescence (LAB-XRF), anodic stripping voltammetry (ASV), and others, which includes inductively coupled plasma-mass spectroscopy.

ELPAT Round 19 Statistical Analysis

Statistical tests were performed on the ELPAT round 19 data as previously described by Schlecht *et al.*⁽¹⁶⁾ The purpose of these statistical tests is to detect bias differences among analytical methods. Biases on round 19 were found in the instrumental methods category. Two-way analysis of variance procedures

found statistically significant biases for dust wipe samples 1 and 4. For dust wipe sample 1 data, differences among instrumental methods were detected at the $p = 0.005$ level. FAA had a positive bias over ICP-AES and ASV, with a maximum difference relative to the mean of 7.5 percent. For dust wipe sample 4 data, differences among instrumental methods were detected at the $p = 0.003$ level. FAA and ASV had a positive bias over ICP-AES, with a maximum difference relative to the mean of 5.5 percent. The positive bias observed with the ASV result may be due to the ultrasonic extraction procedure and not to the analytical measurement technique.

NIOSH ELPAT bias studies have found evidence of bias among the principal instrumental methods used by participating laboratories for all three matrices: paint chips, soils, and dust wipes. The biases range from 2 to 26 percent of the corresponding reference laboratory mean, with the largest biases occurring at low lead levels for dust wipes, generally

TABLE 2. ELPAT Program Summary of Performance: All Laboratories Participated for Round 019

Sample Type	Sample No.	No. of Labs Rated	Acceptable Labs	Low Outlier	High Outlier
Paint chips (%)	1	355	343	7	5
	2	355	334	12	9
	3	355	339	7	9
	4	355	328	17	10
Soil (mg/kg)	1	308	292	6	10
	2	308	290	12	6
	3	308	281	7	20
	4	308	288	15	5
Dust wipes (μg)	1	327	319	3	5
	2	327	309	12	6
	3	327	308	14	5
	4	327	318	6	3

TABLE 3. ELPAT Program Lab Performance Summary for Round 019

Instrument	Digestion	Method	Paint Chips (%)		Soil (mg/kg)		Dust Wipes (μg)	
			Acceptable (%)	Failures (%)	Acceptable (%)	Failures (%)	Acceptable (%)	Failures (%)
FAA	Hotplate	NIOSH-7082/7105	98	2	93	7	96	4
		EPA-SW846-3050A	95	5	93	7	97	3
	Microwave	EPA-SW846-3051	96	4	94	6	98	2
		Ultrasonic	92	8	100	0	100	0
GFAA	Hotplate	NIOSH-7082/7105	75	25	0	0	100	0
		EPA-SW846-3050A	100	0	100	0	100	0
	Microwave	EPA-SW846-3051	58	42	58	42	88	13
		Ultrasonic	0	0	100	0	100	0
ICP-AES	Hotplate	NIOSH-7082/7105	93	8	88	13	92	8
		EPA-SW846-3050A	94	6	95	5	98	2
	Microwave	EPA-SW846-3051	98	2	94	6	91	9
		Ultrasonic	100	0	75	25	100	0
LAB-XRF	XRF Sample Prep		100	0	100	0	0	0
ASV	Ultrasonic		91	9	75	25	88	13
Total			95	5	93	7	96	4

well below HUD and EPA lead action levels. Although it was expected that differences among sample preparation techniques would be found, NIOSH ELPAT bias studies have found no conclusive evidence of bias among the principal sample preparation techniques used by participating laboratories.⁽¹⁶⁾

The results of NIOSH ELPAT bias studies are consistent with the 3 to 18 percent bias found by RTI in an EPA-sponsored collaborative test. In the EPA collaborative test, RTI followed up with participating laboratories and determined that some FAA laboratories failed to perform background corrections, which one would expect to result in a positive bias; also, some ICP-AES laboratories failed to take matrix effects into account, which one would expect to result in a negative bias. NIOSH does not follow up with participating laboratories to determine if each participating ELPAT laboratory has performed all of the steps of the analytical method reported by the laboratory. However, NIOSH has advised both co-operating accrediting organizations that ELPAT bias could be the result of some ELPAT laboratories not following all steps of the analytical method. NIOSH has recommended that accrediting organizations emphasize FAA background correction and ICP-AES matrix effect minimization procedures when evaluating laboratory accreditation applications and in conducting on-site assessments for EPA NLLAP recognition. Laboratories

should refer to the RTI collaborative test for a more complete discussion on how bias can be minimized.⁽¹⁷⁾

Laboratory studies of field portable methods such as ultrasonic extraction and ASV of lead from environmental samples show promise as viable techniques. For a more complete discussion, laboratories can refer to a NIOSH study comparing ultrasonic extraction to hotplate and microwave digestion and field portable ASV to laboratory-based FAA on a series of laboratory-generated air samples and National Institute of Standards and Technology standard reference materials.⁽¹⁸⁾

Lead Reference Materials

The ELPAT Program is designed to supplement, but not replace, a laboratory's internal quality control program. Use of materials of known lead content in suitable matrices is important in obtaining accurate and reliable lead results. Such materials should be used to validate methods when sample preparation techniques or instrumental methods are adopted or modified. In addition, the materials should be used for daily quality control charting of laboratory/analyst performance. A2LA and AIHA should be contacted about potential sources of primary and secondary reference materials.

EPA NLLAP

Under Title X of the Housing and Community Development Act of 1992, EPA, in consultation with the Department of

Health and Human Services, has the responsibility to review and determine if effective voluntary laboratory accreditation systems are in place. If EPA determines that effective voluntary laboratory accreditation systems are not in place, EPA is responsible to establish a federal laboratory certification system.⁽¹⁹⁾

The EPA has established an NLLAP to recognize laboratories performing analysis associated with lead abatement. Published HUD guidelines for the evaluation and control of lead-based paint hazards in housing require the use of NLLAP-recognized laboratories to ensure the consistency and quality of measurements of lead in paints, soils, and dusts.⁽⁴⁾ NLLAP recognition of laboratories analyzing lead in paint chips, soils, and dusts has two requirements: (1) successful participation in proficiency testing using real-world matrices; and (2) laboratory accreditation including on-site assessment of laboratory operations. NLLAP requirements are based on the recommendations of a Federal Interagency Taskforce on Lead Based paint, a group of 17 federal agencies involved with lead issues, that recognition should be based on both proficiency testing and laboratory accreditation.⁽²⁰⁾ Similarly, proficiency testing and laboratory accreditation requirements were also part of the recommendations for environmental laboratories of a 1991 National Conference on Laboratory Issues in Childhood Lead Poisoning Prevention sponsored by the Association of State and

Territorial Public Health Laboratory Directors, the CDC, and EPA.

Laboratory accreditation takes some time to achieve. Laboratory accreditation involves submittal of a description of a laboratory's quality system and manual to the accrediting organization and the on-site evaluation by NLLAP-qualified assessors of laboratory operations, including equipment, facilities, analytical methods, staff, and internal quality control. Laboratories interested in obtaining accreditation information such as the program requirements, time needed to complete the process, and cost should contact the recognized laboratory accreditation organizations. If other laboratory accreditation organizations are recognized, this information will be included in subsequent ELPAT columns.

Lists of laboratories that have performed successfully and are accredited in the ELPAT Program are provided upon request to the public by the Lead Information Clearinghouse (1-800-424-LEAD). The ELPAT proficiency testing program is open to all interested laboratories. This means laboratories outside the United States and laboratories that do not wish to be accredited can continue to participate in ELPAT. However, only accredited laboratories will appear on the NLLAP list provided by the Lead Information Clearinghouse.

Disclaimer

Mention of company names or products does not constitute endorsement by the CDC.

Information

A2LA laboratory accreditation and seminars on environmental lead laboratory accreditation:

American Association for Laboratory Accreditation (A2LA)
656 Quince Orchard Road
Gaithersburg, MD 20878
Phone: (301) 670-1377
FAX: (301) 869-1495

AIHA laboratory accreditation, ELPAT Program information, ELPAT sample orders, and seminars on environmental lead laboratory accreditation:

ELPAT Coordinator
American Industrial Hygiene Association (AIHA)
2700 Prosperity Avenue, Suite #250

Fairfax, VA 22031
Phone: (703) 849-8888
FAX: (703) 207-3561

References

1. American Industrial Hygiene Association and National Institute for Occupational Safety and Health: Cooperative Research and Development Agreement (CRADA) Between the American Industrial Hygiene Association and the National Institute for Occupational Safety and Health. CRADA NIO.C92.001.00. AIHA, Fairfax, VA; and U.S. Department of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention, NIOSH, Cincinnati, OH (1992).
2. U.S. Environmental Protection Agency and National Institute for Occupational Safety and Health; Memorandum of Understanding, project officers J.J. Breen (EPA), J.V. Scalera (EPA), and P. Schlecht (NIOSH). MOU No. PW75935570-01-0. U.S. EPA, Office of Pollution Prevention and Toxics, Washington, DC; and Department of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention, NIOSH, Atlanta, GA (1992).
3. National Lead Laboratory Accreditation Program: Model Memorandum of Understanding (MOU), collaborative effort among U.S. Environmental Protection Agency, Office of Pollution Prevention and Toxics; Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health, Division of Physical Sciences and Engineering; and Accreditation Organizations (1993).
4. U.S. Department of Housing and Urban Development: Guidelines for the Evaluation and Control of Lead-Based Paint Hazards in Housing. U.S. HUD, Washington, DC (1995).
5. American Association for Laboratory Accreditation: A2LA Environmental Lead Program Requirements. A2LA, Gaithersburg, MD (1992).
6. American Industrial Hygiene Association: AIHA Environmental Lead Program Policies. AIHA, Fairfax, VA (1992).
7. National Institute for Occupational Safety and Health: Laboratory Evaluations and Performance Reports for the Proficiency Analytical Testing (PAT) and Environmental Lead Proficiency Analytical Testing (ELPAT) Programs. DHHS (NIOSH) Pub. No. 95-104. U.S. Department of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention, NIOSH, Cincinnati, OH (1995).
8. American Society for Testing and Materials: ASTM E 1792, Standard Specification for Wipe Sampling Materials for Lead in Surface Dust. In: Annual Book of ASTM Standards. ASTM, West Conshohocken, PA (1996).
9. National Institute for Occupational Safety and Health: Method 7082. In: NIOSH Manual of Analytical Methods, 4th ed. DHHS (NIOSH) Pub. No. 94-113. P.M. Eller, Ed. U.S. Department of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention, NIOSH, Cincinnati, OH (1994).
10. U.S. Environmental Protection Agency: Method SW846-3050A, Acid Digestion of Sediments, Sludges and Soils (Metals). In: U.S. EPA Test Methods for Evaluating Solid Waste, Physical/Chemical Method SW846, 3d rev ed. EPA, Washington, DC (1990).
11. American Society of Testing and Materials: ASTM E 1644, Standard Practice for Hot Plate Digestion of Dust Wipe Samples for Determination of Lead by Atomic Spectrometry. In: Annual Book of ASTM Standards. ASTM, Philadelphia, PA (1995).
12. American Society for Testing and Materials: ASTM E 1726, Standard Practice for the Sample Digestion of Soils for the Determination of Lead by Atomic Spectrometry. In: Annual Book of ASTM Standards. ASTM, Philadelphia, PA (1995).
13. U.S. Environmental Protection Agency: Method SW876-3051, Microwave Assisted Acid Digestion of Sediments, Sludges, Soils and Oils (Metals). In: U.S. EPA Test Methods for Evaluating Solid Waste, Physical/Chemical Method SW846, 3d rev ed. EPA, Washington, DC (1990).
14. American Society for Testing and Materials: ASTM E 1645, Standard Practice for Preparation of Dried Paint Samples for Subsequent Lead Analysis by Atomic Spectrometry. In: Annual Book of ASTM Standards. ASTM, Philadelphia, PA (1995).
15. American Society for Testing and Materials: ASTM E 1741, Standard Practice for Preparation of Airborne Particulate Lead Samples Collected During Abatement and Construction Activities for Subsequent Lead Analysis by Atomic Spectrometry. In: Annual Book of ASTM Standards. ASTM, Philadelphia, PA (1996).
16. Schlecht P.C., Groff J.H., Feng A., Song R.: Laboratory and Analytical Method Performance of Lead Measurements in Paint Chips, Soils, and Dusts. Am. Ind. Hyg. Assoc. J. 57(11):1035-1043 (1996).
17. Binstock D.A., Hardison D.L., Grohse P.M., Gutknecht W.F.: Standard Oper-

- ating Procedures for Lead In Paint by Hotplate- or Microwave-based Acid Digestion and Atomic Absorption or Inductively Coupled Plasma Emission Spectrometry. EPA 600/8-91/213. U.S. Environmental Protection Agency, Research Triangle Park, NC (1991). Available from NTIS, Springfield, VA, PB 92-114172.
18. Ashley K. Ultrasonic Extraction and Field-Portable Anodic Stripping Voltammetry of Lead from Environmental Samples. *Electroanalysis* 7(12):1189-1192 (1995).
19. United States Code: Title X, Residential Lead-Based Paint Hazard Reduction Act of 1992. 42 USC 4851 (1992). In: Housing and Community Development Act of 1992. Public Law 102-550 (1992).
20. U.S. Environmental Protection Agency: Task Group on Methods and Standards of the Federal Interagency Lead-Based Paint Task Force. In: Laboratory Accreditation Program Guidelines: Measurement of Lead in Paint, Dust and Soil. Final Report, EPA 747-R-92-001. U.S. EPA, Washington, DC (1992).

EDITORIAL NOTE: Curtis A. Esche and Jensen H. Groff are with the Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health.
