


Supplemental Figure 2: Bland-Altman plot of percent difference in RBC folate concentrations to compare microbiologic assay by Immulite 2000 immunoassay



Bland-Altman plot with percent difference. The thin solid line marks the line of identity, the thick solid line marks the bias. The mean percent bias was 24% and the dashed lines mark the 95% limits of agreement (-26% to 75%).