



Published in final edited form as:

Parasitology. 2018 June ; 145(7): 865–870. doi:10.1017/S0031182017001925.

Molecular Detection of *Cyclospora cayetanensis* in Human Stool Specimens using UNEX-based DNA extraction and real-time PCR

Yvonne Qvarnstrom^{1, #}, Theresa Benedict¹, Paula L. Marcet¹, Ryan E. Wiegand¹, Barbara L. Herwaldt¹, and Alexandre J. da Silva²

¹Division of Parasitic Diseases and Malaria, Center for Global Health, Centers for Disease Control and Prevention, 1600 Clifton Rd, Atlanta, GA 30329, USA

²Office of Applied Research and Safety Assessment, Center for Food Safety and Applied Nutrition, U.S. Food and Drug Administration, 8301 Muirkirk Rd, Laurel, MD 20708, USA

SUMMARY

Cyclospora cayetanensis is a coccidian parasite associated with diarrheal illness. In the United States, foodborne outbreaks of cyclosporiasis have been documented almost every year since the mid-1990s. The typical approach used to identify this parasite in human stools is examination of acid-fast-stained smears under bright-field microscopy. UV fluorescence microscopy of wet mounts is more sensitive and specific than acid-fast staining but requires a fluorescence microscope with a special filter not commonly available in diagnostic laboratories. In this study, we evaluated a new DNA extraction method based on the Universal Nucleic Acid Extraction (UNEX) buffer and compared the performances of four published real-time PCR assays for the specific detection of *C. cayetanensis* in stool. The UNEX-based method had an improved capability to recover DNA from oocysts compared with the FastDNA stool extraction method. The best-performing real-time PCR assay was a *C. cayetanensis*-specific TaqMan PCR that targets the 18S ribosomal RNA gene. This new testing algorithm should be useful for detection of *C. cayetanensis* in human stool samples.

Keywords

Cyclospora cayetanensis; stool DNA extraction; coccidia; molecular diagnostics

INTRODUCTION

Cyclospora cayetanensis is a coccidian parasite associated with the diarrheal illness cyclosporiasis. The majority of reported U.S. cases of cyclosporiasis have been associated with foodborne outbreaks or with international travel to tropical or subtropical areas (Centers for Disease Control and Prevention, 2004; Centers for Disease Control and Prevention, 2013; Hall *et al.*, 2011; Hall *et al.*, 2012; Herwaldt, 2000; Ho *et al.*, 2002).

[#]Address correspondence to: Yvonne Qvarnstrom, Centers for Disease Control and Prevention, MS D64, 1600 Clifton Rd, Atlanta, GA 30329, USA, bvp2@cdc.gov.

Symptoms can persist from weeks to months if the infection is not diagnosed and treated (Herwaldt, 2000).

Laboratory diagnosis of *Cyclospora* infections relies on detecting the oocysts shed in the stool of infected patients. *Cyclospora* oocysts have environmentally resistant outer cell walls that make them acid-fast and autofluorescent under UV light (Erickson & Ortega, 2006). Examination of acid-fast-stained stool smears under bright-field microscopy is the typical approach used by clinical laboratories to identify *Cyclospora* in stools, but this method has suboptimal sensitivity and specificity. UV fluorescence microscopy of wet mounts is an alternative, more accurate approach for diagnosing *Cyclospora* infection (Berlin *et al.*, 1998). However, a UV excitation filter set that is not commonly available in clinical laboratories is needed for this procedure; with the preferred UV filter set (330 to 365 nm), intense blue fluorescence of *Cyclospora* oocysts is obtained. (Berlin *et al.*, 1998; Relman *et al.*, 1996). A molecular method such as PCR can provide sensitive and specific detection, as well as species-level identification (Eberhard *et al.*, 1999; Li *et al.*, 2015). Relman and colleagues developed the first-described PCR assay for *C. cayetanensis* (Relman *et al.*, 1996). This nested PCR assay has since been used for various applications, including confirmatory diagnostic testing (Pieniazek *et al.*, 1996). However, the nested format and the requirement for DNA sequencing analysis of the PCR product for species-level identification limit the usefulness of this method.

To date, four real-time PCR assays for detection of *C. cayetanensis* have been described: two utilize a species-specific TaqMan probe to detect unique regions in the small subunit ribosomal RNA (18S rRNA) gene (Varma *et al.*, 2003; Verweij *et al.*, 2003), and two rely on DNA binding dyes and amplicon melt curve analysis for specificity (Lalonde & Gajadhar, 2011; Marangi *et al.*, 2015). Each of these four assays has the potential to identify the parasite to the species level in a one-step reaction. However, for any PCR-based assay to be successful in detecting *C. cayetanensis* in stool, the preceding DNA extraction step must be able to disrupt the tough outer oocyst wall to recover the DNA. The aims of this study were to evaluate a new DNA extraction method, previously used to detect coccidia in experimentally contaminated food (Shields *et al.*, 2013); and to compare the analytical performances of the four published real-time PCR assays for the detection of *C. cayetanensis* in stool.

MATERIALS AND METHODS

Human specimens

We used 137 human stool specimens to evaluate the DNA extraction and PCR methods. The human stool specimens had been sent to the Centers for Disease Control and Prevention (CDC) for diagnostic confirmation or as part of outbreak investigations during 2004–2015 and were used in accordance with the CDC Human Subjects Research protocol entitled “Use of residual diagnostic specimens from humans for laboratory methods research.” The stool samples were unpreserved or had been collected in commonly used fixatives for parasitology, including Zn-PVA, EcoFix, or TotalFix.

Thirty-two of the 137 human stools were classified as positive for *C. cayetanensis*, based on positive UV fluorescence microscopy (n = 30) or DNA sequencing confirmation of PCR products obtained by the nested PCR assay (n = 2) (Berlin *et al.*, 1998; Relman *et al.*, 1996). One of the microscopy-positive specimen, selected because of a large volume (about 100 ml) and sufficiently high concentration of oocysts (an oocyst count of 1.3×10^4 oocysts per ml, as determined using a hemocytometer), was used to assess the analytical sensitivities (detection limits) of the four PCR assays: four aliquots of this stool were serially diluted in parasite-free stool to obtain four separate dilution series down to 1.3 oocysts per ml. The definition of the detection limit was the lowest number of oocysts that an assay detected in all four aliquots.

For specificity analysis, we used 105 of the 137 human stool specimens, 15 stool samples from non-human primates, 3 stool samples from rats, and 2 DNA samples. The human stool specimens were positive by microscopy for either *Entamoeba histolytica/dispar* (n = 24 specimens); *Enterocytozoon/Encephalitozoon* spp. (n = 13); *Giardia duodenalis* (n = 6); *Cryptosporidium* spp. (n = 3); *Blastocystis hominis*, *Dientamoeba fragilis*, or *Iodamoeba butschlii* (each n = 2); or for *Balantidium coli*, *Chilomastix mesnili*, *Entamoeba coli*, *Entamoeba hartmani*, hookworm, or *Trichomonas hominis* (each n = 1); or were negative by microscopy for parasites and microsporidia (n = 47). The non-human primate samples, positive for simian *Cyclospora* spp. as described elsewhere (Eberhard *et al.*, 1999; Eberhard *et al.*, 2001), included eight *Cyclospora papionis*-positive samples from *Papio anubis* in Ethiopia, two *C. papionis*-positive samples from *P. anubis* in Kenya, two *Cyclospora colobi*-positive samples from *Colobus angolensis* in Kenya, and three *Cyclospora cercopithec*-positive samples from *Cercopithecus aethiops* in Kenya. The rat stools (collected during an environmental investigation unrelated to this study) were positive by microscopy for *Eimeria* spp. (n = 2 samples) and *Hymenolepis nana* (n = 1). DNA samples extracted from *Eimeria tenella* and *Eimeria acervulina* were also included.

DNA extraction

Two methods for extraction of total genomic DNA from stool were compared using fifty human stool samples (25 *C. cayetanensis*-positive and 25 specificity controls). Method 1 used a modification of the FastDNA[®] method (da Silva *et al.*, 1999). Aliquots of ~0.3–0.5 ml of each stool were subjected to bead beating in a FastPrep-24 cell disruptor instrument (MP Biomedicals, Santa Ana, CA, USA). An internal DNA quality control plasmid (Duffy *et al.*, 2013), here called pIC, was added immediately after the bead-beating step to allow for detection of DNA extraction failure (see below). Potential inhibitors remaining after the FastDNA extraction were removed by further purification with the QIAquick PCR purification kit, according to the manufacturer's instructions (QIAGEN Inc., Valencia, CA, USA).

Method 2 was based on the commercially available UNEX (Universal Nucleic Acid Extraction) buffer. An aliquot of ~0.5 ml of stool was added to a matrix E bead-beating tube (MP Biomedicals), along with 60 µl of proteinase K (QIAGEN) and 600 µl of UNEX buffer (Phthisis Diagnostics, Charlottesville, VA, USA). The tube was incubated for 15 min at 56°C for protein digestion, followed by disruption in the FastPrep-24 cell disruptor

instrument at a speed of 6.0 m/s for 1 min. One pg of the internal control plasmid pIC was added, and the tube was then centrifuged at maximum speed ($>13,000 \times g$) for 1 min to pellet the debris. The supernatant was passed through a DNA-binding column (DNeasy mini spin column from QIAGEN). After two wash cycles with ethanol-containing wash buffers, the DNA was eluted from the column in 80 μ l of AE buffer (QIAGEN). To purify the eluate further, it was passed through a Zymo-Spin IV-HRC column (Zymo Research Corp., Irvine, CA, USA).

The remaining 87 human stool specimens and the 18 animal stool samples were extracted with the UNEX-based method only, since that method performed better than the Fast DNA method for extraction of *C. cayetanensis* DNA from stool (see results).

Conventional nested PCR for *Cyclospora*

The nested PCR assay contained 0.3 μ M of each primer (Relman *et al.*, 1996) and the AmpliTaq Gold PCR Master Mix (Life Technologies, Grand Island, NY, USA). The total volume was 50 μ l, with 5 μ l of extracted DNA added to the first reaction and 3 μ l of the undiluted product from the first step added to the nested reaction. The cycling parameters were 95°C for 5 min to activate the polymerase, followed by 30 cycles of 95°C for 15 s, 57°C for 15 s, and 72°C for 1 min for the first step; and 35 cycles of 95°C for 15 s, 65°C for 15 s, and 72°C for 1 min for the nested step. The PCR products were visualized on 1.5% agarose gels stained with ethidium bromide.

Real-time PCR for *C. cayetanensis*

We compared four real-time PCR assays: two TaqMan assays we refer to as the Verweij (Verweij *et al.*, 2003) and the Varma (Varma *et al.*, 2003) assays, which target different parts of the 18S rRNA gene; a SYBR Green assay (Lalonde & Gajadhar, 2011) that targets the 18S rRNA gene; and an EvaGreen assay (Marangi *et al.*, 2015) that targets the internal transcribed region 2 (ITS2). The TaqMan real-time PCR assays were performed in a Mx3000P™ thermocycler, whereas the DNA binding dye assays were performed in an AriaMx thermocycler to facilitate high-resolution melt (HRM) analysis. A 2- μ l aliquot of template DNA was added to each reaction.

We used previously described reaction conditions for the EvaGreen assay (Marangi *et al.*, 2015). For the SYBR Green assay, we only used the *Cyclospora*-specific primers (CycloF and CycloR) (Lalonde & Gajadhar, 2011). We performed the Varma assay with the conditions described in the original publication (Varma *et al.*, 2003) and with various modifications in attempts to improve amplification performance. The results presented here were obtained using QuantiTect Probe PCR Master Mix (QIAGEN) instead of AmpliTaq Master Mix with BSA. We improved the specificity of the Verweij assay by making two modifications: 1) increasing the annealing/extension temperature from 60°C to 67°C, and 2) correcting the reverse primer sequence to 5'-AAT GCC ACG GTA GGC CAA TA-3' (the reverse primer sequence in the original publication by Verweij *et al.* was missing a base, in comparison with the target gene). We also replaced Platinum QPCR Supermix (Life Technologies) for the reagents from Eurogentec and decreased the primer and probe concentrations to 0.5 μ M and 0.1 μ M, respectively.

Detection of the internal DNA quality control plasmid (pIC)

The pIC is a recombinant plasmid with the cDNA of an *Arabidopsis thaliana* gene (GenBank accession number NM_114612.3) inserted into vector pZErO-2. The pIC plasmid was linearized by *Pst*I digestion and diluted to 0.1 pg/μl before adding it to stool during DNA extraction. DNA samples that contained the pIC were analyzed in a real-time PCR assay containing 0.1 μM each of the primers IAC Fw and IAC Rv (Duffy *et al.*, 2013) and 0.05 μM of the TaqMan probe ICP2 (5'-HEX-CCACTGCTAAAGGTAGCCCCACGTC-BHQ1-3'), using standard TaqMan cycling structure.

Statistical analysis

The proportions of samples that were correctly identified were tabulated and confidence intervals were calculated using the Clopper-Pearson method (Clopper & Pearson, 1934). We compared the performances of the FastDNA and UNEX-based extraction methods by calculating the mean Ct value for each method. We assumed that the lower the mean Ct value the better the performance of the DNA extraction method. To evaluate operator-associated variability, two laboratorians processed the same specimens independently. In the statistical analyses, the negative samples, which, by definition, were not assigned Ct values, were considered as right-censored outcomes, in an adaptation of a previously described model for left-censored data (Jin *et al.*, 2011). Thus, a model was fit to the data in SAS software version 9.3 (SAS Institute, Inc., Cary, NC, USA) to estimate the mean difference in the Ct values between the two methods, for each laboratorian; an interaction term between the method and laboratorian was included in the model, and a 5% level of significance was used.

The diagnostic performances of the assays were evaluated as described elsewhere (Griner *et al.*, 1981). In brief, sensitivity was calculated as the probability that the result was positive when *C. cayetanensis*-positive specimens were tested, and specificity as the probability that the result was negative when *C. cayetanensis*-negative specimens/samples were tested.

RESULTS

Evaluation of the UNEX-based method

To compare the FastDNA and UNEX-based methods for DNA extraction from stool, two different laboratorians processed 25 *C. cayetanensis*-positive and 25 specificity-control specimens. Both laboratorians performed both methods in parallel (i.e., for a total of 200 DNA extractions) without knowing which specimens were positive or negative. The results are summarized in Table 1. All of the results obtained with the UNEX-based method agreed with the microscopy results (i.e., no false-positive or false-negative results). In contrast, three *C. cayetanensis*-positive samples tested negative for *C. cayetanensis* after DNA extraction with the FastDNA method, even though the pIC was amplified from all three samples. The statistical modeling supported the conclusion that UNEX was better than FastDNA at recovering *Cyclospora* DNA from stool: the estimated mean difference in the Ct values between the two methods was significantly different from zero for both laboratorians, indicating that the UNEX-based method had a lower estimated Ct mean (Table 2).

Furthermore, the FastDNA method had more manual pipetting steps and was therefore more labor-intensive than the UNEX-based method.

Comparison of real-time PCR methods

We compared the performances of the four real-time PCR assays with that of the conventional nested PCR assay using 32 *Cyclospora*-positive human stool specimens and 125 specificity controls (Table 3). The main difference among the assays was in their analytical sensitivities (i.e., detection limits). The EvaGreen assay was the most sensitive, with a detection limit of two oocysts per DNA extraction (i.e., seven oocysts per ml of stool). The Verweij and SYBR Green assays had the same detection limit as the conventional nested assay (i.e., 15 oocysts per DNA extraction, or 50 per ml), whereas the detection limit for the Varma assay was 200 oocysts per sample (~700 oocysts/ml). The Varma assay also failed to detect *C. cayetanensis* DNA in three of the *Cyclospora*-positive human stools, such that its diagnostic sensitivity was 93%.

The specificities were 100% for both the conventional nested PCR and the Varma assays, as neither assay amplified any of the specificity controls. The specificity of the Verweij assay was dependent on the annealing temperature; the assay amplified DNA from several of the simian *Cyclospora* samples when the previously published annealing temperature of 60°C was used (data not shown) but was specific for *C. cayetanensis* when the annealing temperature was increased to 67°C.

The PCR primers used in the SYBR Green assay were originally designed to detect and distinguish multiple coccidian parasites using HRM analysis. In this study, these primers amplified *C. cayetanensis*, simian *Cyclospora* spp., and *Eimeria* spp. DNA. Fig. 1 depicts the results of the HRM analyses for representative study samples. *Eimeria* spp. displayed a distinct melt curve profile with a main peak at ~86°C. *Cyclospora* spp. displayed a melt curve pattern with three peaks. All of the *C. cayetanensis* specimens had one melt peak in the range of 79.2–79.8°C. Whereas most (14 of 16) of the simian samples had a peak in the range of 80.2–80.8°C, the melt peaks were 79.4°C for one *C. papionis* sample and 79.6°C for one *C. cercopithecii* sample. The other two peaks in the three-peak pattern (at ~85.5°C and 88.5°C) overlapped between *C. cayetanensis* and the simian *Cyclospora* samples. Therefore, the SYBR Green assay could not reliably distinguish *C. cayetanensis* from the simian *Cyclospora* spp., resulting in an analytical specificity of 85%.

The EvaGreen assay was designed to specifically amplify *C. cayetanensis* DNA. All of the *C. cayetanensis* specimens in this study displayed a single melt peak at 85.0–85.5°C. Unfortunately, two of the human specificity controls (one stool positive for microsporidia and one parasite/microsporidia-free stool) also produced PCR products with similar melt curves as those of the *C. cayetanensis*-positive specimens (Fig. 2). Thus, the EvaGreen assay had a 98% diagnostic specificity in this study.

DISCUSSION

We have developed a new molecular testing algorithm for the specific detection of *C. cayetanensis* in human stool specimens. The molecular techniques we used previously—the

FastDNA extraction method followed by amplification with the nested PCR assay—had several limitations, including occasional failure to detect *Cyclospora* DNA in microscopy-positive stool, vulnerability to amplicon contamination, and the need to conduct DNA sequencing analysis for species-level identification. The new algorithm consists of an improved DNA extraction method (UNEX), followed by a real-time PCR assay (the Verweij assay) that detects *C. cayetanensis* to the species level. In our comparison of four published real-time PCR assays, the Verweij assay had the overall best performance. The Varma assay had suboptimal sensitivity, and the two DNA binding dye-based assays were less specific than the Verweij assay.

There are few options available for the laboratory diagnosis of *Cyclospora* infection. There is no *in vitro* culture system or animal model to isolate *C. cayetanensis* in the laboratory (Eberhard *et al.*, 2000). No antigen-based or serologic method for human diagnosis is available. PCR-based molecular detection of *C. cayetanensis* in stool can be helpful to confirm the findings obtained by microscopy and to diagnose the infection in persons with low-level shedding of the parasite. To our knowledge, there is only one FDA-cleared test for diagnosis of *C. cayetanensis* infection: the FilmArray Gastrointestinal Panel, a multiplex PCR-based test that simultaneously detects DNA from 22 different pathogens. However, this test requires stool in Cary Blair transport medium and cannot be applied to stools collected in fixatives commonly used in parasitology.

It can be challenging to extract PCR-ready *C. cayetanensis* DNA from stool. Substances in stool can co-purify with the DNA and inhibit amplification in the subsequent PCR reaction (Monteiro *et al.*, 1997). In addition, *C. cayetanensis* oocysts have tough cell walls, with a thick, carbohydrate-rich outer layer, that can make DNA extraction difficult (Erickson & Ortega, 2006). In a published comparison of four DNA extraction methods for coccidian parasites in experimentally contaminated food items, another method that used the UNEX buffer produced DNA extracts with less inhibitory effect on PCR (Shields *et al.*, 2013). In our comparison using stool specimens, the UNEX-based method yielded better results than the FastDNA method, which likely reflects enhanced capability to break open the oocysts.

In conclusion, we evaluated methods for DNA extraction and real-time PCR detection of *C. cayetanensis* in human stool specimens. The UNEX-based DNA extraction followed by the Verweij TaqMan assay is suitable for molecular detection of *C. cayetanensis* in stool and can be used as a complement to microscopy, for example in cases of inconclusive microscopy findings.

Acknowledgments

We thank colleagues at the U.S. State Public Health Laboratories and Dawn Roellig in the Waterborne Disease Prevention Branch (WDPB) of the CDC for sharing stool specimens with our laboratory. We also thank Michael J. Arrowood in the WDPB for quantifying *Cyclospora* oocysts in the stool specimen used for assessing the detection limits; Palmer Orlandi, Jr., from the U.S. Food and Drug Administration (FDA) for providing the *Eimeria* DNA controls; and Francisca Abanyie for critical reviews of earlier versions of this manuscript. The findings and conclusions in this manuscript are those of the authors and do not necessarily represent the official position of the CDC or the FDA.

References

- Berlin OG, Peter JB, Gagne C, Conteas CN, Ash LR. Autofluorescence and the detection of *Cyclospora* oocysts. *Emerg Infect Dis.* 1998; 4:127–128. [PubMed: 9452408]
- Centers for Disease Control and Prevention. Outbreak of cyclosporiasis associated with snow peas--Pennsylvania, 2004. *MMWR Morb Mortal Wkly Rep.* 2004; 53:876–878. [PubMed: 15385921]
- Centers for Disease Control and Prevention. Outbreaks of cyclosporiasis--United States, June–August 2013. *MMWR Morb Mortal Wkly Rep.* 2013; 62:862. [PubMed: 24172881]
- Clopper CJ, Pearson ES. The use of confidence or fiducial limits illustrated in the case of the binomial. *Biometrika.* 1934; 26:404–413.
- da Silva AJ, Bornay-Llinares FJ, Moura IN, Slemenda SB, Tuttle JL, Pieniazek NJ. Fast and reliable extraction of protozoan parasite DNA from fecal specimens. *Mol Diagn.* 1999; 4:57–64. [PubMed: 10229775]
- Duffy T, Cura CI, Ramirez JC, Abate T, Cayo NM, Parrado R, Bello ZD, Velazquez E, Munoz-Calderon A, Juiz NA, Basile J, Garcia L, Riarte A, Nasser JR, Ocampo SB, Yadon ZE, Torrico F, de Noya BA, Ribeiro I, Schijman AG. Analytical performance of a multiplex real-time PCR assay using TaqMan probes for quantification of *Trypanosoma cruzi* satellite DNA in blood samples. *PLoS Negl Trop Dis.* 2013; 7:e2000.doi: 10.1371/journal.pntd.0002000 [PubMed: 23350002]
- Eberhard ML, da Silva AJ, Lilley BG, Pieniazek NJ. Morphologic and molecular characterization of new *Cyclospora* species from Ethiopian monkeys: *C. cercopithecii* sp.n., *C. colobi* sp.n., and *C. papionis* sp.n. *Emerg Infect Dis.* 1999; 5:651–658. [PubMed: 10511521]
- Eberhard ML, Njenga MN, DaSilva AJ, Owino D, Nace EK, Won KY, Mwenda JM. A survey for *Cyclospora* spp. in Kenyan primates, with some notes on its biology. *J Parasitol.* 2001; 87:1394–1397. DOI: 10.1645/0022-3395(2001)087[1394:ASFCSI]2.0.CO;2 [PubMed: 11780827]
- Eberhard ML, Ortega YR, Hanes DE, Nace EK, Do RQ, Robl MG, Won KY, Gavidia C, Sass NL, Mansfield K, Gozalo A, Griffiths J, Gilman R, Sterling CR, Arrowood MJ. Attempts to establish experimental *Cyclospora cayetanensis* infection in laboratory animals. *J Parasitol.* 2000; 86:577–582. [PubMed: 10864257]
- Erickson MC, Ortega YR. Inactivation of protozoan parasites in food, water, and environmental systems. *J Food Prot.* 2006; 69:2786–2808. [PubMed: 17133829]
- Griner PF, Mayewski RJ, Mushlin AI, Greenland P. Principles of test selection and use. *Ann Intern Med.* 1981; 94:559–563.
- Hall RL, Jones JL, Herwaldt BL. Surveillance for laboratory-confirmed sporadic cases of cyclosporiasis--United States, 1997–2008. *MMWR Surveill Summ.* 2011; 60:1–11.
- Hall RL, Jones JL, Hurd S, Smith G, Mahon BE, Herwaldt BL. Population-based active surveillance for *Cyclospora* infection--United States, Foodborne Diseases Active Surveillance Network (FoodNet), 1997–2009. *Clin Infect Dis.* 2012; 54(Suppl 5):S411–417. DOI: 10.1093/cid/cis049 [PubMed: 22572662]
- Herwaldt BL. *Cyclospora cayetanensis*: a review, focusing on the outbreaks of cyclosporiasis in the 1990s. *Clin Infect Dis.* 2000; 31:1040–1057. [PubMed: 11049789]
- Ho AY, Lopez AS, Eberhart MG, Levenson R, Finkel BS, da Silva AJ, Roberts JM, Orlandi PA, Johnson CC, Herwaldt BL. Outbreak of cyclosporiasis associated with imported raspberries, Philadelphia, Pennsylvania, 2000. *Emerg Infect Dis.* 2002; 8:783–788. [PubMed: 12141962]
- Jin Y, Hein MJ, Deddens JA, Hines CJ. Analysis of lognormally distributed exposure data with repeated measures and values below the limit of detection using SAS. *Ann Occup Hyg.* 2011; 55:97–112. DOI: 10.1093/annhyg/meq061 [PubMed: 21177260]
- Lalonde LF, Gajadhar AA. Detection and differentiation of coccidian oocysts by real-time PCR and melting curve analysis. *J Parasitol.* 2011; 97:725–730. DOI: 10.1645/GE-2706.1 [PubMed: 21506835]
- Li N, Ye J, Arrowood MJ, Ma J, Wang L, Xu H, Feng Y, Xiao L. Identification and morphologic and molecular characterization of *Cyclospora macacae* n. sp. from rhesus monkeys in China. *Parasitol Res.* 2015; 114:1811–1816. DOI: 10.1007/s00436-015-4367-5 [PubMed: 25673080]
- Marangi M, Koehler AV, Zanzani SA, Manfredi MT, Brianti E, Giangaspero A, Gasser RB. Detection of *Cyclospora* in captive chimpanzees and macaques by a quantitative PCR-based mutation

scanning approach. *Parasit Vectors*. 2015; 8:274.doi: 10.1186/s13071-015-0872-8 [PubMed: 25972100]

Monteiro L, Bonnemaïson D, Vekris A, Petry KG, Bonnet J, Vidal R, Cabrita J, Megraud F. Complex polysaccharides as PCR inhibitors in feces: *Helicobacter pylori* model. *J Clin Microbiol*. 1997; 35:995–998. [PubMed: 9157172]

Pieniazek NJ, Slemenda SB, da Silva AJ, Alfano EM, Arrowood MJ. PCR confirmation of infection with *Cyclospora cayetanensis*. *Emerg Infect Dis*. 1996; 2:357–359. [PubMed: 8969255]

Relman DA, Schmidt TM, Gajadhar A, Sogin M, Cross J, Yoder K, Sethabutr O, Echeverria P. Molecular phylogenetic analysis of *Cyclospora*, the human intestinal pathogen, suggests that it is closely related to *Eimeria* species. *J Infect Dis*. 1996; 173:440–445. [PubMed: 8568307]

Shields JM, Joo J, Kim R, Murphy HR. Assessment of three commercial DNA extraction kits and a laboratory-developed method for detecting *Cryptosporidium* and *Cyclospora* in raspberry wash, basil wash and pesto. *J Microbiol Methods*. 2013; 92:51–58. DOI: 10.1016/j.mimet.2012.11.001 [PubMed: 23147278]

Varma M, Hester JD, Schaefer FW 3rd, Ware MW, Lindquist HD. Detection of *Cyclospora cayetanensis* using a quantitative real-time PCR assay. *J Microbiol Methods*. 2003; 53:27–36. [PubMed: 12609720]

Verweij JJ, Laeijendecker D, Brienen EA, van Lieshout L, Polderman AM. Detection of *Cyclospora cayetanensis* in travellers returning from the tropics and subtropics using microscopy and real-time PCR. *Int J Med Microbiol*. 2003; 293:199–202. [PubMed: 12868656]

KEY FINDINGS

- We developed an efficient method for the molecular detection of *Cyclospora cayetanensis* in human stool.
- We compared two DNA extraction methods and four real-time PCR methods.
- The UNEX-based method was better than FastDNA for extraction of DNA from *Cyclospora*-positive stool.
- A TaqMan assay that targets the 18S rRNA gene was the best-performing PCR assay.




Fig. 1. High-resolution melt analysis curves using the SYBR Green assay. The first derivative of the fluorescence multiplied by -1 [$-R'(T)$] is plotted against the temperature ($^{\circ}\text{C}$). Blue = simian *Cyclospora* spp. (*C. cercopithecii*, *C. papionis*, and *C. colobi*); red = *C. cayetanensis*; green = *Eimeria acervulina*.




Fig. 2. High-resolution melt analysis curves using the EvaGreen assay. The first derivative of the fluorescence multiplied by -1 [$-R'(T)$] is plotted against the temperature ($^{\circ}\text{C}$). Red = *C. cayetanensis*-positive stools ($T_m = 85.0$ – 85.5); green = *C. cayetanensis*-negative stools with $T_m = 84.0$ – 84.5 ; black = *C. cayetanensis*-negative stools with $T_m = 85.0$ – 85.5 (i.e., false positives).

Table 1

Comparison of the FastDNA and UNEX-based DNA extraction methods for PCR detection of *Cyclospora cayetanensis*.

	No. of stool samples positive (percentage, and 95% confidence interval) in the Verweij assay after DNA extraction using:			
	The FastDNA method		The UNEX-based method	
	Laboratorian A	Laboratorian B	Laboratorian A	Laboratorian B
Microscopy-positive samples (n = 25)	19 (76, 55–91%)	22 (88, 69–97%)	25 (100, 86–100%)	25 (100, 86–100%)
Microscopy-negative samples (n = 25)	1 (4, 0.1–20%)	2 (8, 1–26%)	0 (0, 0–14%)	0 (0, 0–14%)

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Table 2

Estimated mean differences in the Verweij assay Ct values from a right-censored regression model of the performances of the FastDNA and UNEX-based DNA extraction methods, stratified by laboratorian

Laboratorian	Difference in mean Ct value between FastDNA and UNEX (95% confidence interval)	<i>P</i> value
A	4.6 (3.1–6.1)	<0.001
B	2.3 (0.8–3.8)	0.003

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Analytical and diagnostic performance of the conventional and real-time PCR assays for detection of *Cyclospora cayentanensis* following DNA extraction by the UNEX method.

Table 3

Assay	Target gene	Detection mode	Analytical sensitivity (oocysts per 0.3 ml)	Analytical specificity ^a (95% CI)	Diagnostic specificity ^b (95% CI)	Diagnostic sensitivity ^b (95% CI)	Reference
Nested	18S rRNA	Agarose gel	15	100% (83–100%)	100% (96–100%) ^c	97% (84–99%)	(Relman <i>et al.</i> , 1996)
Varma	18S rRNA	TaqMan probe	200	100% (83–100%)	100% (96–100%)	93% (81–98%)	(Varma <i>et al.</i> , 2003)
Verweij	18S rRNA	TaqMan probe	15	100% (83–100%)	100% (96–100%)	100% (87–100%)	(Verweij <i>et al.</i> , 2003)
SYBR Green	18S rRNA	DNA binding dye	15	85% (62–96%)	100% (96–100%)	100% (87–100%)	(Lalonde & Gajadhar, 2011)
EvaGreen	ITS2	DNA binding dye	2	100% (83–100%)	98% (93–99%)	100% (87–100%)	(Marangi <i>et al.</i> , 2015)

^a As determined using 20 samples positive for *Eimeria* and simian *Cyclospora* spp.

^b As determined using 137 human stool specimens (32 *C. cayentanensis*-positive specimens and 105 negative controls)

^c Requires DNA sequencing analysis of the amplicon